


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM082616\
 Data File : BM007294.D
 Acq On : 26 Aug 2016 14:52
 Operator : UM/SJ
 Sample : DFTPP52
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP52

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM081516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Aug 26 23:46:46 2016



AutoFind: Scans 2554, 2555, 2556; Background Corrected with Scan 2547

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.9	85640	PASS
68	69	0.00	2	1.9	1819	PASS
69	198	0.00	100	34.8	96432	PASS
70	69	0.00	2	0.6	621	PASS
127	198	10	80	42.3	117120	PASS
197	198	0.00	2	0.5	1458	PASS
198	198	100	100	100.0	276992	PASS
199	198	5	9	6.7	18666	PASS
275	198	10	60	31.1	86202	PASS
365	198	1	100	4.9	13618	PASS
441	442	0.01	100	15.4	37626	PASS
442	198	50	100	88.4	244757	PASS
443	442	15	24	19.3	47154	PASS

m/z	Abundance
37.00	498.0
38.00	1411.0
39.10	6878.0
40.10	1161.0
41.00	318.0
43.90	1406.0
48.90	1020.0
50.10	28016.0
51.10	105488.0
52.00	5883.0
55.00	864.0
56.00	3013.0
57.00	7409.0
61.00	1450.0
62.00	2042.0
63.00	4518.0
64.10	670.0
65.00	2624.0
68.00	1861.0
69.00	114296.0
70.00	679.0
73.10	1234.0
74.00	13048.0
75.00	19552.0
76.10	6586.0
77.10	131520.0
78.00	8965.0
79.00	10061.0
80.00	7524.0
81.00	10807.0
82.00	2311.0
83.00	1971.0
83.90	424.0
85.00	2069.0
86.00	3538.0
87.00	2045.0
87.90	396.0
88.80	316.0
91.00	2469.0
92.00	2411.0
93.00	19064.0
94.00	1178.0
94.90	398.0
96.00	617.0
98.00	15054.0
99.00	12661.0
100.00	955.0
101.00	7157.0
101.90	319.0
103.00	1867.0
104.00	3139.0
105.00	3834.0
106.00	1066.0
107.00	45744.0
108.00	6426.0
109.10	1199.0
110.00	73016.0
111.00	10463.0
111.90	1478.0
113.00	576.0
116.00	2990.0
117.00	44648.0
118.00	3026.0
120.10	372.0
122.00	3021.0
123.00	4812.0
124.00	1971.0
124.90	1918.0
126.10	655.0
127.00	131648.0
128.00	10053.0
129.00	58792.0
130.00	5328.0
131.00	876.0
132.00	501.0
134.00	2097.0
135.00	4604.0
136.00	1825.0
137.00	2445.0
138.10	606.0
139.10	345.0
140.00	1068.0
141.00	7044.0
142.10	2474.0
143.00	1261.0
144.00	397.0
145.10	320.0
146.00	1768.0
147.00	4387.0
148.00	10481.0
149.00	2116.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

150.00	748.0
151.00	778.0
151.80	644.0
153.00	2544.0
154.00	1886.0
155.00	4624.0
156.00	6579.0
156.90	1096.0
158.00	2262.0
159.00	1467.0
160.00	2980.0
161.00	3283.0
162.00	1426.0
163.00	804.0
164.00	419.0
165.00	3158.0
166.00	2763.0
167.00	16848.0
168.00	10551.0
169.00	1350.0
169.90	436.0
171.00	496.0
172.00	1497.0
173.00	1991.0
174.00	4056.0
175.10	6912.0
176.00	1525.0
177.00	3003.0
177.90	1132.0
178.90	12678.0
180.00	8886.0
181.00	4321.0
182.00	887.0
183.00	321.0
184.00	1266.0
185.00	6168.0
186.00	46272.0
187.00	12734.0
188.00	1754.0
189.00	3634.0
189.80	451.0
191.10	1783.0
192.00	3855.0
193.00	4328.0
193.90	1159.0
194.90	302.0
195.90	6995.0
197.10	1506.0
198.00	298816.0
199.00	21168.0
200.00	1484.0
201.50	1362.0
203.00	2158.0
204.00	12469.0
205.00	22016.0
206.00	86136.0
207.10	11699.0
208.00	3546.0
209.00	1267.0
210.00	1584.0
211.00	3789.0
213.00	360.0
214.90	1408.0
216.00	2162.0
217.00	25632.0
218.00	3311.0
219.80	364.0
221.00	9322.0
221.80	3420.0
223.00	5341.0
224.00	49720.0
225.00	13174.0
226.00	1831.0
227.00	23432.0
228.00	4143.0
229.00	4651.0
229.90	342.0
231.00	1629.0
233.00	505.0
233.90	1497.0
235.10	1289.0
235.90	1014.0
237.00	2437.0
238.00	451.0
238.90	834.0
240.00	805.0
241.00	1480.0
242.00	2488.0
243.10	2588.0
244.10	35496.0
245.10	4554.0
246.00	8555.0
247.00	1745.0
248.00	501.0
248.90	1824.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

252.10	326.0
253.00	958.0
254.10	1345.0
255.00	198464.0
256.00	31320.0
257.00	2969.0
258.00	13391.0
259.00	2270.0
259.80	354.0
261.00	302.0
263.80	346.0
265.00	5785.0
266.00	761.0
270.00	360.0
271.00	512.0
272.00	509.0
273.00	5343.0
274.00	15858.0
275.00	84952.0
276.00	11971.0
277.00	8421.0
278.00	1466.0
278.90	412.0
283.00	1057.0
284.00	601.0
285.00	1500.0
286.00	391.0
288.90	434.0
291.80	439.0
293.00	1786.0
294.00	505.0
296.00	27152.0
297.00	3946.0
297.90	420.0
301.00	373.0
301.90	368.0
303.00	3545.0
304.00	1179.0
309.90	461.0
314.00	974.0
315.00	3446.0
315.90	1749.0
321.00	1023.0
322.00	373.0
323.10	7418.0
324.00	1581.0
327.00	1591.0
328.00	643.0
332.00	648.0
333.00	729.0
334.10	5318.0
335.10	1698.0
341.00	1057.0
346.00	1787.0
347.00	435.0
352.00	2355.0
353.00	1425.0
354.00	2424.0
355.00	389.0
365.00	13156.0
366.00	1344.0
371.00	401.0
372.00	4111.0
373.10	1020.0
382.90	1154.0
390.00	463.0
391.00	335.0
402.00	1461.0
403.00	2070.0
404.10	1215.0
421.00	1525.0
421.90	1635.0
423.10	11146.0
424.00	3332.0
425.10	412.0
441.00	30840.0
442.10	212224.0
443.00	40896.0
444.00	4429.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

m/z	Abundance
37.00	306.0
38.10	1139.0
39.10	6909.0
39.90	1014.0
43.90	1414.0
48.90	1056.0
50.10	29448.0
51.10	98080.0
52.10	4735.0
53.10	393.0
55.10	577.0
56.00	3002.0
57.00	7050.0
61.10	1347.0
62.00	1784.0
63.00	4374.0
64.00	466.0
65.00	2128.0
68.10	2213.0
69.00	112672.0
70.00	768.0
73.10	1187.0
74.00	13171.0
75.00	18376.0
76.00	6044.0
77.00	127160.0
78.10	9014.0
79.10	9207.0
80.00	7322.0
81.00	11498.0
82.10	2767.0
83.00	2259.0
83.90	457.0
85.00	1661.0
86.00	3153.0
87.00	1217.0
88.00	456.0
91.00	1972.0
92.00	2854.0
93.00	18584.0
94.00	1128.0
96.10	838.0
98.00	15615.0
99.00	10825.0
99.90	976.0
101.00	6170.0
103.00	1777.0
104.00	3572.0
105.00	3589.0
106.10	1047.0
107.00	46800.0
108.00	6282.0
109.10	944.0
110.00	70336.0
111.00	12001.0
112.10	1183.0
113.00	831.0
116.00	2919.0
117.00	47880.0
118.00	3402.0
119.00	490.0
120.00	538.0
122.00	4067.0
123.00	5640.0
123.90	1629.0
125.00	1699.0
126.20	557.0
127.00	139392.0
128.00	10547.0
129.00	57272.0
130.00	5281.0
131.00	1018.0
131.90	330.0
133.90	1945.0
135.00	4931.0
136.00	1812.0
137.00	2718.0
137.80	527.0
138.90	402.0
139.10	409.0
140.00	859.0
141.00	7886.0
142.00	2798.0
143.00	1556.0
143.90	439.0
145.00	448.0
145.90	1619.0
147.00	5073.0
148.00	11488.0
149.00	2501.0
150.00	695.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

151.00	1077.0
151.80	684.0
153.00	2455.0
154.00	1945.0
155.00	4652.0
156.00	7549.0
157.00	1275.0
157.90	1752.0
159.00	1132.0
160.00	3045.0
161.00	3914.0
162.00	1281.0
163.00	414.0
164.00	434.0
165.00	3504.0
166.00	2912.0
167.00	19640.0
168.00	10454.0
169.00	1592.0
170.00	780.0
171.00	785.0
172.00	1255.0
173.00	2558.0
174.10	3691.0
175.10	7327.0
176.10	1616.0
176.90	3603.0
178.00	1441.0
179.00	14091.0
180.00	9512.0
181.00	3809.0
181.90	791.0
182.90	314.0
183.90	1134.0
185.00	7630.0
186.00	52568.0
187.00	15313.0
188.10	1508.0
189.00	3093.0
190.00	664.0
191.00	1629.0
192.10	4578.0
193.00	5343.0
194.00	1219.0
194.90	672.0
196.00	8772.0
197.10	1600.0
198.00	326080.0
199.00	22016.0
200.00	2110.0
201.30	1214.0
202.00	592.0
203.00	2864.0
204.10	14067.0
205.00	23320.0
206.00	96760.0
207.00	14412.0
208.00	3922.0
209.00	1289.0
210.20	1385.0
211.00	4590.0
212.00	624.0
213.00	302.0
215.00	1231.0
216.00	2720.0
217.00	28304.0
218.00	3297.0
219.00	447.0
220.00	482.0
221.00	11752.0
221.90	3165.0
223.00	6003.0
224.00	56696.0
225.00	13749.0
226.00	1714.0
227.00	27392.0
228.00	4456.0
229.00	5410.0
230.00	592.0
231.00	1415.0
233.00	571.0
233.90	1581.0
235.00	2298.0
235.90	1705.0
237.00	2606.0
237.90	309.0
239.00	1164.0
239.90	908.0
241.00	1318.0
242.00	3039.0
243.00	3284.0
244.00	42912.0
245.00	5593.0
246.00	10914.0
247.00	2149.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

248.10	417.0
248.90	1655.0
250.90	434.0
252.00	464.0
252.90	700.0
254.00	1392.0
255.00	230976.0
256.00	32288.0
257.00	2494.0
258.00	16233.0
258.90	2397.0
261.10	403.0
263.90	376.0
265.00	6446.0
266.00	1131.0
270.90	433.0
272.00	718.0
273.00	6802.0
274.00	18264.0
275.00	103552.0
276.00	13383.0
277.00	10767.0
277.90	1917.0
279.00	384.0
281.10	419.0
281.90	323.0
283.00	1182.0
284.00	838.0
285.10	2141.0
286.10	517.0
289.00	356.0
289.90	417.0
291.90	378.0
293.00	2722.0
294.00	547.0
295.10	895.0
296.00	36896.0
297.00	5082.0
298.00	547.0
301.90	528.0
303.00	3758.0
304.00	999.0
308.10	406.0
308.90	458.0
310.00	517.0
314.00	1613.0
315.00	4663.0
316.00	2077.0
317.00	515.0
321.00	921.0
322.00	633.0
323.00	8875.0
324.00	1260.0
327.00	1892.0
327.90	1019.0
332.00	869.0
333.00	1243.0
334.00	7020.0
334.90	1622.0
341.00	976.0
342.00	480.0
346.00	1785.0
347.10	343.0
352.00	2542.0
353.00	1645.0
354.10	3430.0
355.00	677.0
365.00	15283.0
366.00	2395.0
370.90	659.0
372.00	5031.0
373.00	1302.0
383.00	1225.0
384.10	435.0
390.10	641.0
390.80	315.0
391.10	311.0
391.90	434.0
402.00	1735.0
403.00	3101.0
404.00	1309.0
421.00	2380.0
422.00	2050.0
423.10	16145.0
424.10	3455.0
425.00	614.0
441.00	45584.0
442.00	290304.0
443.00	55560.0
444.00	5138.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

m/z	Abundance
38.00	829.0
39.00	3894.0
39.90	917.0
44.00	1600.0
49.00	685.0
50.10	13788.0
51.00	53352.0
52.10	2623.0
55.00	316.0
56.00	1473.0
57.00	3568.0
61.00	892.0
62.00	705.0
63.10	2545.0
63.90	348.0
65.10	1482.0
68.00	1383.0
69.00	62328.0
70.00	416.0
73.10	843.0
74.00	7604.0
75.00	10266.0
76.10	3672.0
77.10	69160.0
78.00	4246.0
79.00	4914.0
80.00	3448.0
81.00	6070.0
82.00	1448.0
83.00	1411.0
85.00	1297.0
86.00	1885.0
87.10	826.0
87.90	310.0
91.00	1195.0
91.90	1606.0
93.00	10435.0
94.00	699.0
96.00	542.0
97.20	353.0
98.00	8720.0
99.00	6169.0
100.00	550.0
101.00	3915.0
103.00	1169.0
104.00	1974.0
105.00	2224.0
106.00	744.0
107.00	26704.0
108.00	3902.0
109.10	647.0
110.00	41936.0
111.00	6143.0
111.90	731.0
112.80	499.0
116.00	1227.0
117.00	29576.0
118.00	1750.0
119.90	429.0
122.10	2166.0
123.00	2301.0
124.00	1225.0
125.00	979.0
126.20	493.0
127.00	80320.0
128.00	6546.0
129.00	34856.0
130.00	2825.0
130.90	632.0
134.00	1176.0
135.00	2900.0
136.00	1326.0
137.10	2036.0
140.10	477.0
141.00	5436.0
142.00	1401.0
143.10	747.0
145.90	1184.0
147.00	3154.0
148.00	6076.0
149.00	1424.0
149.90	304.0
151.00	566.0
152.00	321.0
153.00	1698.0
154.00	1320.0
155.00	2868.0
156.10	4411.0
157.00	889.0
157.90	1591.0
159.00	807.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

160.00	1459.0
161.00	2563.0
162.00	676.0
164.90	2004.0
166.00	1280.0
167.00	11518.0
168.00	6603.0
169.00	1257.0
170.90	485.0
172.00	1733.0
173.00	1550.0
174.00	2256.0
175.00	3812.0
176.00	1070.0
177.00	2452.0
178.00	786.0
179.00	8616.0
180.00	5593.0
181.00	2563.0
181.90	570.0
183.90	1173.0
185.00	4193.0
186.00	31704.0
187.00	8794.0
188.10	1063.0
189.00	2096.0
189.90	530.0
190.90	1587.0
192.00	3389.0
193.00	3679.0
194.00	664.0
196.00	6225.0
197.00	1269.0
198.00	206080.0
199.00	12816.0
199.90	1174.0
201.60	737.0
203.00	1649.0
204.00	8923.0
205.00	12860.0
206.00	60104.0
207.00	9106.0
208.00	2800.0
209.00	930.0
210.10	980.0
211.10	2540.0
211.90	393.0
215.00	851.0
215.90	1059.0
217.00	18832.0
218.00	2567.0
219.90	412.0
221.00	7350.0
222.00	2601.0
223.00	4092.0
224.00	36880.0
225.00	9072.0
226.00	925.0
227.00	18248.0
227.90	2580.0
229.00	3768.0
230.10	307.0
231.10	1180.0
234.00	1202.0
235.00	1128.0
235.90	712.0
237.10	1250.0
237.80	386.0
239.00	633.0
239.90	465.0
241.00	777.0
242.00	1707.0
243.00	2023.0
244.00	27224.0
245.10	3215.0
246.00	6520.0
246.90	1069.0
249.00	914.0
251.00	355.0
251.90	308.0
252.90	696.0
254.00	733.0
255.00	155776.0
256.00	22000.0
257.00	2002.0
258.00	11989.0
259.00	1689.0
261.00	347.0
265.00	4471.0
266.10	927.0
270.90	393.0
272.00	490.0
273.00	4510.0
274.00	11731.0
275.00	70104.0

Instrument :

BNA_M

ClientSampleId :

DFTPP52

276.00	8857.0
277.00	6326.0
278.00	1092.0
281.90	345.0
283.00	901.0
284.00	450.0
285.10	1383.0
288.90	380.0
290.00	358.0
292.00	393.0
293.00	2008.0
294.00	569.0
295.00	751.0
296.00	24624.0
297.00	3688.0
301.90	405.0
303.00	2988.0
304.10	679.0
307.80	341.0
314.10	1140.0
315.00	2058.0
316.00	1692.0
321.00	832.0
323.00	7094.0
324.00	1396.0
327.00	1324.0
328.00	884.0
332.00	512.0
332.90	816.0
334.00	4638.0
334.90	1087.0
341.00	1057.0
345.90	1735.0
350.90	340.0
352.00	2501.0
353.00	1544.0
354.10	2783.0
355.10	515.0
365.00	12417.0
366.00	1530.0
371.00	646.0
372.10	4055.0
373.10	914.0
383.00	854.0
383.90	492.0
390.00	527.0
390.90	303.0
401.00	314.0
402.00	1830.0
403.00	2424.0
404.00	970.0
421.00	1820.0
422.00	1809.0
423.00	13579.0
424.00	2864.0
424.90	477.0
441.00	36456.0
442.00	231744.0
443.10	45008.0
444.00	3997.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

Scan 2547 (17.568 min): BM007294.D
DFTPP52

m/z	Abundance
39.90	764.0
44.00	1762.0
73.00	346.0
207.00	1283.0
208.00	351.0
280.90	397.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52