

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082622\
 Data File : BM036434.D
 Acq On : 26 Aug 2022 20:55
 Operator : CG/JU
 Sample : PB147281BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK281

Quant Time: Aug 26 23:10:10 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM082622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 14:26:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.791	152	4112	0.400	ng/ul	0.00
4) Naphthalene-d8	10.589	136	16425	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.432	164	9808	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.174	188	21033	0.400	ng/ul	0.00
17) Chrysene-d12	21.371	240	22152	0.400	ng/ul	# 0.00
23) Perylene-d12	23.683	264	20951	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.201	96	33548	6.058	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.206	152	8650	0.329	ng/ul	0.02
18) Fluoranthene-d10	19.205	212	19495	0.712	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082622\
Data File : BM036434.D
Acq On : 26 Aug 2022 20:55
Operator : CG/JU
Sample : PB147281BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK281

Quant Time: Aug 26 23:10:10 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM082622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 26 14:26:34 2022
Response via : Initial Calibration

