

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082624\
 Data File : BM047352.D
 Acq On : 26 Aug 2024 18:23
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM082624

Quant Time: Aug 27 04:27:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 26 18:11:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.427	0.421	1.4	104	0.00
3	Pyridine	1.184	1.181	0.3	103	0.00
4	n-Nitrosodimethylamine	0.482	0.490	-1.7	103	0.00
5 S	2-Fluorophenol	1.042	1.031	1.1	103	0.00
6	Aniline	1.286	1.311	-1.9	103	0.00
7 S	Phenol-d6	1.420	1.426	-0.4	104	0.00
8	2-Chlorophenol	1.167	1.163	0.3	104	0.00
9	Benzaldehyde	0.783	0.793	-1.3	109	0.00
10 C	Phenol	1.397	1.398	-0.1	104	0.00
11	bis(2-Chloroethyl)ether	1.211	1.216	-0.4	104	0.00
12	1,3-Dichlorobenzene	1.427	1.434	-0.5	103	0.00
13 C	1,4-Dichlorobenzene	1.445	1.449	-0.3	103	0.00
14	1,2-Dichlorobenzene	1.369	1.365	0.3	103	0.00
15	Benzyl Alcohol	0.985	1.033	-4.9	105	0.00
16	2,2'-oxybis(1-Chloropropane	1.452	1.469	-1.2	102	-0.02
17	2-Methylphenol	0.971	0.996	-2.6	103	0.00
18	Hexachloroethane	0.540	0.547	-1.3	104	0.00
19 P	n-Nitroso-di-n-propylamine	0.976	0.963	1.3	104	0.00
20	3+4-Methylphenols	1.350	1.392	-3.1	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
22	Acetophenone	0.464	0.470	-1.3	103	0.00
23 S	Nitrobenzene-d5	0.360	0.372	-3.3	103	0.00
24	Nitrobenzene	0.371	0.383	-3.2	102	0.00
25	Isophorone	0.681	0.700	-2.8	103	0.00
26 C	2-Nitrophenol	0.182	0.195	-7.1	103	0.00
27	2,4-Dimethylphenol	0.206	0.216	-4.9	105	0.00
28	bis(2-Chloroethoxy)methane	0.417	0.434	-4.1	104	0.00
29 C	2,4-Dichlorophenol	0.319	0.345	-8.2	104	0.00
30	1,2,4-Trichlorobenzene	0.369	0.381	-3.3	104	0.00
31	Naphthalene	0.977	0.997	-2.0	104	0.00
32	Benzoic acid	0.246	0.266	-8.1	107	0.00
33	4-Chloroaniline	0.363	0.383	-5.5	105	0.00
34 C	Hexachlorobutadiene	0.230	0.239	-3.9	104	0.00
35	Caprolactam	0.103	0.106	-2.9	105	0.00
36 C	4-Chloro-3-methylphenol	0.324	0.337	-4.0	104	0.00
37	2-Methylnaphthalene	0.712	0.726	-2.0	104	0.00
38	1-Methylnaphthalene	0.704	0.714	-1.4	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	0.592	0.633	-6.9	104	0.00
41 P	Hexachlorocyclopentadiene	0.187	0.200	-7.0	100	0.00
42 S	2,4,6-Tribromophenol	0.241	0.249	-3.3	104	0.00
43 C	2,4,6-Trichlorophenol	0.406	0.437	-7.6	104	0.00
44	2,4,5-Trichlorophenol	0.451	0.484	-7.3	104	0.00
45 S	2-Fluorobiphenyl	1.302	1.359	-4.4	104	0.00
46	1,1'-Biphenyl	1.397	1.462	-4.7	104	0.00
47	2-Chloronaphthalene	1.094	1.138	-4.0	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082624\
 Data File : BM047352.D
 Acq On : 26 Aug 2024 18:23
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM082624

Quant Time: Aug 27 04:27:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 26 18:11:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.307	0.326	-6.2	103	0.00
49	Acenaphthylene	1.591	1.654	-4.0	103	0.00
50	Dimethylphthalate	1.378	1.419	-3.0	104	0.00
51	2,6-Dinitrotoluene	0.321	0.340	-5.9	104	0.00
52 C	Acenaphthene	1.010	1.043	-3.3	104	0.00
53	3-Nitroaniline	0.296	0.316	-6.8	104	0.00
54 P	2,4-Dinitrophenol	0.198	0.205	-3.5	101	0.00
55	Dibenzofuran	1.626	1.663	-2.3	103	0.00
56 P	4-Nitrophenol	0.231	0.246	-6.5	105	0.00
57	2,4-Dinitrotoluene	0.415	0.428	-3.1	103	0.00
58	Fluorene	1.316	1.346	-2.3	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.370	0.387	-4.6	103	0.00
60	Diethylphthalate	1.359	1.388	-2.1	103	0.00
61	4-Chlorophenyl-phenylether	0.670	0.681	-1.6	103	0.00
62	4-Nitroaniline	0.274	0.285	-4.0	102	0.00
63	Azobenzene	1.184	1.214	-2.5	103	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.123	0.133	-8.1	101	0.00
66 c	n-Nitrosodiphenylamine	0.547	0.569	-4.0	103	0.00
67	4-Bromophenyl-phenylether	0.211	0.222	-5.2	104	0.00
68	Hexachlorobenzene	0.244	0.255	-4.5	104	0.00
69	Atrazine	0.179	0.158	11.7	102	0.00
70 C	Pentachlorophenol	0.158	0.169	-7.0	103	0.00
71	Phenanthrene	0.971	0.993	-2.3	103	0.00
72	Anthracene	0.947	0.979	-3.4	103	0.00
73	Carbazole	0.909	0.937	-3.1	103	0.00
74	Di-n-butylphthalate	1.106	1.138	-2.9	102	0.00
75 C	Fluoranthene	1.140	1.153	-1.1	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
77	Benzydine	0.334	0.407	-21.9	83	0.00
78	Pyrene	1.534	1.580	-3.0	102	0.00
79 S	Terphenyl-d14	1.124	1.174	-4.4	103	0.00
80	Butylbenzylphthalate	0.600	0.627	-4.5	101	0.00
81	Benzo(a)anthracene	1.304	1.359	-4.2	103	0.00
82	3,3'-Dichlorobenzidine	0.439	0.471	-7.3	101	0.00
83	Chrysene	1.227	1.267	-3.3	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.848	0.880	-3.8	102	0.00
85 c	Di-n-octyl phthalate	1.428	1.495	-4.7	101	0.00
86 I	Perylene-d12	1.000	1.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	1.252	1.317	-5.2	104	0.00
88	Benzo(b)fluoranthene	1.167	1.196	-2.5	102	0.00
89	Benzo(k)fluoranthene	1.106	1.147	-3.7	105	0.00
90 C	Benzo(a)pyrene	1.009	1.052	-4.3	103	0.00
91	Dibenzo(a,h)anthracene	1.014	1.064	-4.9	104	0.00
92	Benzo(g,h,i)perylene	1.054	1.116	-5.9	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082624\
Data File : BM047352.D
Acq On : 26 Aug 2024 18:23
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM082624

Quant Time: Aug 27 04:27:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 26 18:11:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0