

Data Path : Z:\HPCHEM1\BNA_M\Data\BM082715\
 Data File : BM002557.D
 Acq On : 27 Aug 2015 16:14
 Operator : UM/IZ
 Sample : SSTD1.660
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD1.660

Quant Time: Aug 27 16:46:37 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM082715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 27 16:29:33 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.78	152	9424	0.40	ng/ul	0.00
2) Naphthalene-d8	11.64	136	38040	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.36	164	18862	0.40	ng/ul	0.00
10) Phenanthrene-d10	18.08	188	36436	0.40	ng/ul	0.00
16) Chrysene-d12	22.16	240	26444	0.40	ng/ul	0.00
20) Perylene-d12	24.78	264	23807	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	13.17	152	80822	1.55	ng/ul	0.00
14) Fluoranthene-d10	20.04	212	129219	1.51	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	11.69	128	156262	1.59	ng/ul	98
5) 2-Methylnaphthalene	13.25	142	105448	1.57	ng/ul	100
7) Acenaphthylene	15.09	152	159486	1.64	ng/ul	94
8) Acenaphthene	15.42	153	111539	1.61	ng/ul	86
9) Fluorene	16.39	166	123684	1.55	ng/ul	93
11) Pentachlorophenol	17.72	266	7100	2.34	ng/ul	99
12) Phenanthrene	18.11	178	175264	1.61	ng/ul	98
13) Anthracene	18.20	178	173723	1.65	ng/ul	98
15) Fluoranthene	20.06	202	174747	1.50	ng/ul	97
17) Pyrene	20.42	202	174501	1.57	ng/ul	94
18) Benzo(a)anthracene	22.14	228	136484	1.59	ng/ul	93
19) Chrysene	22.19	228	132512	1.62	ng/ul	97
21) Benzo(b)fluoranthene	23.96	252	130055	1.46	ng/ul	97
22) Benzo(k)fluoranthene	24.01	252	128188	1.62	ng/ul	98
23) Benzo(a)pyrene	24.65	252	124851	1.62	ng/ul	96
24) Indeno(1,2,3-cd)pyrene	27.53	276	147491	1.72	ng/ul	98
25) Dibenzo(a,h)anthracene	27.54	278	120942	1.73	ng/ul	95
26) Benzo(g,h,i)perylene	28.39	276	125904	1.71	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM082715\
Data File : BM002557.D
Acq On : 27 Aug 2015 16:14
Operator : UM/IZ
Sample : SSTD1.660
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD1.660

Quant Time: Aug 27 16:46:37 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM082715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 27 16:29:33 2015
Response via : Initial Calibration

