

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM082721\  
 Data File : BM031924.D  
 Acq On : 27 Aug 2021 12:34  
 Operator : CG/JU  
 Sample : PB138731BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 PB138731BL

Quant Time: Aug 27 14:21:13 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM082621.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 26 14:24:44 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.510  | 152  | 66893    | 20.000  | ng    | -0.01    |
| 21) Naphthalene-d8          | 10.269 | 136  | 275458   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.151 | 164  | 173626   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 16.910 | 188  | 356804   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.115 | 240  | 300906   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 23.262 | 264  | 273668   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.152  | 112  | 693915   | 165.688 | ng    | 0.00     |
| 7) Phenol-d6                | 6.716  | 99   | 878131   | 146.227 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 8.663  | 82   | 573037   | 94.101  | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 15.657 | 330  | 289783   | 154.612 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 12.769 | 172  | 1218357  | 94.900  | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.562 | 244  | 1904551  | 101.751 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM082721\  
Data File : BM031924.D  
Acq On : 27 Aug 2021 12:34  
Operator : CG/JU  
Sample : PB138731BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
PB138731BL

Quant Time: Aug 27 14:21:13 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM082621.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Aug 26 14:24:44 2021  
Response via : Initial Calibration

