

Data Path : Z:\HPCHEM1\BNA_M\Data\BM082915\
 Data File : BM002560.D
 Acq On : 29 Aug 2015 12:27
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02062

Quant Time: Aug 31 01:07:48 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 31 01:07:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	138	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
3	Naphthalene	1.021	0.962	5.8	126	0.00
4	2-Methylnaphthalene	0.765	0.708	7.5	123	0.00
5	1-Methylnaphthalene	0.756	0.693	8.3	121	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
7 S	Acenaphthylene-d8	2.045	1.948	4.7	119	0.00
8	Acenaphthylene	2.140	2.042	4.6	120	0.00
9 C	Acenaphthene	1.410	1.348	4.4	121	0.00
10 S	Fluorene-d10	1.436	1.343	6.5	118	0.00
11	Fluorene	1.615	1.535	5.0	119	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
13	Phenanthrene	1.155	1.110	3.9	115	0.00
14 S	Anthracene-d10	0.965	0.919	4.8	114	0.00
15	Anthracene	1.170	1.113	4.9	113	0.00
16 C	Fluoranthene	1.308	1.145	12.5	107	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
18 S	Pyrene-d10	0.887	1.015	-14.4	106	0.00
19	Pyrene	1.172	1.355	-15.6	106	0.00
20	Benzo(a)anthracene	1.196	1.122	6.2	93	0.00
21	Chrysene	1.119	1.047	6.4	92	0.00
22 I	Perylene-d12	1.000	1.000	0.0	77	0.00
23	Benzo(b)fluoranthene	1.209	1.181	2.3	77	0.00
24	Benzo(k)fluoranthene	1.146	1.133	1.1	76	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.989	5.9	74	0.00
26 C	Benzo(a)pyrene	1.155	1.092	5.5	73	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.175	11.5	69	-0.02
28	Dibenzo(a,h)anthracene	1.110	0.988	11.0	69	-0.02
29	Benzo(g,h,i)perylene	1.106	0.976	11.8	69	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0