

Data Path : Z:\HPCHEM1\BNA_M\Data\BM082915\
 Data File : BM002576.D
 Acq On : 29 Aug 2015 22:09
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02063

Quant Time: Aug 31 01:13:59 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 31 01:07:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	156#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
3	Naphthalene	1.021	0.964	5.6	119	0.00
4	2-Methylnaphthalene	0.765	0.636	16.9	105	0.00
5	1-Methylnaphthalene	0.756	0.614	18.8	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
7 S	Acenaphthylene-d8	2.045	1.963	4.0	88	0.00
8	Acenaphthylene	2.140	2.059	3.8	89	0.00
9 C	Acenaphthene	1.410	1.328	5.8	87	0.00
10 S	Fluorene-d10	1.436	1.258	12.4	81	0.00
11	Fluorene	1.615	1.435	11.1	82	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	76	0.00
13	Phenanthrene	1.155	1.093	5.4	73	0.00
14 S	Anthracene-d10	0.965	0.917	5.0	73	0.00
15	Anthracene	1.170	1.115	4.7	73	0.00
16 C	Fluoranthene	1.308	1.128	13.8	67	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	72	0.00
18 S	Pyrene-d10	0.887	0.891	-0.5	68	0.00
19	Pyrene	1.172	1.185	-1.1	68	0.00
20	Benzo(a)anthracene	1.196	1.131	5.4	69	0.00
21	Chrysene	1.119	1.059	5.4	68	0.00
22 I	Perylene-d12	1.000	1.000	0.0	76	0.00
23	Benzo(b)fluoranthene	1.209	1.137	6.0	74	0.00
24	Benzo(k)fluoranthene	1.146	1.043	9.0	69	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.981	6.7	72	0.00
26 C	Benzo(a)pyrene	1.155	1.086	6.0	72	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.307	1.6	76	0.00
28	Dibenzo(a,h)anthracene	1.110	1.110	0.0	77	0.00
29	Benzo(g,h,i)perylene	1.106	1.095	1.0	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0