

Data Path : Z:\HPCHEM1\BNA_M\Data\BM082915\
 Data File : BM002585.D
 Acq On : 30 Aug 2015 03:35
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02064

Quant Time: Aug 31 01:37:37 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 31 01:34:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	188#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	161#	0.00
3	Naphthalene	1.021	0.965	5.5	156#	0.00
4	2-Methylnaphthalene	0.765	0.660	13.7	142	0.00
5	1-Methylnaphthalene	0.756	0.642	15.1	139	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	127	0.00
7 S	Acenaphthylene-d8	2.045	1.941	5.1	121	0.00
8	Acenaphthylene	2.140	2.046	4.4	122	0.00
9 C	Acenaphthene	1.410	1.332	5.5	122	0.00
10 S	Fluorene-d10	1.436	1.243	13.4	111	0.00
11	Fluorene	1.615	1.413	12.5	112	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
13	Phenanthrene	1.155	1.098	4.9	93	0.00
14 S	Anthracene-d10	0.965	0.906	6.1	92	0.00
15	Anthracene	1.170	1.101	5.9	92	0.00
16 C	Fluoranthene	1.308	1.004	23.2	76	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
18 S	Pyrene-d10	0.887	0.919	-3.6	75	0.00
19	Pyrene	1.172	1.209	-3.2	74	0.00
20	Benzo(a)anthracene	1.196	1.121	6.3	73	0.00
21	Chrysene	1.119	1.056	5.6	73	0.00
22 I	Perylene-d12	1.000	1.000	0.0	80	0.01
23	Benzo(b)fluoranthene	1.209	1.103	8.8	75	0.01
24	Benzo(k)fluoranthene	1.146	1.079	5.8	75	0.01
25 S	Benzo(a)pyrene-d12	1.051	0.984	6.4	76	0.01
26 C	Benzo(a)pyrene	1.155	1.085	6.1	76	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.312	1.2	80	0.02
28	Dibenzo(a,h)anthracene	1.110	1.104	0.5	80	0.02
29	Benzo(g,h,i)perylene	1.106	1.106	0.0	81	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0