

Data Path : Z:\HPCHEM1\BNA_M\Data\BM083016\
 Data File : BM007342.D
 Acq On : 30 Aug 2016 13:16
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02038

Quant Time: Aug 30 14:48:41 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM081516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 23:46:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	61	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
3	Naphthalene	1.004	1.017	-1.3	73	0.00
4 SURR	2-Methylnaphthalene-d10	0.500	0.558	-11.6	80	-0.01
5	2-Methylnaphthalene	0.676	0.755	-11.7	81	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
7	Acenaphthylene	2.166	1.978	8.7	90	0.00
8 C	Acenaphthene	1.392	1.467	-5.4	103	0.00
9	Fluorene	1.578	1.764	-11.8	111	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	123	-0.02
11	Pentachlorophenol	0.100	0.099	1.0	129	0.00
12	Phenanthrene	1.136	1.191	-4.8	121	0.00
13	Anthracene	1.085	1.113	-2.6	120	0.00
14 SURR	Fluoranthene-d10	0.901	1.000	-11.0	129	0.00
15 C	Fluoranthene	1.320	1.455	-10.2	129	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	141	0.00
17	Pyrene	1.329	1.380	-3.8	132	0.00
18	Benzo(a)anthracene	1.264	1.306	-3.3	133	0.00
19	Chrysene	1.215	1.349	-11.0	140	0.00
20 I	Perylene-d12	1.000	1.000	0.0	138	0.00
21	Benzo(b)fluoranthene	1.482	1.417	4.4	122	0.00
22	Benzo(k)fluoranthene	1.221	1.431	-17.2	143	0.00
23 C	Benzo(a)pyrene	1.272	1.365	-7.3	133	0.00
24	Indeno(1,2,3-cd)pyrene	1.584	1.436	9.3	114	0.00
25	Dibenzo(a,h)anthracene	1.246	1.138	8.7	114	0.00
26	Benzo(g,h,i)perylene	1.345	1.214	9.7	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0