

Data Path : Z:\HPCHEM1\BNA_M\Data\BM083016\
 Data File : BM007351.D
 Acq On : 30 Aug 2016 21:05
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02039

Quant Time: Aug 31 10:30:24 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM081516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 23:46:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	53	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
3	Naphthalene	1.004	1.014	-1.0	65	0.00
4 SURR	2-Methylnaphthalene-d10	0.500	0.559	-11.8	72	-0.01
5	2-Methylnaphthalene	0.676	0.757	-12.0	73	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.02
7	Acenaphthylene	2.166	1.900	12.3	80	0.00
8 C	Acenaphthene	1.392	1.445	-3.8	94	0.00
9	Fluorene	1.578	1.760	-11.5	102	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	116	-0.02
11	Pentachlorophenol	0.100	0.089	11.0	108	0.00
12	Phenanthrene	1.136	1.214	-6.9	117	0.00
13	Anthracene	1.085	1.085	0.0	110	0.00
14 SURR	Fluoranthene-d10	0.901	1.001	-11.1	122	-0.01
15 C	Fluoranthene	1.320	1.488	-12.7	124	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	128	-0.01
17	Pyrene	1.329	1.429	-7.5	125	0.00
18	Benzo(a)anthracene	1.264	1.277	-1.0	119	-0.01
19	Chrysene	1.215	1.408	-15.9	134	-0.01
20 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
21	Benzo(b)fluoranthene	1.482	1.523	-2.8	117	-0.01
22	Benzo(k)fluoranthene	1.221	1.418	-16.1	126	-0.01
23 C	Benzo(a)pyrene	1.272	1.372	-7.9	119	-0.01
24	Indeno(1,2,3-cd)pyrene	1.584	1.452	8.3	103	-0.01
25	Dibenzo(a,h)anthracene	1.246	1.144	8.2	101	-0.02
26	Benzo(g,h,i)perylene	1.345	1.236	8.1	103	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0