

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM083021\
 Data File : BM031961.D
 Acq On : 30 Aug 2021 22:13
 Operator : CG/JU
 Sample : PB138565BS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS565

Manual Integrations
 APPROVED

mohammad
 8/31/2021 11:37:28 AM

Quant Time: Aug 31 02:37:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM082821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 30 09:21:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.510	152	82066	20.000	ng/u1	0.00
20) Naphthalene-d8	10.269	136	362175	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.151	164	279189	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.910	188	603095	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.115	240	658095	20.000	ng/u1	# 0.00
88) Perylene-d12	23.250	264	632050	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.134	96	12048	6.774	ng/uL	0.00
4) Pyridine-d5	3.522	84	174883	34.625	ng/u1	0.00
7) Phenol-d5	6.704	99	242171	35.586	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.869	67	135121	34.317	ng/u1	0.00
11) 2-Chlorophenol-d4	7.051	132	199753	38.145	ng/u1	0.00
15) 4-Methylphenol-d8	8.222	113	206297	35.334	ng/u1	0.00
21) Nitrobenzene-d5	8.663	128	100048	38.033	ng/u1	0.00
24) 2-Nitrophenol-d4	9.369	143	122195	39.251	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.898	165	242667	38.851	ng/u1	0.00
31) 4-Chloroaniline-d4	10.422	131	300657	34.124	ng/u1	0.00
46) Dimethylphthalate-d6	13.580	166	770335	36.465	ng/u1	0.00
49) Acenaphthylene-d8	13.839	160	943797	37.724	ng/u1	0.00
54) 4-Nitrophenol-d4	14.392	143	137805	36.038	ng/u1	0.00
60) Fluorene-d10	15.151	176	680637	36.652	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.298	200	152853	36.885	ng/u1	0.00
73) Anthracene-d10	17.004	188	1092827	38.176	ng/u1	0.00
81) Pyrene-d10	19.309	212	1242388	40.305	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.121	264	1250929	37.658	ng/u1	0.00
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.163	88	27121	12.648	ng/uL	88
5) Pyridine	3.540	79	187621	36.132	ng/u1	91
6) Benzaldehyde	6.675	77	149020m	48.371	ng/u1	
8) Phenol	6.728	94	269981	39.426	ng/u1	95
10) Bis(2-Chloroethyl)ether	6.957	93	202361	38.799	ng/u1	91
12) 2-Chlorophenol	7.081	128	220467	42.160	ng/u1	96
13) 2-Methylphenol	7.957	108	205968	39.010	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.034	45	257472	37.986	ng/u1#	93
16) Acetophenone	8.334	105	348238	38.082	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.322	70	184728	41.408	ng/u1	94
18) 4-Methylphenol	8.287	108	229620	39.163	ng/u1	97
19) Hexachloroethane	8.557	117	91086	40.662	ng/u1	92
22) Nitrobenzene	8.704	77	267354	40.257	ng/u1	99
23) Isophorone	9.228	82	497666	40.716	ng/u1#	96
25) 2-Nitrophenol	9.404	139	138677	42.761	ng/u1#	92
26) 2,4-Dimethylphenol	9.469	107	264053	39.999	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.710	93	296255	40.888	ng/u1	99
29) 2,4-Dichlorophenol	9.928	162	247708	41.693	ng/u1	98
30) Naphthalene	10.316	128	728070	40.156	ng/u1	99
32) 4-Chloroaniline	10.445	127	325557	37.041	ng/u1	98
33) Hexachlorobutadiene	10.592	225	159316	40.421	ng/u1	97
34) Caprolactam	11.245	113	79869m	40.330	ng/u1	
35) 4-Chloro-3-methylphenol	11.581	107	265579	42.626	ng/u1	99
36) 2-Methylnaphthalene	11.939	142	580322	40.594	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM083021\
 Data File : BM031961.D
 Acq On : 30 Aug 2021 22:13
 Operator : CG/JU
 Sample : PB138565BS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS565


Manual Integrations
 APPROVED

mohammad
 8/31/2021 11:37:28 AM

Quant Time: Aug 31 02:37:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM082821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 30 09:21:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.157	142	558176	38.334	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.310	216	328776	41.024	ng/ul	99
40) Hexachlorocyclopentadiene	12.280	237	154933	35.917	ng/ul	98
41) 2,4,6-Trichlorophenol	12.569	196	223619	41.651	ng/ul	98
42) 2,4,5-Trichlorophenol	12.639	196	246702	42.313	ng/ul	99
43) 1,1'-Biphenyl	12.975	154	781693	39.804	ng/ul	98
44) 2-Chloronaphthalene	13.010	162	627537	40.334	ng/ul	98
45) 2-Nitroaniline	13.239	65	181713	42.393	ng/ul	98
47) Dimethylphthalate	13.627	163	832318	40.124	ng/ul	99
48) 2,6-Dinitrotoluene	13.751	165	178565	43.634	ng/ul	96
50) Acenaphthylene	13.869	152	928820	40.347	ng/ul	98
51) 3-Nitroaniline	14.080	138	165533	43.224	ng/ul	96
52) Acenaphthene	14.216	153	674649	40.351	ng/ul	98
53) 2,4-Dinitrophenol	14.298	184	111814	39.078	ng/ul	95
55) 4-Nitrophenol	14.404	109	141129	39.978	ng/ul	86
56) Dibenzofuran	14.557	168	986506	40.009	ng/ul	100
57) 2,4-Dinitrotoluene	14.551	165	263860	42.990	ng/ul#	82
58) 2,3,4,6-Tetrachlorophenol	14.786	232	222926	42.592	ng/ul#	96
59) Diethylphthalate	15.004	149	871556	41.109	ng/ul	98
61) Fluorene	15.210	166	847625	40.628	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.210	204	428726	40.124	ng/ul	98
63) 4-Nitroaniline	15.257	138	164471m	47.945	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.316	198	162364	40.295	ng/ul#	99
67) N-Nitrosodiphenylamine	15.433	169	740543	42.758	ng/ul	97
68) 4-Bromophenyl-phenylether	16.104	248	265278	42.923	ng/ul	96
69) Hexachlorobenzene	16.210	284	306430	42.704	ng/ul	99
70) Atrazine	16.398	200	284827	43.648	ng/ul	96
71) Pentachlorophenol	16.563	266	192589	41.043	ng/ul	97
72) Phenanthrene	16.951	178	1376261	41.782	ng/ul	99
74) Anthracene	17.039	178	1416844	42.029	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.933	216	331126	42.192	ng/uL	99
76) Pentachlorobenzene	14.469	250	336977	40.928	ng/uL	98
77) Carbazole	17.321	167	1254379	40.656	ng/ul	99
78) Di-n-butylphthalate	17.892	149	1571574	44.452	ng/ul	99
80) Fluoranthene	18.974	202	1669483	44.572	ng/ul	95
82) Pyrene	19.339	202	1732675	43.693	ng/ul	95
83) Butylbenzylphthalate	20.262	149	745396	47.870	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.045	252	633672	42.771	ng/ul	95
85) Benzo(a)anthracene	21.098	228	1724427	42.125	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.045	149	1200444	47.803	ng/ul#	97
87) Chrysene	21.150	228	1666649	42.328	ng/ul	99
89) Di-n-octyl phthalate	21.903	149	2046458	47.459	ng/ul	100
90) Benzo(b)fluoranthene	22.621	252	1751243	43.488	ng/ul	98
91) Benzo(k)fluoranthene	22.662	252	1757664	45.086	ng/ul	98
93) Benzo(a)pyrene	23.162	252	1518233	41.751	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	25.368	276	1676246	36.017	ng/ul	96
95) Dibenzo(a,h)anthracene	25.374	278	1445291	36.925	ng/ul#	96
96) Benzo(g,h,i)perylene	26.003	276	1320431	34.262	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
1,4-Dioxane
Concen: 12.648 ng/uL
RT: 3.163 min Scan# 13
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion: 88 Resp: 27121
Ion Ratio Lower Upper
88 100
43 34.9 31.5 47.3
58 68.1 64.1 96.1

Instrument :
BNA_M
ClientSampled :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#5
Pyridine
Concen: 36.132 ng/uL
RT: 3.540 min Scan# 77
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion: 79 Resp: 187621
Ion Ratio Lower Upper
79 100
52 77.4 69.9 104.9
51 39.5 33.8 50.6





#6
Benzaldehyde
Concen: 48.371 ng/ul m
RT: 6.675 min Scan# 610
Delta R.T. -0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampleId :
SLCS565



Tgt Ion: 77 Resp: 149020
Ion Ratio Lower Upper
77 100
105 98.1 82.2 123.4
106 91.3 75.0 112.4




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#8
Phenol
Concen: 39.426 ng/ul
RT: 6.728 min Scan# 619
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion: 94 Resp: 269981
Ion Ratio Lower Upper
94 100
65 31.0 24.2 36.4
66 44.3 31.4 47.0





#10
Bis(2-Chloroethyl)ether
Concen: 38.799 ng/ul
RT: 6.957 min Scan# 658
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion: 93 Resp: 202361
Ion Ratio Lower Upper
93 100
63 74.8 68.0 102.0
95 31.1 25.8 38.6

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#12
2-Chlorophenol
Concen: 42.160 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion: 128 Resp: 220467
Ion Ratio Lower Upper
128 100
64 44.0 39.2 58.8
130 31.5 25.3 37.9









#13
2-Methylphenol
Concen: 39.010 ng/ul
RT: 7.957 min Scan# 828
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion: 108 Resp: 205968
Ion Ratio Lower Upper
108 100
107 90.1 70.6 105.8

Instrument :
BNA_M
ClientSampleId :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#14
2,2'-oxybis(1-Chloropropane)
Concen: 37.986 ng/ul
RT: 8.034 min Scan# 841
Delta R.T. -0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion: 45 Resp: 257472
Ion Ratio Lower Upper
45 100
77 16.1 10.2 15.4#
79 11.4 7.8 11.8









#16
Acetophenone
Concen: 38.082 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion: 105 Resp: 348238
Ion Ratio Lower Upper
105 100
77 78.0 63.7 95.5
51 28.7 25.4 38.0

Instrument :
BNA_M
ClientSampleId :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#17
N-Nitroso-di-n-propylamine
Concen: 41.408 ng/ul
RT: 8.322 min Scan# 890
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion: 70 Resp: 184728
Ion Ratio Lower Upper
70 100
42 61.0 53.4 80.0
101 10.1 7.4 11.0
130 22.0 16.8 25.2









#18
4-Methylphenol
Concen: 39.163 ng/ul
RT: 8.287 min Scan# 884
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:108 Resp: 229620
Ion Ratio Lower Upper
108 100
107 111.7 87.0 130.6

Instrument :
BNA_M
ClientSampleId :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#19
Hexachloroethane
Concen: 40.662 ng/ul
RT: 8.557 min Scan# 930
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion:117 Resp: 91086
Ion Ratio Lower Upper
117 100
201 109.4 80.1 120.1
199 67.9 50.8 76.2





#22
Nitrobenzene
Concen: 40.257 ng/ul
RT: 8.704 min Scan# 955
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion: 77 Resp: 267354
Ion Ratio Lower Upper
77 100
123 44.6 34.8 52.2
65 13.4 10.4 15.6

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#23
Isophorone
Concen: 40.716 ng/ul
RT: 9.228 min Scan# 1044
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Tgt Ion: 82 Resp: 497666
Ion Ratio Lower Upper
82 100
95 8.6 5.3 7.9#
138 18.4 13.6 20.4





#25
2-Nitrophenol
Concen: 42.761 ng/ul
RT: 9.404 min Scan# 1074
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:139 Resp: 138677
Ion Ratio Lower Upper
139 100
65 50.4 43.7 65.5
109 36.1 24.0 36.0#

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#26
2,4-Dimethylphenol
Concen: 39.999 ng/ul
RT: 9.469 min Scan# 1085
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion:107 Resp: 264053
Ion Ratio Lower Upper
107 100
121 49.8 41.0 61.6
122 85.3 68.2 102.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 40.888 ng/ul
 RT: 9.710 min Scan# 1126
 Delta R.T. 0.000 min
 Lab File: BM031961.D
 Acq: 30 Aug 2021 22:13

Tgt Ion: 93 Resp: 296255
 Ion Ratio Lower Upper
 93 100
 95 31.7 25.9 38.9
 123 11.4 8.5 12.7

Instrument :
 BNA_M
 ClientSampled :
 SLCS565




Manual Integrations
 APPROVED


mohammad
 8/31/2021 11:37:28 AM



#29
 2,4-Dichlorophenol
 Concen: 41.693 ng/ul
 RT: 9.928 min Scan# 1163
 Delta R.T. 0.000 min
 Lab File: BM031961.D
 Acq: 30 Aug 2021 22:13


Tgt Ion: 162 Resp: 247708
 Ion Ratio Lower Upper
 162 100
 164 65.4 51.6 77.4
 98 32.7 28.2 42.4





#30
Naphthalene
Concen: 40.156 ng/ul
RT: 10.316 min Scan# 1229
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:128 Resp: 728070
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.4
127 13.3 10.6 15.8

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#32
4-Chloroaniline
Concen: 37.041 ng/ul
RT: 10.445 min Scan# 1251
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Tgt Ion:127 Resp: 325557
Ion Ratio Lower Upper
127 100
129 31.0 25.6 38.4





#33
Hexachlorobutadiene
Concen: 40.421 ng/ul
RT: 10.592 min Scan# 1276
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:225 Resp: 159316
Ion Ratio Lower Upper
225 100
223 61.1 50.7 76.1
227 66.5 50.7 76.1

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#34
Caprolactam
Concen: 40.330 ng/ul m
RT: 11.245 min Scan# 1387
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Tgt Ion:113 Resp: 79869
Ion Ratio Lower Upper
113 100
55 152.6 155.2 232.8#
56 114.1 101.9 152.9





#35
4-Chloro-3-methylphenol
Concen: 42.626 ng/ul
RT: 11.581 min Scan# 1444
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:107 Resp: 265579
Ion Ratio Lower Upper
107 100
144 27.0 22.3 33.5
142 82.8 65.6 98.4

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#36
2-Methylnaphthalene
Concen: 40.594 ng/ul
RT: 11.939 min Scan# 1505
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion:142 Resp: 580322
Ion Ratio Lower Upper
142 100
141 88.3 70.6 105.8









#37
1-Methylnaphthalene
Concen: 38.334 ng/ul
RT: 12.157 min Scan# 1542
Delta R.T. -0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:142 Resp: 558176
Ion Ratio Lower Upper
142 100
141 91.4 72.8 109.2
116 4.1 3.1 4.7

Instrument :
BNA_M
ClientSampled :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#39
1,2,4,5-Tetrachlorobenzene
Concen: 41.024 ng/ul
RT: 12.310 min Scan# 1568
Delta R.T. -0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:216 Resp: 328776
Ion Ratio Lower Upper
216 100
214 77.3 62.2 93.4
179 21.1 16.6 24.8
108 17.2 13.5 20.3









#40
Hexachlorocyclopentadiene
Concen: 35.917 ng/ul
RT: 12.280 min Scan# 1563
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:237 Resp: 154933
Ion Ratio Lower Upper
237 100
235 62.8 49.0 73.4
272 14.0 11.0 16.4

Instrument :
BNA_M
ClientSampleId :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#41
2,4,6-Trichlorophenol
Concen: 41.651 ng/ul
RT: 12.569 min Scan# 1612
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:196 Resp: 223619
Ion Ratio Lower Upper
196 100
198 101.5 78.8 118.2
200 32.3 26.0 39.0









#42
2,4,5-Trichlorophenol
Concen: 42.313 ng/ul
RT: 12.639 min Scan# 1624
Delta R.T. -0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:196 Resp: 246702
Ion Ratio Lower Upper
196 100
198 94.7 76.1 114.1
200 31.1 25.4 38.2

Instrument :
BNA_M
ClientSampled :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#43
1,1'-Biphenyl
Concen: 39.804 ng/ul
RT: 12.975 min Scan# 1681
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:154 Resp: 781693
Ion Ratio Lower Upper
154 100
153 40.2 32.3 48.5
76 11.0 10.6 15.8









#44
2-Chloronaphthalene
Concen: 40.334 ng/ul
RT: 13.010 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:162 Resp: 627537
Ion Ratio Lower Upper
162 100
164 32.1 26.2 39.4
127 35.0 29.4 44.0

Instrument :
BNA_M
ClientSampled :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#45
2-Nitroaniline
Concen: 42.393 ng/ul
RT: 13.239 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion: 65 Resp: 181713
Ion Ratio Lower Upper
65 100
92 67.2 53.9 80.9
138 107.1 83.3 124.9





#47
Dimethylphthalate
Concen: 40.124 ng/ul
RT: 13.627 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampleId :
SLCS565




Tgt Ion:163 Resp: 832318
Ion Ratio Lower Upper
163 100
194 4.4 3.2 4.8
164 9.8 8.2 12.4

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#48
2,6-Dinitrotoluene
Concen: 43.634 ng/ul
RT: 13.751 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Tgt Ion:165 Resp: 178565
Ion Ratio Lower Upper
165 100
89 57.9 48.6 72.8
121 19.9 18.1 27.1





#50
Acenaphthylene
Concen: 40.347 ng/ul
RT: 13.869 min Scan# 1833
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:152 Resp: 928820
Ion Ratio Lower Upper
152 100
151 19.2 16.2 24.2
153 13.3 10.2 15.4

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#51
3-Nitroaniline
Concen: 43.224 ng/ul
RT: 14.080 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Tgt Ion:138 Resp: 165533
Ion Ratio Lower Upper
138 100
108 9.2 7.4 11.2
92 115.8 96.5 144.7





#52
Acenaphthene
Concen: 40.351 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:153 Resp: 674649
Ion Ratio Lower Upper
153 100
152 46.7 37.4 56.2
154 89.2 69.4 104.0

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#53
2,4-Dinitrophenol
Concen: 39.078 ng/ul
RT: 14.298 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion:184 Resp: 111814
Ion Ratio Lower Upper
184 100
63 65.5 56.9 85.3
154 64.2 49.0 73.6









#55
4-Nitrophenol
Concen: 39.978 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:109 Resp: 141129
Ion Ratio Lower Upper
109 100
139 107.2 100.3 150.5
65 117.1 104.6 156.8

Instrument :
BNA_M
ClientSampleId :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#56
Dibenzofuran
Concen: 40.009 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion:168 Resp: 986506
Ion Ratio Lower Upper
168 100
139 36.9 29.5 44.3





#57
2,4-Dinitrotoluene
Concen: 42.990 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:165 Resp: 263860
Ion Ratio Lower Upper
165 100
63 59.3 37.6 56.4#
182 3.0 1.8 2.8#

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#58
2,3,4,6-Tetrachlorophenol
Concen: 42.592 ng/ul
RT: 14.786 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion:232 Resp: 222926
Ion Ratio Lower Upper
232 100
131 46.2 35.0 52.6
130 2.7 1.7 2.5#
166 33.2 24.4 36.6









#59
Diethylphthalate
Concen: 41.109 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:149 Resp: 871556
Ion Ratio Lower Upper
149 100
177 22.3 17.0 25.4
150 12.5 9.9 14.9

Instrument :
BNA_M
ClientSampleId :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#61
Fluorene
Concen: 40.628 ng/ul
RT: 15.210 min Scan# 2061
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion:166 Resp: 847625
Ion Ratio Lower Upper
166 100
165 96.9 78.4 117.6
167 13.5 10.6 15.8





#62
4-Chlorophenyl-phenylether
Concen: 40.124 ng/ul
RT: 15.210 min Scan# 2061
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:204 Resp: 428726
Ion Ratio Lower Upper
204 100
206 33.0 26.0 39.0
141 58.7 45.6 68.4

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#63
4-Nitroaniline
Concen: 47.945 ng/ul m
RT: 15.257 min Scan# 2069
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion:138 Resp: 164471
Ion Ratio Lower Upper
138 100
92 53.1 41.5 62.3
108 99.4 61.4 92.2#









#66
4,6-Dinitro-2-methylphenol
Concen: 40.295 ng/ul
RT: 15.316 min Scan# 2079
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:198 Resp: 162364
Ion Ratio Lower Upper
198 100
182 3.8 3.8 5.6#
77 24.8 20.1 30.1

Instrument :
BNA_M
ClientSampled :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#67
N-Nitrosodiphenylamine
Concen: 42.758 ng/ul
RT: 15.433 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion:169 Resp: 740543
Ion Ratio Lower Upper
169 100
168 64.8 53.8 80.6
167 34.8 28.6 43.0





#68
4-Bromophenyl-phenylether
Concen: 42.923 ng/ul
RT: 16.104 min Scan# 2213
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:248 Resp: 265278
Ion Ratio Lower Upper
248 100
250 97.4 78.6 118.0
141 76.9 56.2 84.2

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#69
Hexachlorobenzene
Concen: 42.704 ng/ul
RT: 16.210 min Scan# 2231
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion:284 Resp: 306430
Ion Ratio Lower Upper
284 100
142 35.5 29.0 43.4
249 31.6 25.0 37.4









#70
Atrazine
Concen: 43.648 ng/ul
RT: 16.398 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:200 Resp: 284827
Ion Ratio Lower Upper
200 100
173 25.9 18.7 28.1
215 46.6 38.9 58.3

Instrument :
BNA_M
ClientSampled :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#71
Pentachlorophenol
Concen: 41.043 ng/ul
RT: 16.563 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:266 Resp: 192589
Ion Ratio Lower Upper
266 100
264 62.9 48.2 72.4
268 63.1 51.6 77.4









#72
Phenanthrene
Concen: 41.782 ng/u1
RT: 16.951 min Scan# 2357
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:178 Resp: 1376261
Ion Ratio Lower Upper
178 100
179 14.9 12.1 18.1
176 18.7 15.2 22.8

Instrument :
BNA_M
ClientSampleId :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#74
Anthracene
Concen: 42.029 ng/u1
RT: 17.039 min Scan# 2372
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion:178 Resp: 1416844
Ion Ratio Lower Upper
178 100
179 15.5 12.1 18.1
176 18.3 15.2 22.8





#75
1,2,3,4-Tetrachlorobenzene
Concen: 42.192 ng/uL
RT: 12.933 min Scan# 1674
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:216 Resp: 331126
Ion Ratio Lower Upper
216 100
214 77.3 61.9 92.9
218 48.4 39.5 59.3

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#76
Pentachlorobenzene
Concen: 40.928 ng/uL
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Tgt Ion:250 Resp: 336977
Ion Ratio Lower Upper
250 100
248 64.1 50.2 75.4
252 63.9 52.2 78.2





#77
 Carbazole
 Concen: 40.656 ng/ul
 RT: 17.321 min Scan# 2420
 Delta R.T. 0.000 min
 Lab File: BM031961.D
 Acq: 30 Aug 2021 22:13

Instrument :
 BNA_M
 ClientSampleId :
 SLCS565






Tgt Ion:167 Resp: 1254379


Ion	Ratio	Lower	Upper
167	100		
166	20.8	17.0	25.6
139	12.6	9.9	14.9

Manual Integrations
 APPROVED

mohammad
 8/31/2021 11:37:28 AM






#78
 Di-n-butylphthalate
 Concen: 44.452 ng/ul
 RT: 17.892 min Scan# 2517
 Delta R.T. 0.000 min
 Lab File: BM031961.D
 Acq: 30 Aug 2021 22:13



Tgt Ion:149 Resp: 1571574

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	5.7	4.1	6.1









#80
Fluoranthene
Concen: 44.572 ng/u1
RT: 18.974 min Scan# 2701
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:202 Resp: 1669483
Ion Ratio Lower Upper
202 100
101 9.0 8.8 13.2
100 7.1 6.8 10.2

Instrument :
BNA_M
ClientSampleId :
SLCS565




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM




#82
Pyrene
Concen: 43.693 ng/u1
RT: 19.339 min Scan# 2763
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Tgt Ion:202 Resp: 1732675
Ion Ratio Lower Upper
202 100
101 10.7 10.0 15.0
100 8.7 8.5 12.7






#83
Butylbenzylphthalate
Concen: 47.870 ng/ul
RT: 20.262 min Scan# 2920
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13





Tgt Ion:149 Resp: 745396
Ion Ratio Lower Upper
149 100
91 71.5 57.4 86.2
206 20.2 15.9 23.9




Instrument :
BNA_M
ClientSampled :
SLCS565

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#84
3,3'-Dichlorobenzidine
Concen: 42.771 ng/ul
RT: 21.045 min Scan# 3053
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Tgt Ion:252 Resp: 633672
Ion Ratio Lower Upper
252 100
254 65.9 49.9 74.9
126 9.4 8.9 13.3





#85
Benzo(a)anthracene
Concen: 42.125 ng/ul
RT: 21.098 min Scan# 3062
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13

Instrument :
BNA_M
ClientSampled :
SLCS565






Tgt Ion:228 Resp: 1724427


Ion	Ratio	Lower	Upper
228	100		
229	19.9	16.0	24.0
226	27.0	22.2	33.2

Manual Integrations
APPROVED

mohammad
8/31/2021 11:37:28 AM






#86
Bis(2-ethylhexyl)phthalate
Concen: 47.803 ng/ul
RT: 21.045 min Scan# 3053
Delta R.T. 0.000 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion:149 Resp: 1200444


Ion	Ratio	Lower	Upper
149	100		
167	27.5	20.7	31.1
279	4.7	3.0	4.4#





#87
Chrysene
Concen: 42.328 ng/ul
RT: 21.150 min Scan# 3071
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565



Tgt Ion:228 Resp: 1666649
Ion Ratio Lower Upper
228 100
226 30.0 23.6 35.4
229 20.0 15.9 23.9




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#89
Di-n-octyl phthalate
Concen: 47.459 ng/ul
RT: 21.903 min Scan# 3199
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion:149 Resp: 2046458





#90
Benzo(b)fluoranthene
Concen: 43.488 ng/ul
RT: 22.621 min Scan# 3321
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampleId :
SLCS565



Tgt Ion:252 Resp: 1751243
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 7.9 7.7 11.5




Manual Integrations
APPROVED


mohammad
8/31/2021 11:37:28 AM



#91
Benzo(k)fluoranthene
Concen: 45.086 ng/ul
RT: 22.662 min Scan# 3328
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13


Tgt Ion:252 Resp: 1757664
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 8.2 7.9 11.9





#93
Benzo(a)pyrene
Concen: 41.751 ng/ul
RT: 23.162 min Scan# 3413
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:252 Resp: 1518233
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 8.8 8.6 13.0

Manual Integrations
APPROVED



mohammad
8/31/2021 11:37:28 AM




#94
Indeno(1,2,3-cd)pyrene
Concen: 36.017 ng/ul
RT: 25.368 min Scan# 3788
Delta R.T. 0.012 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Tgt Ion:276 Resp: 1676246
Ion Ratio Lower Upper
276 100
138 20.8 19.5 29.3
277 25.1 19.6 29.4





#95
Dibenzo(a,h)anthracene
Concen: 36.925 ng/ul
RT: 25.374 min Scan# 3789
Delta R.T. 0.012 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13




Instrument :
BNA_M
ClientSampled :
SLCS565




Tgt Ion:278 Resp: 1445291
Ion Ratio Lower Upper
278 100
139 12.7 12.8 19.2#
279 23.2 19.1 28.7

Manual Integrations
APPROVED

mohammad
8/31/2021 11:37:28 AM



#96
Benzo(g,h,i)perylene
Concen: 34.262 ng/ul
RT: 26.003 min Scan# 3896
Delta R.T. 0.006 min
Lab File: BM031961.D
Acq: 30 Aug 2021 22:13



Tgt Ion:276 Resp: 1320431
Ion Ratio Lower Upper
276 100
138 17.0 16.5 24.7
277 23.7 18.5 27.7

