

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM083115\
 Data File : BM002587.D
 Acq On : 31 Aug 2015 11:32
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02065

Quant Time: Sep 01 01:37:18 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 31 01:34:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	188#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	157#	0.00
3	Naphthalene	1.021	0.960	6.0	151#	0.00
4	2-Methylnaphthalene	0.765	0.648	15.3	136	0.00
5	1-Methylnaphthalene	0.756	0.627	17.1	133	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
7 S	Acenaphthylene-d8	2.045	1.941	5.1	113	0.00
8	Acenaphthylene	2.140	2.028	5.2	113	0.00
9 C	Acenaphthene	1.410	1.329	5.7	113	0.00
10 S	Fluorene-d10	1.436	1.220	15.0	102	0.00
11	Fluorene	1.615	1.391	13.9	103	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
13	Phenanthrene	1.155	1.083	6.2	85	0.00
14 S	Anthracene-d10	0.965	0.901	6.6	84	0.00
15	Anthracene	1.170	1.093	6.6	84	0.00
16 C	Fluoranthene	1.308	1.064	18.7	74	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
18 S	Pyrene-d10	0.887	0.895	-0.9	74	0.00
19	Pyrene	1.172	1.189	-1.5	74	0.00
20	Benzo(a)anthracene	1.196	1.121	6.3	74	0.00
21	Chrysene	1.119	1.052	6.0	73	0.00
22 I	Perylene-d12	1.000	1.000	0.0	80	0.00
23	Benzo(b)fluoranthene	1.209	1.102	8.9	76	0.00
24	Benzo(k)fluoranthene	1.146	1.076	6.1	75	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.989	5.9	77	0.00
26 C	Benzo(a)pyrene	1.155	1.088	5.8	76	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.305	1.7	80	0.01
28	Dibenzo(a,h)anthracene	1.110	1.104	0.5	81	0.01
29	Benzo(g,h,i)perylene	1.106	1.101	0.5	81	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0