

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM083115\
 Data File : BM002596.D
 Acq On : 31 Aug 2015 19:23
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02066

Quant Time: Sep 01 01:42:25 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 01 01:38:52 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	181#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	155#	0.00
3	Naphthalene	1.021	0.961	5.9	150	0.00
4	2-Methylnaphthalene	0.765	0.660	13.7	137	0.00
5	1-Methylnaphthalene	0.756	0.642	15.1	134	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
7 S	Acenaphthylene-d8	2.045	1.938	5.2	117	0.00
8	Acenaphthylene	2.140	2.032	5.0	118	0.00
9 C	Acenaphthene	1.410	1.323	6.2	117	0.00
10 S	Fluorene-d10	1.436	1.235	14.0	107	0.00
11	Fluorene	1.615	1.385	14.2	106	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
13	Phenanthrene	1.155	1.086	6.0	88	0.00
14 S	Anthracene-d10	0.965	0.894	7.4	86	0.00
15	Anthracene	1.170	1.090	6.8	86	0.00
16 C	Fluoranthene	1.308	1.027	21.5	74	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
18 S	Pyrene-d10	0.887	0.908	-2.4	74	0.00
19	Pyrene	1.172	1.200	-2.4	73	0.00
20	Benzo(a)anthracene	1.196	1.120	6.4	73	0.00
21	Chrysene	1.119	1.052	6.0	72	0.00
22 I	Perylene-d12	1.000	1.000	0.0	80	0.00
23	Benzo(b)fluoranthene	1.209	1.130	6.5	77	0.00
24	Benzo(k)fluoranthene	1.146	1.053	8.1	73	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.979	6.9	75	0.00
26 C	Benzo(a)pyrene	1.155	1.092	5.5	76	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.303	1.9	79	0.01
28	Dibenzo(a,h)anthracene	1.110	1.105	0.5	80	0.00
29	Benzo(g,h,i)perylene	1.106	1.097	0.8	80	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0