

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM083115\
 Data File : BM002614.D
 Acq On : 01 Sep 2015 06:21
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02067

Quant Time: Sep 01 07:04:21 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 01 02:00:58 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	180#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	165#	0.00
3	Naphthalene	1.021	0.953	6.7	158#	0.00
4	2-Methylnaphthalene	0.765	0.674	11.9	149	0.00
5	1-Methylnaphthalene	0.756	0.656	13.2	146	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
7 S	Acenaphthylene-d8	2.045	1.943	5.0	136	0.00
8	Acenaphthylene	2.140	2.027	5.3	136	0.00
9 C	Acenaphthene	1.410	1.334	5.4	137	0.00
10 S	Fluorene-d10	1.436	1.309	8.8	132	0.00
11	Fluorene	1.615	1.489	7.8	133	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
13	Phenanthrene	1.155	1.093	5.4	93	0.00
14 S	Anthracene-d10	0.965	0.907	6.0	92	0.00
15	Anthracene	1.170	1.104	5.6	92	0.00
16 C	Fluoranthene	1.308	1.012	22.6	77	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	68	0.00
18 S	Pyrene-d10	0.887	1.034	-16.6	74	0.00
19	Pyrene	1.172	1.378	-17.6	74	0.00
20	Benzo(a)anthracene	1.196	1.137	4.9	65	0.00
21	Chrysene	1.119	1.052	6.0	64	0.00
22 I	Perylene-d12	1.000	1.000	0.0	66	0.00
23	Benzo(b)fluoranthene	1.209	1.111	8.1	62	0.00
24	Benzo(k)fluoranthene	1.146	1.095	4.5	63	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.989	5.9	63	0.00
26 C	Benzo(a)pyrene	1.155	1.088	5.8	62	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.307	1.6	66	0.00
28	Dibenzo(a,h)anthracene	1.110	1.097	1.2	66	0.00
29	Benzo(g,h,i)perylene	1.106	1.093	1.2	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0