

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM083115\
 Data File : BM002621.D
 Acq On : 01 Sep 2015 10:36
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02068

Quant Time: Sep 01 11:16:57 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 01 07:11:38 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	138	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
3	Naphthalene	1.021	0.953	6.7	117	0.00
4	2-Methylnaphthalene	0.765	0.679	11.2	111	0.00
5	1-Methylnaphthalene	0.756	0.663	12.3	110	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
7 S	Acenaphthylene-d8	2.045	1.935	5.4	101	0.00
8	Acenaphthylene	2.140	2.040	4.7	102	0.00
9 C	Acenaphthene	1.410	1.338	5.1	102	0.00
10 S	Fluorene-d10	1.436	1.313	8.6	98	0.00
11	Fluorene	1.615	1.491	7.7	99	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
13	Phenanthrene	1.155	1.086	6.0	88	0.00
14 S	Anthracene-d10	0.965	0.903	6.4	87	0.00
15	Anthracene	1.170	1.100	6.0	87	0.00
16 C	Fluoranthene	1.308	1.004	23.2	72	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	65	0.00
18 S	Pyrene-d10	0.887	1.007	-13.5	70	0.00
19	Pyrene	1.172	1.336	-14.0	69	0.00
20	Benzo(a)anthracene	1.196	1.118	6.5	62	0.00
21	Chrysene	1.119	1.048	6.3	61	0.00
22 I	Perylene-d12	1.000	1.000	0.0	64	0.00
23	Benzo(b)fluoranthene	1.209	1.121	7.3	61	0.00
24	Benzo(k)fluoranthene	1.146	1.064	7.2	59	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.980	6.8	61	0.00
26 C	Benzo(a)pyrene	1.155	1.084	6.1	61	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.292	2.7	63	0.00
28	Dibenzo(a,h)anthracene	1.110	1.084	2.3	63	0.00
29	Benzo(g,h,i)perylene	1.106	1.087	1.7	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0