

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM083115\
 Data File : BM002637.D
 Acq On : 01 Sep 2015 20:23
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02069

Quant Time: Sep 02 00:46:23 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 02 00:45:29 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	154#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	137	0.00
3	Naphthalene	1.021	0.954	6.6	131	0.00
4	2-Methylnaphthalene	0.765	0.686	10.3	126	0.00
5	1-Methylnaphthalene	0.756	0.676	10.6	125	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
7 S	Acenaphthylene-d8	2.045	1.925	5.9	127	0.00
8	Acenaphthylene	2.140	2.012	6.0	127	0.00
9 C	Acenaphthene	1.410	1.344	4.7	129	0.00
10 S	Fluorene-d10	1.436	1.211	15.7	114	0.00
11	Fluorene	1.615	1.361	15.7	114	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
13	Phenanthrene	1.155	1.088	5.8	107	0.00
14 S	Anthracene-d10	0.965	0.901	6.6	106	0.00
15	Anthracene	1.170	1.093	6.6	106	0.00
16 C	Fluoranthene	1.308	0.983	24.8	87	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
18 S	Pyrene-d10	0.887	0.969	-9.2	83	0.00
19	Pyrene	1.172	1.288	-9.9	83	0.00
20	Benzo(a)anthracene	1.196	1.127	5.8	77	0.00
21	Chrysene	1.119	1.053	5.9	76	0.00
22 I	Perylene-d12	1.000	1.000	0.0	82	0.00
23	Benzo(b)fluoranthene	1.209	1.123	7.1	79	0.00
24	Benzo(k)fluoranthene	1.146	1.068	6.8	76	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.989	5.9	78	0.00
26 C	Benzo(a)pyrene	1.155	1.087	5.9	78	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.304	1.8	82	0.00
28	Dibenzo(a,h)anthracene	1.110	1.104	0.5	82	0.00
29	Benzo(g,h,i)perylene	1.106	1.098	0.7	83	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0