


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM083115\
 Data File : BM002586.D
 Acq On : 31 Aug 2015 10:56
 Operator : UM/IZ
 Sample : DFTPP09
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP09

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Sep 01 07:11:38 2015



AutoFind: Scans 2607, 2608, 2609; Background Corrected with Scan 2602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.2	25016	PASS
68	69	0.00	2	1.7	444	PASS
69	198	0.00	100	35.5	25930	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.8	37157	PASS
197	198	0.00	2	0.4	297	PASS
198	198	100	100	100.0	73117	PASS
199	198	5	9	6.6	4846	PASS
275	198	10	60	25.9	18950	PASS
365	198	1	100	2.8	2058	PASS
441	443	0.01	100	78.1	10218	PASS
442	198	50	100	92.1	67360	PASS
443	442	15	24	19.4	13078	PASS

Scan 2602 (18.392 min): BM002586.D
DFTPP09

m/z	Abundance
39.90	523.0
43.00	315.0
44.00	2926.0
55.00	357.0
73.00	339.0
207.00	940.0
280.90	360.0
326.10	323.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
38.00	320.0
39.00	2026.0
40.00	640.0
41.00	310.0
43.10	401.0
44.00	3195.0
50.00	6088.0
51.10	24544.0
52.00	1398.0
55.10	483.0
56.00	783.0
57.00	1784.0
61.00	312.0
62.00	353.0
63.00	993.0
65.00	554.0
68.00	436.0
69.00	25432.0
73.10	535.0
74.00	2531.0
75.00	3920.0
76.00	1323.0
77.00	28720.0
78.10	2247.0
79.00	1862.0
79.90	1379.0
81.00	2509.0
82.00	527.0
83.00	708.0
85.00	447.0
86.00	571.0
87.00	333.0
91.00	559.0
91.90	433.0
93.00	2996.0
94.00	307.0
95.00	325.0
98.00	2405.0
99.00	1780.0
101.00	1178.0
102.90	402.0
103.90	699.0
105.00	765.0
107.00	8651.0
107.90	1448.0
109.10	371.0
110.00	17008.0
111.00	2535.0
111.90	338.0
116.00	531.0
117.00	5947.0
118.00	502.0
121.90	599.0
122.90	981.0
123.90	389.0
125.00	422.0
127.00	33752.0
128.00	2667.0
129.00	12063.0
130.00	998.0
133.90	394.0
134.90	1080.0
135.90	414.0
137.10	594.0
141.00	1598.0
142.00	661.0
142.90	402.0
146.00	322.0
146.90	854.0
147.90	1577.0
149.00	598.0
152.90	522.0
153.90	337.0
154.90	894.0
156.00	1284.0
157.90	312.0
160.00	474.0
161.00	896.0
161.90	386.0
162.90	313.0
165.00	692.0
165.90	516.0
167.00	2923.0
167.90	1361.0
169.00	377.0
171.90	312.0
173.00	351.0
174.00	708.0
175.00	1300.0
176.10	392.0
176.90	540.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

179.00	2169.0
180.00	1421.0
181.00	798.0
184.90	1090.0
186.00	8391.0
187.00	2581.0
187.90	308.0
189.00	641.0
191.00	451.0
192.00	718.0
193.00	730.0
196.00	2245.0
196.80	410.0
197.90	62096.0
199.00	4175.0
200.00	429.0
201.30	391.0
203.00	481.0
203.90	1862.0
205.00	3888.0
206.00	15949.0
207.00	3023.0
207.90	682.0
209.00	349.0
211.10	645.0
217.00	3976.0
217.90	613.0
221.00	2817.0
223.00	937.0
224.00	7988.0
225.00	2047.0
227.00	3329.0
227.90	541.0
229.00	626.0
231.00	429.0
236.90	310.0
242.00	503.0
243.10	521.0
244.00	6648.0
245.00	939.0
246.00	1179.0
248.90	317.0
254.10	322.0
255.00	32344.0
256.00	5189.0
256.90	405.0
258.00	1634.0
258.90	306.0
264.90	752.0
273.00	929.0
274.00	2545.0
275.00	14834.0
276.00	2057.0
277.00	1249.0
296.00	3624.0
296.90	501.0
303.00	404.0
315.00	406.0
323.00	1198.0
324.10	432.0
326.20	335.0
326.90	374.0
333.90	716.0
346.10	324.0
352.00	534.0
354.00	417.0
364.90	1504.0
371.90	650.0
403.10	421.0
423.00	2230.0
424.00	531.0
441.00	6720.0
442.00	43200.0
443.00	8326.0
444.00	859.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
38.00	360.0
39.00	2473.0
40.00	617.0
41.00	405.0
43.00	307.0
44.00	3215.0
50.00	7228.0
51.00	29816.0
52.00	1437.0
55.00	464.0
56.00	1037.0
57.00	2246.0
61.00	424.0
62.00	399.0
63.00	1153.0
65.00	629.0
68.10	571.0
69.00	31256.0
73.00	514.0
74.00	2980.0
75.00	5053.0
76.10	1582.0
77.10	35768.0
78.00	2572.0
79.00	2294.0
80.00	1551.0
81.00	2528.0
82.00	716.0
83.00	731.0
84.90	625.0
86.00	676.0
87.00	354.0
91.00	605.0
92.00	634.0
93.00	3765.0
93.90	312.0
95.00	312.0
98.00	2681.0
99.00	2160.0
101.00	1508.0
102.90	607.0
104.00	1017.0
105.00	990.0
106.00	382.0
107.00	11133.0
108.00	1958.0
109.10	558.0
110.00	22152.0
111.00	3405.0
112.00	482.0
116.00	690.0
117.00	7324.0
117.90	600.0
122.00	752.0
123.00	1210.0
124.00	520.0
125.00	445.0
126.20	308.0
127.00	44688.0
128.00	3353.0
129.00	15376.0
129.90	1229.0
133.90	518.0
135.00	1512.0
136.00	693.0
137.00	684.0
141.00	1739.0
142.00	695.0
142.90	561.0
145.90	319.0
147.00	1179.0
148.00	2079.0
149.00	587.0
151.20	389.0
152.90	709.0
153.90	442.0
155.00	1081.0
156.00	1757.0
157.10	385.0
158.00	431.0
158.90	395.0
160.00	610.0
161.10	1058.0
161.90	506.0
164.90	751.0
166.00	580.0
167.00	3830.0
168.00	1752.0
168.90	365.0
171.90	430.0
173.00	488.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

174.00	999.0
175.00	1786.0
176.00	514.0
176.90	743.0
178.00	364.0
179.00	3102.0
180.00	2131.0
181.00	1057.0
185.00	1337.0
186.00	11612.0
187.00	3370.0
188.00	401.0
188.90	756.0
191.00	578.0
192.00	923.0
193.00	1138.0
196.00	2812.0
197.10	621.0
197.90	87824.0
199.00	5808.0
199.90	367.0
201.30	514.0
202.90	612.0
204.00	2982.0
205.00	4928.0
206.00	22192.0
207.00	3648.0
207.90	855.0
208.90	304.0
211.00	891.0
215.00	307.0
215.90	537.0
217.00	5089.0
218.00	792.0
221.00	4092.0
223.00	1081.0
224.00	11793.0
225.00	3212.0
226.00	355.0
227.00	4323.0
228.00	606.0
228.90	991.0
231.00	375.0
233.90	396.0
234.90	386.0
236.90	355.0
242.00	623.0
243.00	564.0
244.00	9717.0
245.00	1226.0
246.00	1832.0
249.00	319.0
254.10	382.0
255.00	49096.0
256.00	6980.0
257.00	539.0
258.00	2413.0
259.00	453.0
265.00	1041.0
272.90	1455.0
274.00	3535.0
275.00	22576.0
276.00	3298.0
277.00	1591.0
284.90	383.0
292.90	457.0
296.00	5142.0
297.00	837.0
303.00	634.0
314.00	342.0
314.90	694.0
316.00	343.0
323.00	1998.0
324.00	498.0
326.90	480.0
334.00	997.0
335.00	424.0
341.00	321.0
345.90	432.0
352.00	767.0
352.90	522.0
354.00	728.0
364.90	2405.0
366.00	468.0
372.00	1230.0
383.00	395.0
402.00	405.0
403.00	699.0
421.00	477.0
422.00	601.0
423.00	3928.0
424.00	901.0
441.00	11899.0
442.00	78000.0
443.00	15312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

444.00

1480.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
39.00	1590.0
39.90	630.0
41.00	348.0
43.00	319.0
44.00	3035.0
50.00	4786.0
51.00	20688.0
52.00	1028.0
55.00	540.0
56.00	687.0
57.00	1594.0
63.00	858.0
65.00	494.0
67.00	311.0
68.10	327.0
69.00	21104.0
73.00	494.0
74.00	1856.0
75.00	3563.0
76.00	1089.0
77.00	25840.0
78.00	1748.0
79.00	1781.0
80.00	1301.0
81.00	1823.0
82.00	547.0
83.00	623.0
85.00	417.0
85.90	519.0
91.00	626.0
92.00	514.0
93.00	3055.0
95.10	355.0
98.00	1830.0
99.00	1828.0
101.00	1061.0
103.00	400.0
103.90	694.0
105.00	720.0
107.00	7943.0
108.00	1259.0
109.10	475.0
110.00	15982.0
111.00	2235.0
116.00	489.0
117.00	5566.0
117.90	376.0
121.90	731.0
123.00	933.0
124.00	395.0
124.90	350.0
127.00	33032.0
128.00	2467.0
129.00	12018.0
130.00	916.0
133.90	392.0
135.00	877.0
136.00	431.0
137.00	520.0
141.00	1539.0
142.00	571.0
142.90	427.0
146.00	336.0
147.00	976.0
147.90	1510.0
148.90	448.0
151.10	320.0
152.90	521.0
154.00	366.0
155.00	853.0
156.00	1230.0
156.90	357.0
158.00	355.0
160.00	457.0
160.90	915.0
162.00	316.0
164.90	782.0
166.00	545.0
167.00	3199.0
168.00	1424.0
169.00	337.0
173.00	439.0
174.00	690.0
175.00	1328.0
176.00	444.0
176.90	679.0
178.90	2242.0
180.00	1648.0
181.00	838.0
185.00	1349.0
186.00	8964.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

187.00	2516.0
187.90	376.0
188.90	663.0
191.00	436.0
192.00	728.0
193.00	958.0
196.00	2130.0
196.80	482.0
197.90	69432.0
198.90	4556.0
201.40	349.0
203.00	528.0
204.00	2394.0
205.00	4094.0
206.00	17712.0
207.00	3015.0
208.00	838.0
210.10	312.0
210.80	676.0
216.00	331.0
217.00	4424.0
218.00	616.0
220.90	3115.0
221.70	692.0
223.00	986.0
224.00	9969.0
225.00	2307.0
227.00	3626.0
228.00	495.0
228.90	833.0
233.90	333.0
237.00	425.0
241.90	458.0
243.00	604.0
244.00	7772.0
245.10	888.0
246.00	1446.0
248.90	368.0
254.10	398.0
255.00	39208.0
256.00	5620.0
257.00	419.0
257.90	1971.0
264.90	689.0
273.00	1075.0
274.00	3445.0
275.00	19440.0
276.00	2378.0
276.90	1430.0
277.90	327.0
282.90	314.0
292.90	508.0
296.00	4640.0
296.90	607.0
303.00	659.0
315.00	611.0
315.90	347.0
323.00	1829.0
324.00	369.0
326.10	311.0
327.00	398.0
334.00	1184.0
345.90	415.0
352.00	521.0
353.00	479.0
354.00	778.0
365.00	2265.0
372.00	1086.0
373.00	432.0
383.00	317.0
402.00	499.0
403.00	659.0
404.00	319.0
421.00	599.0
421.90	521.0
423.00	4113.0
424.00	762.0
441.00	12036.0
442.00	80880.0
443.00	15596.0
444.00	1145.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09