

Data Path : Z:\HPCHEM1\BNA M\DATA\BM090116\
 Data File : BM007360.D
 Acq On : 01 Sep 2016 15:48
 Operator : UM/SJ
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC3.2

Manual Integrations
 APPROVED

sohil
 9/2/2016 4:37:25 PM

Quant Time: Sep 02 12:34:53 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM090116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 05:02:21 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	2737	5.00	ng	-0.02
7) Naphthalene-d8	10.58	136	13281	5.00	ng	0.00
13) Acenaphthene-d10	14.42	164	9465	5.00	ng	0.00
19) Phenanthrene-d10	17.17	188	23260	5.00	ng	0.00
26) Chrysene-d12	21.35	240	26543	5.00	ng	0.00
35) Perylene-d12	23.61	264	23515	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.39	112	1634	2.58	ng	0.00
5) Phenol-d6	6.97	99	2200	2.84	ng	0.00
8) Nitrobenzene-d5	8.95	82	2118	3.07	ng	-0.01
14) 2,4,6-Tribromophenol	15.90	330	1092m	2.94	ng	-0.02
15) 2-Fluorobiphenyl	13.05	172	8747	3.05	ng	0.00
29) Terphenyl-d14	19.79	244	11571	2.53	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.35	88	580	2.02	ng	# 81
3) n-Nitrosodimethylamine	3.65	42	868	3.08	ng	# 91
6) bis(2-Chloroethyl)ether	7.23	93	1848	2.70	ng	98
9) Nitrobenzene	8.99	77	2323	2.99	ng	97
10) Naphthalene	10.61	128	8893	2.95	ng	# 86
11) Hexachlorobutadiene	10.91	225	1580m	3.35	ng	
12) 2-Methylnaphthalene	12.24	142	6607	3.26	ng	96
16) Acenaphthylene	14.13	152	13082	3.10	ng	100
17) Acenaphthene	14.48	154	8206	2.96	ng	100
18) Fluorene	15.46	166	9927	3.02	ng	# 95
20) 4-Bromophenyl-phenylether	16.35	248	3157	2.52	ng	# 86
21) Hexachlorobenzene	16.48	284	3786	3.04	ng	# 94
22) Pentachlorophenol	16.83	266	1354	3.84	ng	# 86
23) Phenanthrene	17.19	178	18574	2.45	ng	99
24) Anthracene	17.29	178	18742	3.29	ng	98
25) Fluoranthene	19.23	202	21769	2.98	ng	99
27) Benzidine	19.41	184	6582m	4.60	ng	
28) Pyrene	19.59	202	23972	2.65	ng	100
30) Benzo(a)anthracene	21.33	228	23377	2.88	ng	97
31) 3,3'-Dichlorobenzidine	21.27	252	8499	3.79	ng	# 90
32) Chrysene	21.37	228	23135m	2.83	ng	
33) Bis(2-ethylhexyl)phthalate	21.25	149	15520	3.06	ng	# 97
34) Indeno(1,2,3-cd)pyrene	25.89	276	19896	3.38	ng	100
36) Benzo(b)fluoranthene	22.93	252	23208	3.06	ng	# 92
37) Benzo(k)fluoranthene	22.97	252	20066	2.88	ng	# 92
38) Benzo(a)pyrene	23.52	252	19894	3.06	ng	# 91
39) Dibenzo(a,h)anthracene	25.91	278	16030	3.15	ng	# 86
40) Benzo(g,h,i)perylene	26.60	276	16179	3.03	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA M\DATA\BM090116\
Data File : BM007360.D
Acq On : 01 Sep 2016 15:48
Operator : UM/SJ
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM

Quant Time: Sep 02 12:34:53 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM090116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 05:02:21 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.79 min Scan# 320

Delta R.T. -0.02 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Instrument :

BNA_M





ClientSampleId :

SSTDICC3.2

Tot Ion:	152	Resp:	2737
Ion Ratio	Lower	Upper	
152	100		
150	172.6	109.8	164.8#
115	73.0	39.0	58.4#

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM



#2

1,4-Dioxane

Concen: 2.02 ng




RT: 3.35 min Scan# 44


Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Tgt Ion:	88	Resp:	580
Ion Ratio	Lower	Upper	
88	100		
58	68.1	41.0	61.4#
43	35.5	24.2	36.2





#3





n-Nitrosodimethylamine
 Concen: 3.08 ng
 RT: 3.65 min Scan# 63
 Delta R.T. -0.02 min
 Lab File: BM007360.D
 Acq: 01 Sep 2016 15:48

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
42	100		
74	96.2	78.9	118.3
44	5.8	19.1	28.7#

Manual Integrations
 APPROVED




sohil
 9/2/2016 4:37:25 PM






#4

2-Fluorophenol
 Concen: 2.58 ng
 RT: 5.39 min Scan# 176
 Delta R.T. 0.00 min
 Lab File: BM007360.D
 Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
112	100		
64	48.2	36.9	55.3
63	30.2	22.6	34.0





#5

Phenol-d6

Concen: 2.84 ng

RT: 6.97 min Scan# 272

Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Tot Ion:	99	Resp:	2200
Ion	Ratio	Lower	Upper
99	100		
42	23.0	19.8	29.8
71	36.3	28.1	42.1

Instrument :





BNA_M

ClientSampled :

SSTDICC3.2

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM



#6

bis(2-Chloroethyl)ether

Concen: 2.70 ng


RT: 7.23 min Scan# 287


Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Tgt Ion:	93	Resp:	1848
Ion	Ratio	Lower	Upper
93	100		
63	69.7	54.3	81.5
95	32.4	25.7	38.5





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.58 min Scan# 558

Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Instrument :

BNA_M





ClientSampleId :

SSTDICC3.2

Tot Ion	Ratio	Resp	Lower	Upper
136	100	13281		
137	11.7		9.4	14.0
54	4.3		4.0	6.0
68	3.0		2.7	4.1

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM



#8

Nitrobenzene-d5

Concen: 3.07 ng




RT: 8.95 min Scan# 423


Delta R.T. -0.01 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	2118		
128	51.5		54.9	82.3#
54	50.2		53.8	80.6#





#9

Nitrobenzene

Concen: 2.99 ng

RT: 8.99 min Scan# 427

Delta R.T. -0.01 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Instrument :

BNA_M





ClientSampled :

SSTDICC3.2

Tot Ion:	77	Resp:	2323
Ion	Ratio	Lower	Upper
77	100		
123	51.9	40.0	60.0
65	14.4	10.6	15.8

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM



#10

Naphthalene

Concen: 2.95 ng




RT: 10.61 min Scan# 560


Delta R.T. -0.02 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Tgt Ion:	128	Resp:	8893
Ion	Ratio	Lower	Upper
128	100		
129	11.1	15.5	23.3#
127	15.3	15.5	23.3#









#11
Hexachlorobutadiene
Concen: 3.35 ng m
RT: 10.91 min Scan# 577
Delta R.T. -0.02 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampled :
SSTDICC3.2

Tot Ion: 225 Resp: 1580
Ion Ratio Lower Upper
225 100
223 83.5 197.7 296.5#
227 80.7 196.2 294.4#




Manual Integrations
APPROVED


sohil
9/2/2016 4:37:25 PM



#12
2-Methylnaphthalene
Concen: 3.26 ng
RT: 12.24 min Scan# 701
Delta R.T. -0.01 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion: 142 Resp: 6607
Ion Ratio Lower Upper
142 100
141 90.2 70.2 105.4
115 41.0 35.7 53.5









#13
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.42 min Scan# 864
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
164	100		
162	106.6	86.2	129.4
160	49.4	40.5	60.7




Manual Integrations
APPROVED


sohil
9/2/2016 4:37:25 PM



#14
2,4,6-Tribromophenol
Concen: 2.94 ng m
RT: 15.90 min Scan# 931
Delta R.T. -0.02 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
330	100		
332	0.0	79.2	118.8#
141	59.6	0.0	0.0#









#15
2-Fluorobiphenyl
Concen: 3.05 ng
RT: 13.05 min Scan# 779
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
172	100		
171	33.1	32.2	48.2
170	20.9	23.5	35.3




Manual Integrations
APPROVED


sohil
9/2/2016 4:37:25 PM



#16
Acenaphthylene
Concen: 3.10 ng
RT: 14.13 min Scan# 851
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.7	23.5
153	12.7	10.1	15.1









#17
Acenaphthene
Concen: 2.96 ng
RT: 14.48 min Scan# 867
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
154	100		
153	106.5	85.1	127.7
152	50.7	40.9	61.3




Manual Integrations
APPROVED


sohil
9/2/2016 4:37:25 PM



#18
Fluorene
Concen: 3.02 ng
RT: 15.46 min Scan# 911
Delta R.T. -0.02 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.0	78.8	118.2
167	0.0	10.8	16.2#





#19

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.17 min Scan# 981

Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Instrument :

BNA_M


ClientSampleId :

SSTDICC3.2

Tot Ion:188	Resp:	23260
Ion Ratio	Lower	Upper
188 100		
94 2.1	2.0	3.0
80 1.7	1.6	2.4

Manual Integrations
APPROVED


sohil
9/2/2016 4:37:25 PM





Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM0

Ion 80.00 (79.70 to 80.70): BM0



Time-->



#20

4-Bromophenyl-phenylether

Concen: 2.52 ng


RT: 16.35 min Scan# 949

Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48


Tgt Ion:248	Resp:	3157
Ion Ratio	Lower	Upper
248 100		
250 83.7	0.0	0.0#
141 107.1	98.2	147.4




Abundance Ion 248.00 (247.70 to 248.70): E


Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



Time-->









#21
Hexachlorobenzene
Concen: 3.04 ng
RT: 16.48 min Scan# 954
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
284	100		
142	0.0	0.0	0.0
249	30.0	26.6	40.0




Manual Integrations
APPROVED




sohil
9/2/2016 4:37:25 PM



#22
Pentachlorophenol
Concen: 3.84 ng
RT: 16.83 min Scan# 968
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
266	100		
268	77.8	66.4	99.6
264	56.8	61.6	92.4





#23

Phenanthrene

Concen: 2.45 ng

RT: 17.19 min Scan# 982

Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Tot Ion:178	Resp:	18574
Ion Ratio	Lower	Upper
178	100	
176	20.4	16.6 24.8
179	14.5	12.5 18.7

Instrument :





BNA_M

ClientSampleId :

SSTDICC3.2

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM



#24

Anthracene

Concen: 3.29 ng


RT: 17.29 min Scan# 986


Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Tgt Ion:178	Resp:	18742
Ion Ratio	Lower	Upper
178	100	
176	17.7	14.8 22.2
179	15.7	12.2 18.2






#25
Fluoranthene
Concen: 2.98 ng
RT: 19.23 min Scan# 1085
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampleId :
SSTDICC3.2




Tot Ion	Ratio	Lower	Upper
202	100		
101	8.3	6.5	9.7
203	17.3	14.2	21.2

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM






Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E




#26
Chrysene-d12
Concen: 5.00 ng
RT: 21.35 min Scan# 1201
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
240	100		
120	1.4	1.3	1.9
236	23.8	19.1	28.7

Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#27

Benzidine

Concen: 4.60 ng/mL

RT: 19.41 min Scan# 1096

Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Instrument :

BNA_M


ClientSampled :

SSTDICC3.2




Tot Ion:	184	Resp:	6582
Ion Ratio	Lower	Upper	
184	100		
185	16.8	29.8	44.8#
183	11.1	29.8	44.8#

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 185.00 (184.70 to 185.70): E
Ion 183.00 (182.70 to 183.70): E



#28

Pyrene

Concen: 2.65 ng


RT: 19.59 min Scan# 1106

Delta R.T. 0.00 min



Lab File: BM007360.D


Acq: 01 Sep 2016 15:48

Tgt Ion:	202	Resp:	23972
Ion Ratio	Lower	Upper	
202	100		
200	20.8	16.6	25.0
203	17.6	13.8	20.8



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E





#29

Terphenyl-d14

Concen: 2.53 ng

RT: 19.79 min Scan# 1118

Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Instrument :

BNA_M





ClientSampleId :

SSTDICC3.2

Tot Ion:	244	Resp:	11571
Ion Ratio	Lower	Upper	
244	100		
212	10.2	12.2	18.4#
122	3.7	9.6	14.4#

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM



#30

Benzo(a)anthracene

Concen: 2.88 ng




RT: 21.33 min Scan# 1200


Delta R.T. 0.00 min

Lab File: BM007360.D

Acq: 01 Sep 2016 15:48

Tgt Ion:	228	Resp:	23377
Ion Ratio	Lower	Upper	
228	100		
226	24.0	20.4	30.6
229	21.9	19.0	28.4









#31
 3,3'-Dichlorobenzidine
 Concen: 3.79 ng
 RT: 21.27 min Scan# 1197
 Delta R.T. 0.00 min
 Lab File: BM007360.D
 Acq: 01 Sep 2016 15:48

Instrument :
 BNA_M
 ClientSampled :
 SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
252	100		
254	63.4	49.4	74.0
126	4.0	16.3	24.5




Manual Integrations
 APPROVED


sohil
 9/2/2016 4:37:25 PM



#32
 Chrysene
 Concen: 2.83 ng m
 RT: 21.37 min Scan# 1202
 Delta R.T. 0.00 min
 Lab File: BM007360.D
 Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
228	100		
226	34.2	28.3	42.5
229	16.8	15.7	23.5









#33
Bis(2-ethylhexyl)phthalate
Concen: 3.06 ng
RT: 21.25 min Scan# 1196
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
149	100		
167	24.8	21.0	31.4
279	0.0	0.0	0.0




Manual Integrations
APPROVED


sohil
9/2/2016 4:37:25 PM



#34
Indeno(1,2,3-cd)pyrene
Concen: 3.38 ng
RT: 25.89 min Scan# 1573
Delta R.T. -0.01 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
276	100		
138	26.4	21.0	31.4
277	24.7	19.7	29.5









#35
Pervlene-d12
Concen: 5.00 ng
RT: 23.61 min Scan# 1354
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampled :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
264	100		
260	26.7	21.2	31.8
265	21.0	17.1	25.7




Manual Integrations
APPROVED


sohil
9/2/2016 4:37:25 PM



#36
Benzo(b)fluoranthene
Concen: 3.06 ng
RT: 22.93 min Scan# 1289
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.1	21.4	32.2
125	9.0	10.5	15.7#









#37
Benzo(k)fluoranthene
Concen: 2.88 ng
RT: 22.97 min Scan# 1293
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.4	20.6	30.8
125	9.3	10.7	16.1#




Manual Integrations
APPROVED


sohil
9/2/2016 4:37:25 PM



#38
Benzo(a)pyrene
Concen: 3.06 ng
RT: 23.52 min Scan# 1345
Delta R.T. 0.01 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.9	20.6	31.0
125	9.9	13.0	19.4#









#39
Dibenzo(a,h)anthracene
Concen: 3.15 ng
RT: 25.91 min Scan# 1575
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Instrument :
BNA_M
ClientSampled :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
278	100		
139	17.2	20.4	30.6#
279	24.1	24.2	36.4#

Manual Integrations
APPROVED

sohil
9/2/2016 4:37:25 PM



#40
Benzo(g,h,i)perylene
Concen: 3.03 ng
RT: 26.60 min Scan# 1641
Delta R.T. 0.00 min
Lab File: BM007360.D
Acq: 01 Sep 2016 15:48

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.3	24.2	36.2
138	22.6	23.1	34.7#

