

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090119\
 Data File : BM022480.D
 Acq On : 01 Sep 2019 16:58
 Operator : HP/JU
 Sample : K4619-15
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP5-COMP

Quant Time: Sep 02 04:20:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 02 03:54:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	10762	20.00	ng	0.00
21) Naphthalene-d8	10.29	136	36146	20.00	ng	0.00
39) Acenaphthene-d10	14.17	164	23789	20.00	ng	0.00
64) Phenanthrene-d10	16.95	188	56672	20.00	ng	0.01
76) Chrysene-d12	21.16	240	89669	20.00	ng	0.00
87) Perylene-d12	23.34	264	110588	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.13	112	57879	94.40	ng	0.00
7) Phenol-d6	6.70	99	71750	87.51	ng	0.00
23) Nitrobenzene-d5	8.69	82	49345	59.79	ng	0.00
42) 2,4,6-Tribromophenol	15.69	330	36864	75.73	ng	0.00
45) 2-Fluorobiphenyl	12.77	172	110456	52.60	ng	0.00
79) Terphenyl-d14	19.59	244	307364	45.24	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.67	163	8324	4.023	ng	Qvalue # 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090119\
Data File : BM022480.D
Acq On : 01 Sep 2019 16:58
Operator : HP/JU
Sample : K4619-15
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP5-COMP

Quant Time: Sep 02 04:20:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 02 03:54:42 2019
Response via : Initial Calibration

