

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090119\
 Data File : BM022481.D
 Acq On : 01 Sep 2019 17:35
 Operator : HP/JU
 Sample : K4619-18
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP6-COMP

Quant Time: Sep 02 04:23:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 02 03:54:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	12259	20.00	ng	0.00
21) Naphthalene-d8	10.29	136	40560	20.00	ng	0.00
39) Acenaphthene-d10	14.17	164	27357	20.00	ng	0.00
64) Phenanthrene-d10	16.95	188	61615	20.00	ng	0.01
76) Chrysene-d12	21.16	240	89866	20.00	ng	0.00
87) Perylene-d12	23.34	264	109520	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.13	112	57965	83.00	ng	0.00
7) Phenol-d6	6.70	99	70521	75.51	ng	0.00
23) Nitrobenzene-d5	8.69	82	49529	53.49	ng	0.00
42) 2,4,6-Tribromophenol	15.69	330	35887	64.11	ng	0.00
45) 2-Fluorobiphenyl	12.77	172	103875	43.01	ng	0.00
79) Terphenyl-d14	19.59	244	258896	38.02	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.67	163	7540	3.169	ng	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090119\
Data File : BM022481.D
Acq On : 01 Sep 2019 17:35
Operator : HP/JU
Sample : K4619-18
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP6-COMP

Quant Time: Sep 02 04:23:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 02 03:54:42 2019
Response via : Initial Calibration

