

Data Path : Z:\HPCHEM1\BNA_M\Data\BM090215\
 Data File : BM002639.D
 Acq On : 02 Sep 2015 11:35
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02070

Quant Time: Sep 02 17:00:04 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 02 00:45:29 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	155#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	134	0.00
3	Naphthalene	1.021	0.962	5.8	129	0.00
4	2-Methylnaphthalene	0.765	0.672	12.2	121	0.00
5	1-Methylnaphthalene	0.756	0.654	13.5	118	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
7 S	Acenaphthylene-d8	2.045	1.945	4.9	104	0.00
8	Acenaphthylene	2.140	2.024	5.4	104	0.00
9 C	Acenaphthene	1.410	1.335	5.3	105	0.00
10 S	Fluorene-d10	1.436	1.245	13.3	96	0.00
11	Fluorene	1.615	1.411	12.6	96	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
13	Phenanthrene	1.155	1.102	4.6	82	0.00
14 S	Anthracene-d10	0.965	0.907	6.0	80	0.00
15	Anthracene	1.170	1.110	5.1	80	0.00
16 C	Fluoranthene	1.308	1.061	18.9	70	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	72	0.00
18 S	Pyrene-d10	0.887	0.909	-2.5	69	0.00
19	Pyrene	1.172	1.218	-3.9	69	0.00
20	Benzo(a)anthracene	1.196	1.124	6.0	68	0.00
21	Chrysene	1.119	1.049	6.3	67	0.00
22 I	Perylene-d12	1.000	1.000	0.0	74	0.00
23	Benzo(b)fluoranthene	1.209	1.108	8.4	70	0.00
24	Benzo(k)fluoranthene	1.146	1.065	7.1	68	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.976	7.1	70	0.00
26 C	Benzo(a)pyrene	1.155	1.083	6.2	70	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.285	3.2	72	0.00
28	Dibenzo(a,h)anthracene	1.110	1.085	2.3	73	0.00
29	Benzo(g,h,i)perylene	1.106	1.083	2.1	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0