

Data Path : Z:\HPCHEM1\BNA_M\Data\BM090215\
 Data File : BM002642.D
 Acq On : 02 Sep 2015 13:42
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02071

Quant Time: Sep 02 17:55:07 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 02 17:21:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	140	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
3	Naphthalene	1.021	0.959	6.1	123	0.00
4	2-Methylnaphthalene	0.765	0.691	9.7	119	0.00
5	1-Methylnaphthalene	0.756	0.674	10.8	116	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
7 S	Acenaphthylene-d8	2.045	1.941	5.1	110	0.00
8	Acenaphthylene	2.140	2.039	4.7	110	0.00
9 C	Acenaphthene	1.410	1.338	5.1	111	0.00
10 S	Fluorene-d10	1.436	1.314	8.5	106	0.00
11	Fluorene	1.615	1.497	7.3	107	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
13	Phenanthrene	1.155	1.094	5.3	97	0.00
14 S	Anthracene-d10	0.965	0.915	5.2	97	0.00
15	Anthracene	1.170	1.107	5.4	96	0.00
16 C	Fluoranthene	1.308	1.029	21.3	82	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
18 S	Pyrene-d10	0.887	1.011	-14.0	79	0.00
19	Pyrene	1.172	1.348	-15.0	79	0.00
20	Benzo(a)anthracene	1.196	1.120	6.4	70	0.00
21	Chrysene	1.119	1.056	5.6	69	0.00
22 I	Perylene-d12	1.000	1.000	0.0	70	0.00
23	Benzo(b)fluoranthene	1.209	1.112	8.0	66	0.00
24	Benzo(k)fluoranthene	1.146	1.099	4.1	67	0.00
25 S	Benzo(a)pyrene-d12	1.051	0.981	6.7	66	0.00
26 C	Benzo(a)pyrene	1.155	1.086	6.0	66	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.287	3.1	68	0.00
28	Dibenzo(a,h)anthracene	1.110	1.081	2.6	69	0.00
29	Benzo(g,h,i)perylene	1.106	1.081	2.3	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0