

Data Path : Z:\HPCHEM1\BNA_M\Data\BM090216\
 Data File : BM007368.D
 Acq On : 02 Sep 2016 15:00
 Operator : UM/SJ
 Sample : H4612-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 NC-TS-003

Quant Time: Sep 02 15:37:54 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM090116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 12:51:17 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	3869	5.00	ng	-0.02
7) Naphthalene-d8	10.58	136	17416	5.00	ng	0.00
13) Acenaphthene-d10	14.42	164	10464	5.00	ng	0.00
19) Phenanthrene-d10	17.16	188	21711	5.00	ng	0.00
26) Chrysene-d12	21.35	240	21157	5.00	ng	0.00
35) Perylene-d12	23.62	264	16533	5.00	ng	0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.39	112	6906	8.65	ng	0.00
5) Phenol-d6	6.95	99	8868	8.12	ng	-0.02
8) Nitrobenzene-d5	8.94	82	3924	4.01	ng	-0.02
14) 2,4,6-Tribromophenol	15.90	330	3411	8.42	ng	-0.02
15) 2-Fluorobiphenyl	13.04	172	10454	3.22	ng	-0.01
29) Terphenyl-d14	19.79	244	8283	2.92	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) n-Nitrosodimethylamine	3.63	42	10	0.16	ng	# 1
6) bis(2-Chloroethyl)ether	7.23	93	31	0.04	ng	# 24
10) Naphthalene	10.61	128	423	0.11	ng	# 78
12) 2-Methylnaphthalene	12.24	142	199	0.07	ng	# 88
16) Acenaphthylene	14.13	152	1350	0.30	ng	95
17) Acenaphthene	14.48	154	313	0.10	ng	# 72
18) Fluorene	15.46	166	475	0.13	ng	89
22) Pentachlorophenol	16.83	266	58	0.17	ng	87
23) Phenanthrene	17.19	178	6836	1.23	ng	99
24) Anthracene	17.19	178	6786	1.22	ng	98
25) Fluoranthene	19.23	202	18285	2.80	ng	# 92
28) Pyrene	19.59	202	18498	3.07	ng	# 94
30) Benzo(a)anthracene	21.33	228	21747	3.55	ng	97
32) Chrysene	21.33	228	21704	3.64	ng	# 88
34) Indeno(1,2,3-cd)pyrene	25.90	276	6150	1.17	ng	97
36) Benzo(b)fluoranthene	22.93	252	19085	3.70	ng	# 93
37) Benzo(k)fluoranthene	22.93	252	19085	4.01	ng	# 92
38) Benzo(a)pyrene	23.52	252	9047	1.98	ng	98
39) Dibenzo(a,h)anthracene	25.90	278	1585	0.43	ng	# 60
40) Benzo(g,h,i)perylene	26.60	276	6393	1.66	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM090216\
Data File : BM007368.D
Acq On : 02 Sep 2016 15:00
Operator : UM/SJ
Sample : H4612-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
NC-TS-003

Quant Time: Sep 02 15:37:54 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM090116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 12:51:17 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.79 min Scan# 320

Delta R.T. -0.02 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion: 152 Resp: 3869

Ion Ratio Lower Upper

152 100

150 157.7 109.8 164.8




115 60.4 39.0 58.4#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#3

n-Nitrosodimethylamine

Concen: 0.16 ng

RT: 3.63 min Scan# 62

Delta R.T. -0.03 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00


Tgt Ion: 42 Resp: 10

Ion Ratio Lower Upper

42 100

74 10.0 78.9 118.3#



44 160.0 19.1 28.7#




Abundance Ion 42.00 (41.70 to 42.70): BM0

Ion 74.00 (73.70 to 74.70): BM0

Ion 44.00 (43.70 to 44.70): BM0





#4

2-Fluorophenol

Concen: 8.65 ng

RT: 5.39 min Scan# 176

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion: 112 Resp: 6906

Ion Ratio Lower Upper

112 100

64 48.9 36.9 55.3




63 30.1 22.6 34.0



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): B

Ion 63.00 (62.70 to 63.70): B



#5

Phenol-d6

Concen: 8.12 ng

RT: 6.95 min Scan# 271

Delta R.T. -0.02 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00


Tgt Ion: 99 Resp: 8868

Ion Ratio Lower Upper

99 100

42 22.7 19.8 29.8



71 37.9 28.1 42.1




Abundance Ion 99.00 (98.70 to 99.70): B

Ion 42.00 (41.70 to 42.70): B

Ion 71.00 (70.70 to 71.70): B





#6

bis(2-Chloroethyl)ether

Concen: 0.04 ng

RT: 7.23 min Scan# 287

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion: 93 Resp: 31

Ion Ratio Lower Upper

93 100

63 0.0 54.3 81.5#

95 0.0 25.7 38.5#






Abundance Ion 93.00 (92.70 to 93.70): BM007358.D (-285) (-)

Ion 63.00 (62.70 to 63.70): BM007358.D (-285) (-)

Ion 95.00 (94.70 to 95.70): BM007358.D (-285) (-)

Time-->



#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.58 min Scan# 558

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Tgt Ion: 136 Resp: 17416


Ion Ratio Lower Upper

136 100

137 12.3 9.4 14.0

54 3.7 4.0 6.0#

68 2.9 2.7 4.1





Abundance Ion 136.00 (135.70 to 136.70): BM007358.D (-553) (-)


Ion 137.00 (136.70 to 137.70): BM007358.D (-553) (-)

Ion 54.00 (53.70 to 54.70): BM007358.D (-553) (-)

Ion 68.00 (67.70 to 68.70): BM007358.D (-553) (-)

Time-->





#8

Nitrobenzene-d5

Concen: 4.01 ng

RT: 8.94 min Scan# 422

Delta R.T. -0.02 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion: 82 Resp: 3924

Ion Ratio Lower Upper

82 100

128 45.7 54.9 82.3#

54 58.9 53.8 80.6




Abundance Ion 82.00 (81.70 to 82.70): BM007368.D (-405) (-)


Ion 128.00 (127.70 to 128.70): BM007368.D (-405) (-)

Ion 54.00 (53.70 to 54.70): BM007368.D (-405) (-)

Time-->



Time-->



#10

Naphthalene

Concen: 0.11 ng

RT: 10.61 min Scan# 560

Delta R.T. -0.02 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00


Tgt Ion: 128 Resp: 423

Ion Ratio Lower Upper

128 100

129 28.1 15.5 23.3#

127 30.7 15.5 23.3#




Abundance Ion 128.00 (127.70 to 128.70): BM007368.D (-549) (-)


Ion 129.00 (128.70 to 129.70): BM007368.D (-549) (-)

Ion 127.00 (126.70 to 127.70): BM007368.D (-549) (-)

Time-->



Time-->



#12

2-Methylnaphthalene

Concen: 0.07 ng

RT: 12.24 min Scan# 701

Delta R.T. -0.01 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003





Tgt Ion:142 Resp: 199

Ion Ratio Lower Upper

142 100

141 90.9 70.2 105.4

115 62.9 35.7 53.5#



#13

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.42 min Scan# 864

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00




Tgt Ion:164 Resp: 10464


Ion Ratio Lower Upper

164 100

162 101.7 86.2 129.4

160 45.1 40.5 60.7









#14
 2,4,6-Tribromophenol
 Concen: 8.42 ng
 RT: 15.90 min Scan# 931
 Delta R.T. -0.02 min
 Lab File: BM007368.D
 Acq: 02 Sep 2016 15:00




Instrument :
 BNA_M
 ClientSampleId :
 NC-TS-003


Tgt Ion: 330 Resp: 3411
 Ion Ratio Lower Upper
 330 100
 332 95.1 79.2 118.8
 141 57.0 0.0 0.0#



#15
 2-Fluorobiphenyl
 Concen: 3.22 ng
 RT: 13.04 min Scan# 778
 Delta R.T. -0.01 min
 Lab File: BM007368.D
 Acq: 02 Sep 2016 15:00

Tgt Ion: 172 Resp: 10454
 Ion Ratio Lower Upper
 172 100
 171 37.3 32.2 48.2
 170 26.7 23.5 35.3









#16
Acenaphthylene
Concen: 0.30 ng
RT: 14.13 min Scan# 851
Delta R.T. 0.00 min
Lab File: BM007368.D
Acq: 02 Sep 2016 15:00




Instrument :
BNA_M
ClientSampleId :
NC-TS-003


Tgt Ion: 152 Resp: 1350
Ion Ratio Lower Upper
152 100
151 21.9 15.7 23.5
153 15.0 10.1 15.1



#17
Acenaphthene
Concen: 0.10 ng
RT: 14.48 min Scan# 867
Delta R.T. 0.00 min
Lab File: BM007368.D
Acq: 02 Sep 2016 15:00

Tgt Ion: 154 Resp: 313
Ion Ratio Lower Upper
154 100
153 100.3 85.1 127.7
152 0.0 40.9 61.3#









#18
Fluorene
 Concen: 0.13 ng
 RT: 15.46 min Scan# 911
 Delta R.T. -0.02 min
 Lab File: BM007368.D
 Acq: 02 Sep 2016 15:00




Instrument :
 BNA_M
 ClientSampleId :
 NC-TS-003


Tgt Ion:166 Resp: 475
 Ion Ratio Lower Upper
 166 100
 165 110.1 78.8 118.2
 167 15.2 10.8 16.2



#19
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.16 min Scan# 981
 Delta R.T. 0.00 min
 Lab File: BM007368.D
 Acq: 02 Sep 2016 15:00

Tgt Ion:188 Resp: 21711
 Ion Ratio Lower Upper
 188 100
 94 2.4 2.0 3.0
 80 1.9 1.6 2.4





#22

Pentachlorophenol

Concen: 0.17 ng

RT: 16.83 min Scan# 968

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion: 266 Resp: 58

Ion Ratio Lower Upper

266 100

268 93.5 66.4 99.6




264 90.3 61.6 92.4



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#23

Phenanthrene

Concen: 1.23 ng

RT: 17.19 min Scan# 982

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00


Tgt Ion: 178 Resp: 6836

Ion Ratio Lower Upper

178 100

176 20.2 16.6 24.8



179 15.9 12.5 18.7




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#24
 Anthracene
 Concen: 1.22 ng
 RT: 17.19 min Scan# 982
 Delta R.T. -0.10 min
 Lab File: BM007368.D
 Acq: 02 Sep 2016 15:00




Instrument :
 BNA_M
 ClientSampleId :
 NC-TS-003


Tgt Ion:178 Resp: 6786
 Ion Ratio Lower Upper
 178 100
 176 20.0 14.8 22.2
 179 15.6 12.2 18.2



#25
 Fluoranthene
 Concen: 2.80 ng
 RT: 19.23 min Scan# 1085
 Delta R.T. 0.00 min
 Lab File: BM007368.D
 Acq: 02 Sep 2016 15:00

Tgt Ion:202 Resp: 18285
 Ion Ratio Lower Upper
 202 100
 101 0.0 6.5 9.7#
 203 18.0 14.2 21.2









#26
Chrysene-d12
Concen: 5.00 ng
RT: 21.35 min Scan# 1201
Delta R.T. 0.00 min
Lab File: BM007368.D
Acq: 02 Sep 2016 15:00




Instrument :
BNA_M
ClientSampleId :
NC-TS-003


Tgt Ion:240 Resp: 21157
Ion Ratio Lower Upper
240 100
120 2.9 1.3 1.9#
236 25.6 19.1 28.7



#28
Pyrene
Concen: 3.07 ng
RT: 19.59 min Scan# 1106
Delta R.T. 0.00 min
Lab File: BM007368.D
Acq: 02 Sep 2016 15:00

Tgt Ion:202 Resp: 18498
Ion Ratio Lower Upper
202 100
200 21.2 16.6 25.0
203 22.5 13.8 20.8#





#29

Terphenyl-d14

Concen: 2.92 ng

RT: 19.79 min Scan# 1118

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion:244 Resp: 8283

Ion Ratio Lower Upper

244 100

212 11.4 12.2 18.4#




122 6.5 9.6 14.4#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#30

Benzo(a)anthracene

Concen: 3.55 ng

RT: 21.33 min Scan# 1200

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00


Tgt Ion:228 Resp: 21747

Ion Ratio Lower Upper

228 100

226 28.1 20.4 30.6



229 24.5 19.0 28.4




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#32

Chrysene

Concen: 3.64 ng

RT: 21.33 min Scan# 1200

Delta R.T. -0.04 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion:228 Resp: 21704

Ion Ratio Lower Upper

228 100

226 28.1 28.3 42.5#




229 24.5 15.7 23.5#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#34

Indeno(1,2,3-cd)pyrene

Concen: 1.17 ng

RT: 25.90 min Scan# 1573

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00


Tgt Ion:276 Resp: 6150

Ion Ratio Lower Upper

276 100

138 23.7 21.0 31.4



277 24.1 19.7 29.5




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#35
Perylene-d12
 Concen: 5.00 ng
 RT: 23.62 min Scan# 1354
 Delta R.T. 0.01 min
 Lab File: BM007368.D
 Acq: 02 Sep 2016 15:00




Instrument :
 BNA_M
 ClientSampleId :
 NC-TS-003


Tgt Ion: 264 Resp: 16533
 Ion Ratio Lower Upper
 264 100
 260 26.3 21.2 31.8
 265 23.6 17.1 25.7



#36
Benzo(b)fluoranthene
 Concen: 3.70 ng
 RT: 22.93 min Scan# 1288
 Delta R.T. 0.00 min
 Lab File: BM007368.D
 Acq: 02 Sep 2016 15:00

Tgt Ion: 252 Resp: 19085
 Ion Ratio Lower Upper
 252 100
 253 29.3 21.4 32.2
 125 17.2 10.5 15.7#





#37

Benzo(k)fluoranthene

Concen: 4.01 ng

RT: 22.93 min Scan# 1288

Delta R.T. -0.04 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion: 252 Resp: 19085

Ion Ratio Lower Upper




252 100

253 29.3 20.6 30.8

125 17.2 10.7 16.1#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#38

Benzo(a)pyrene

Concen: 1.98 ng

RT: 23.52 min Scan# 1344

Delta R.T. 0.01 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00


Tgt Ion: 252 Resp: 9047

Ion Ratio Lower Upper



252 100


253 26.0 20.6 31.0

125 18.4 13.0 19.4



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#39

Dibenzo(a,h)anthracene

Concen: 0.43 ng

RT: 25.90 min Scan# 1573

Delta R.T. -0.01 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00

Instrument :

BNA_M

ClientSampleId :

NC-TS-003


Tgt Ion: 278 Resp: 1585

Ion Ratio Lower Upper

278 100

139 57.2 20.4 30.6#




279 41.8 24.2 36.4#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#40

Benzo(g,h,i)perylene

Concen: 1.66 ng

RT: 26.60 min Scan# 1640

Delta R.T. 0.00 min

Lab File: BM007368.D

Acq: 02 Sep 2016 15:00


Tgt Ion: 276 Resp: 6393

Ion Ratio Lower Upper

276 100

277 27.5 24.2 36.2

138 28.5 23.1 34.7



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

