

Data Path : Z:\HPCHEM1\BNA M\DATA\BM090315\
 Data File : BM002661.D
 Acq On : 03 Sep 2015 23:09
 Operator : UM/IZ
 Sample : MDL-02-S
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-02-S

Quant Time: Sep 04 04:37:57 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM90315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 04 04:35:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.78	152	7694	0.40	ng/ul	0.00
4) Naphthalene-d8	11.64	136	31541	0.40	ng/ul	0.00
8) Acenaphthene-d10	15.36	164	13635	0.40	ng/ul	0.00
12) Phenanthrene-d10	18.08	188	23127	0.40	ng/ul	0.00
18) Chrysene-d12	22.16	240	18942	0.40	ng/ul	0.00
22) Perylene-d12	24.77	264	18582	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 1,4-Dioxane-d8	3.94	96	12	0.00	ng/uL	0.12
6) 2-Methylnaphthalene-d10	13.17	152	18799	0.44	ng/ul	0.00
16) Fluoranthene-d10	20.03	212	24305	0.46	ng/ul	-0.01


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Naphthalene	11.69	128	4621	0.06	ng/ul	95
7) 2-Methylnaphthalene	13.25	142	3482	0.06	ng/ul	98
9) Acenaphthylene	15.09	152	4033	0.06	ng/ul	97
10) Acenaphthene	15.42	153	2907	0.06	ng/ul	97
11) Fluorene	16.39	166	3022	0.06	ng/ul	92
13) Pentachlorophenol	17.74	266	8	0.00	ng/ul#	47
14) Phenanthrene	18.11	178	4159	0.06	ng/ul	96
15) Anthracene	18.20	178	3976	0.06	ng/ul	97
17) Fluoranthene	20.07	202	4353	0.06	ng/ul	90
19) Pyrene	20.42	202	4347	0.06	ng/ul#	93
20) Benzo(a)anthracene	22.14	228	4014	0.06	ng/ul#	90
21) Chrysene	22.19	228	3915	0.06	ng/ul#	89
23) Benzo(b)fluoranthene	23.96	252	4485	0.07	ng/ul#	54
24) Benzo(k)fluoranthene	24.01	252	3786	0.06	ng/ul#	52
25) Benzo(a)pyrene	24.66	252	3790	0.06	ng/ul#	50
26) Indeno(1,2,3-cd)pyrene	27.54	276	4472	0.06	ng/ul	97
27) Dibenzo(a,h)anthracene	27.55	278	3608	0.06	ng/ul#	69
28) Benzo(a,h,i)perylene	28.40	276	3928	0.06	ng/ul#	76


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM090315\
Data File : BM002661.D
Acq On : 03 Sep 2015 23:09
Operator : UM/IZ
Sample : MDL-02-S
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MDL-02-S

Quant Time: Sep 04 04:37:57 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM90315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 04 04:35:12 2015
Response via : Initial Calibration





#5

Naphthalene

Concen: 0.06 ng/ul

RT: 11.69 min Scan# 546

Delta R.T. 0.00 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09

Instrument :

BNA_M

ClientSampleId :

MDL-02-S


Tot Ion:128 Resp: 4621

Ion Ratio Lower Upper

128 100

129 14.2 9.8 14.8




127 15.4 10.7 16.1



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#7

2-Methylnaphthalene

Concen: 0.06 ng/ul

RT: 13.25 min Scan# 678

Delta R.T. 0.00 min

Lab File: BM002661.D


Acq: 03 Sep 2015 23:09

Tgt Ion:142 Resp: 3482

Ion Ratio Lower Upper



142 100


141 89.7 70.3 105.5



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E





#9

Acenaphthylene

Concen: 0.06 ng/ul

RT: 15.09 min Scan# 852

Delta R.T. 0.00 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09

Instrument :

BNA_M

ClientSampleId :

MDL-02-S





Tot Ion:152 Resp: 4033

Ion Ratio Lower Upper

152 100

151 20.6 16.8 25.2

153 16.4 11.3 16.9



#10

Acenaphthene

Concen: 0.06 ng/ul

RT: 15.42 min Scan# 871

Delta R.T. 0.00 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09




Tgt Ion:153 Resp: 2907


Ion Ratio Lower Upper

153 100

152 50.8 42.3 63.5

154 83.7 64.7 97.1









#11
Fluorene
 Concen: 0.06 ng/ul
 RT: 16.39 min Scan# 928
 Delta R.T. 0.00 min
 Lab File: BM002661.D
 Acq: 03 Sep 2015 23:09




Instrument :
 BNA_M
 ClientSampleId :
 MDL-02-S


Tot Ion:166 Resp: 3022
 Ion Ratio Lower Upper
 166 100
 165 100.3 87.6 131.4
 167 16.6 11.1 16.7



#13
Pentachlorophenol
 Concen: 0.00 ng/ul
 RT: 17.74 min Scan# 1011
 Delta R.T. 0.00 min
 Lab File: BM002661.D
 Acq: 03 Sep 2015 23:09

Tgt Ion:266 Resp: 8
 Ion Ratio Lower Upper
 266 100
 264 50.0 53.5 80.3#
 268 0.0 54.2 81.2#





#14

Phenanthrene

Concen: 0.06 ng/ul

RT: 18.11 min Scan# 1035

Delta R.T. 0.00 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09

Instrument :

BNA_M

ClientSampleId :

MDL-02-S


Tot Ion:178 Resp: 4159

Ion Ratio Lower Upper

178 100

176 21.4 16.2 24.4




179 18.5 12.8 19.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#15

Anthracene

Concen: 0.06 ng/ul

RT: 18.20 min Scan# 1041

Delta R.T. 0.00 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09


Tgt Ion:178 Resp: 3976

Ion Ratio Lower Upper

178 100

176 20.8 15.8 23.8



179 18.7 13.3 19.9




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#17
Fluoranthene
 Concen: 0.06 ng/ul
 RT: 20.07 min Scan# 1181
 Delta R.T. 0.00 min
 Lab File: BM002661.D
 Acq: 03 Sep 2015 23:09




Instrument :
 BNA_M
 ClientSampleId :
 MDL-02-S


Tot Ion:202 Resp: 4353
 Ion Ratio Lower Upper
 202 100
 101 10.4 0.0 33.1
 203 22.9 0.0 37.2



#19
Pyrene
 Concen: 0.06 ng/ul
 RT: 20.42 min Scan# 1210
 Delta R.T. 0.00 min
 Lab File: BM002661.D
 Acq: 03 Sep 2015 23:09

Tgt Ion:202 Resp: 4347
 Ion Ratio Lower Upper
 202 100
 200 23.5 18.0 27.0
 203 22.7 13.8 20.6#





#20

Benzo(a)anthracene

Concen: 0.06 ng/ul

RT: 22.14 min Scan# 1360

Delta R.T. 0.00 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09

Instrument :

BNA_M

ClientSampleId :

MDL-02-S


Tot Ion:228 Resp: 4014

Ion Ratio Lower Upper

228 100

226 31.3 23.0 34.4




229 27.3 15.2 22.8#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#21

Chrysene

Concen: 0.06 ng/ul

RT: 22.19 min Scan# 1365

Delta R.T. -0.01 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09


Tgt Ion:228 Resp: 3915

Ion Ratio Lower Upper

228 100

226 36.5 25.7 38.5



229 25.8 15.2 22.8#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#23

Benzo(b)fluoranthene

Concen: 0.07 ng/ul

RT: 23.96 min Scan# 1535

Delta R.T. -0.01 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09

Instrument :

BNA_M

ClientSampleId :

MDL-02-S


Tot Ion:252 Resp: 4485

Ion Ratio Lower Upper




252 100

253 41.6 0.0 48.0

125 44.7 0.0 36.6#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#24

Benzo(k)fluoranthene

Concen: 0.06 ng/ul

RT: 24.01 min Scan# 1540

Delta R.T. -0.01 min

Lab File: BM002661.D

Acq: 03 Sep 2015 23:09


Tgt Ion:252 Resp: 3786

Ion Ratio Lower Upper



252 100


253 42.7 20.8 31.2#

125 46.2 12.2 18.4#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E









#25
Benzo(a)pyrene
Concen: 0.06 ng/ul
RT: 24.66 min Scan# 1602
Delta R.T. 0.00 min
Lab File: BM002661.D
Acq: 03 Sep 2015 23:09




Instrument :
BNA_M
ClientSampled :
MDL-02-S


Tot Ion:252 Resp: 3790
Ion Ratio Lower Upper
252 100
253 45.5 21.8 32.8#
125 50.1 14.5 21.7#



#26
Indeno(1,2,3-cd)pyrene
Concen: 0.06 ng/ul
RT: 27.54 min Scan# 1878
Delta R.T. 0.00 min
Lab File: BM002661.D
Acq: 03 Sep 2015 23:09

Tgt Ion:276 Resp: 4472
Ion Ratio Lower Upper
276 100
138 35.0 27.5 41.3
277 27.0 19.3 28.9









#27
Dibenzo(a,h)anthracene
Concen: 0.06 ng/ul
RT: 27.55 min Scan# 1879
Delta R.T. 0.00 min
Lab File: BM002661.D
Acq: 03 Sep 2015 23:09

Instrument :
BNA_M
ClientSampleId :
MDL-02-S

Tot Ion:278 Resp: 3608
Ion Ratio Lower Upper
278 100
139 42.6 21.7 32.5#
279 44.3 21.9 32.9#



#28
Benzo(g,h,i)perylene
Concen: 0.06 ng/ul
RT: 28.40 min Scan# 1961
Delta R.T. -0.01 min
Lab File: BM002661.D
Acq: 03 Sep 2015 23:09

Tgt Ion:276 Resp: 3928
Ion Ratio Lower Upper
276 100
277 36.7 21.3 31.9#
138 47.4 25.5 38.3#

