

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM090315\
 Data File : BM002658.D
 Acq On : 03 Sep 2015 21:19
 Operator : UM/IZ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.468

Quant Time: Sep 04 04:32:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM90315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 03 14:33:37 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 SURR	1,4-Dioxane-d8	0.403	0.366	9.2	99	0.00
3	1,4-Dioxane	0.465	0.411	11.6	99	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
5	Naphthalene	1.037	0.977	5.8	90	0.00
6 SURR	2-Methylnaphthalene-d10	0.537	0.486	9.5	86	0.00
7	2-Methylnaphthalene	0.691	0.634	8.2	87	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
9	Acenaphthylene	2.132	2.019	5.3	82	0.00
10 C	Acenaphthene	1.493	1.398	6.4	80	0.00
11	Fluorene	1.588	1.424	10.3	77	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
13	Pentachlorophenol	0.039	0.037	5.1	96	0.00
14	Phenanthrene	1.197	1.128	5.8	78	0.00
15	Anthracene	1.191	1.103	7.4	77	0.00
16 SURR	Fluoranthene-d10	0.913	0.883	3.3	83	0.00
17 C	Fluoranthene	1.252	1.200	4.2	85	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
19	Pyrene	1.512	1.348	10.8	85	0.00
20	Benzo(a)anthracene	1.326	1.246	6.0	91	0.00
21	Chrysene	1.283	1.206	6.0	92	0.00
22 I	Perylene-d12	1.000	1.000	0.0	99	0.00
23	Benzo(b)fluoranthene	1.413	1.326	6.2	97	0.00
24	Benzo(k)fluoranthene	1.326	1.267	4.4	94	0.00
25 C	Benzo(a)pyrene	1.320	1.235	6.4	93	0.00
26	Indeno(1,2,3-cd)pyrene	1.563	1.478	5.4	94	0.00
27	Dibenzo(a,h)anthracene	1.280	1.204	5.9	93	0.00
28	Benzo(g,h,i)perylene	1.339	1.275	4.8	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0