

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM090315\
 Data File : BM002667.D
 Acq On : 04 Sep 2015 02:49
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.469

Quant Time: Sep 04 04:38:56 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM90315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 04 04:35:12 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2 SURR	1,4-Dioxane-d8	0.403	0.344	14.6	90	0.00
3	1,4-Dioxane	0.465	0.390	16.1	91	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
5	Naphthalene	1.037	0.978	5.7	90	0.00
6 SURR	2-Methylnaphthalene-d10	0.537	0.491	8.6	87	0.00
7	2-Methylnaphthalene	0.691	0.643	6.9	88	-0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
9	Acenaphthylene	2.132	1.946	8.7	81	0.00
10 C	Acenaphthene	1.493	1.417	5.1	84	0.00
11	Fluorene	1.588	1.452	8.6	81	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
13	Pentachlorophenol	0.039	0.034	12.8	90	0.00
14	Phenanthrene	1.197	1.129	5.7	80	0.00
15	Anthracene	1.191	1.109	6.9	79	0.00
16 SURR	Fluoranthene-d10	0.913	0.861	5.7	83	-0.01
17 C	Fluoranthene	1.252	1.164	7.0	84	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
19	Pyrene	1.512	1.391	8.0	84	0.00
20	Benzo(a)anthracene	1.326	1.237	6.7	87	0.00
21	Chrysene	1.283	1.229	4.2	90	-0.01
22 I	Perylene-d12	1.000	1.000	0.0	94	-0.01
23	Benzo(b)fluoranthene	1.413	1.277	9.6	89	-0.01
24	Benzo(k)fluoranthene	1.326	1.292	2.6	91	-0.01
25 C	Benzo(a)pyrene	1.320	1.255	4.9	90	0.00
26	Indeno(1,2,3-cd)pyrene	1.563	1.476	5.6	89	0.00
27	Dibenzo(a,h)anthracene	1.280	1.208	5.6	89	-0.01
28	Benzo(g,h,i)perylene	1.339	1.274	4.9	91	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0