


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM090315\
Data File : BM002643.D
Acq On : 03 Sep 2015 09:37
Operator : UM/IZ
Sample : DFTPP11
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP11

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM90315.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Sep 04 04:35:12 2015



AutoFind: Scans 2606, 2607, 2608; Background Corrected with Scan 2600

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.5	17885	PASS
68	69	0.00	2	0.7	123	PASS
69	198	0.00	100	33.9	18664	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.2	27045	PASS
197	198	0.00	1	0.5	278	PASS
198	198	100	100	100.0	55005	PASS
199	198	5	9	7.0	3854	PASS
275	198	10	60	26.5	14553	PASS
365	198	1	100	3.0	1674	PASS
441	443	0.01	100	78.0	7652	PASS
442	198	50	100	92.7	51002	PASS
443	442	15	24	19.2	9808	PASS

Scan 2600 (18.380 min): BM002643.D
DFTPP11

m/z	Abundance
39.90	364.0
44.00	1604.0
207.00	674.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

m/z	Abundance
39.00	1354.0
39.90	366.0
44.00	1770.0
50.00	3985.0
51.00	16299.0
52.00	904.0
56.00	546.0
57.00	1282.0
63.00	623.0
65.00	438.0
69.00	16632.0
74.00	1343.0
75.00	2622.0
76.10	931.0
77.00	19344.0
78.00	1275.0
79.00	1297.0
80.00	950.0
81.00	1558.0
83.00	459.0
85.90	362.0
91.00	340.0
91.90	340.0
93.00	1993.0
98.00	1384.0
99.00	1562.0
100.90	831.0
103.90	397.0
105.00	560.0
107.00	5686.0
108.00	956.0
110.00	11540.0
111.00	1719.0
117.00	3767.0
118.00	334.0
121.90	398.0
122.90	634.0
124.90	343.0
127.00	23080.0
128.00	1733.0
129.00	8187.0
130.00	626.0
135.00	724.0
137.10	330.0
141.00	986.0
146.90	537.0
148.00	1123.0
153.90	327.0
155.00	589.0
156.00	876.0
160.90	556.0
164.90	481.0
167.00	1859.0
167.90	1097.0
174.00	379.0
175.00	827.0
177.00	406.0
179.00	1448.0
180.00	1019.0
181.00	468.0
185.00	756.0
186.00	5459.0
187.00	1491.0
191.90	383.0
192.90	429.0
196.00	1247.0
197.90	42368.0
198.90	2769.0
202.90	305.0
204.00	1292.0
205.00	2509.0
206.00	10534.0
207.00	1912.0
208.00	452.0
210.80	421.0
216.90	2490.0
218.00	379.0
221.00	1774.0
223.00	455.0
224.00	5670.0
225.00	1481.0
227.00	2046.0
227.90	325.0
228.90	533.0
244.00	4538.0
245.10	628.0
245.90	819.0
255.00	22888.0
256.00	3311.0
258.00	1118.0
264.90	435.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

272.90	540.0
274.00	1723.0
275.00	10639.0
275.90	1460.0
277.00	823.0
296.00	2387.0
296.90	479.0
303.00	301.0
323.00	840.0
334.00	572.0
352.00	301.0
365.00	1015.0
372.00	394.0
423.00	1737.0
424.00	384.0
441.00	4347.0
442.00	30256.0
443.00	6040.0
444.00	679.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

m/z	Abundance
39.00	1741.0
40.00	404.0
44.00	1977.0
50.00	5235.0
51.00	21120.0
52.00	1033.0
56.00	744.0
57.00	1558.0
63.00	922.0
65.00	439.0
69.00	22304.0
73.10	315.0
74.10	2168.0
75.00	3649.0
76.00	1094.0
77.10	25688.0
78.00	1674.0
79.00	1574.0
80.00	1209.0
81.00	1904.0
81.90	416.0
82.90	436.0
85.00	342.0
85.90	478.0
90.90	513.0
92.00	383.0
93.00	2664.0
97.90	2092.0
98.90	1575.0
101.00	1187.0
103.00	432.0
104.00	572.0
105.00	603.0
107.00	8101.0
108.00	1265.0
110.00	15996.0
111.00	2348.0
112.00	346.0
116.00	482.0
117.00	5593.0
118.00	396.0
121.90	505.0
123.00	829.0
123.90	410.0
125.00	450.0
127.00	31112.0
128.00	2462.0
129.00	11726.0
129.90	977.0
133.90	370.0
135.00	1200.0
136.00	390.0
137.00	428.0
140.90	1302.0
141.90	447.0
145.90	341.0
147.00	924.0
148.00	1528.0
149.00	373.0
151.10	312.0
152.90	497.0
154.00	358.0
155.00	866.0
156.00	1079.0
160.00	434.0
161.00	646.0
164.90	535.0
165.90	420.0
167.00	2726.0
168.00	1421.0
173.00	398.0
174.00	682.0
175.00	1305.0
175.90	327.0
176.90	637.0
179.00	1824.0
180.00	1348.0
180.90	654.0
184.90	1086.0
186.00	8776.0
187.00	2661.0
188.90	591.0
190.90	314.0
191.90	659.0
193.00	830.0
196.00	2269.0
197.10	485.0
197.90	65488.0
199.00	4637.0
200.00	395.0
201.20	307.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

202.90	397.0
204.00	2096.0
205.00	3520.0
206.00	16309.0
207.00	2825.0
207.90	595.0
210.80	604.0
216.90	3820.0
217.90	600.0
221.00	3287.0
221.80	663.0
223.00	972.0
224.00	8519.0
225.00	2117.0
227.00	3248.0
228.00	364.0
228.90	644.0
236.90	344.0
242.00	374.0
242.90	524.0
244.00	7052.0
245.00	876.0
246.00	1191.0
254.10	320.0
255.00	36072.0
256.00	5427.0
257.10	455.0
257.90	1698.0
264.90	777.0
273.00	1107.0
274.00	2985.0
275.00	16648.0
276.00	2155.0
277.00	1102.0
292.90	322.0
296.00	4029.0
296.90	509.0
302.90	389.0
314.00	329.0
314.90	481.0
323.00	1332.0
323.90	300.0
334.10	868.0
346.00	323.0
352.00	563.0
353.00	414.0
354.00	577.0
365.00	1856.0
372.00	794.0
402.00	385.0
403.00	539.0
421.00	424.0
422.00	515.0
423.00	2855.0
424.00	673.0
441.00	8333.0
442.00	56912.0
443.00	10967.0
444.00	1004.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

m/z	Abundance
39.00	1349.0
39.90	411.0
44.00	1911.0
50.00	4009.0
51.00	16236.0
52.00	886.0
56.00	569.0
57.00	1089.0
63.00	653.0
65.00	368.0
68.00	369.0
69.00	17056.0
74.00	1557.0
75.00	2767.0
76.10	849.0
77.00	20168.0
78.00	1549.0
79.00	1383.0
79.90	1103.0
81.00	1354.0
82.00	375.0
83.00	389.0
86.00	354.0
91.00	363.0
92.10	361.0
93.00	2054.0
98.00	1538.0
99.00	1399.0
100.90	840.0
103.00	306.0
104.00	481.0
104.90	628.0
107.00	6321.0
108.00	938.0
110.00	12982.0
111.00	1785.0
116.00	430.0
117.00	4258.0
122.00	432.0
122.90	720.0
123.90	337.0
125.00	300.0
127.00	26944.0
128.00	2180.0
129.00	9553.0
130.00	875.0
135.00	770.0
135.90	334.0
137.00	338.0
140.90	1184.0
142.00	445.0
142.90	300.0
147.00	654.0
148.00	1360.0
152.90	407.0
154.00	305.0
155.00	718.0
155.90	1111.0
159.90	405.0
160.90	655.0
164.90	451.0
165.90	522.0
167.00	2299.0
168.00	1190.0
173.00	358.0
174.00	676.0
175.00	1003.0
176.00	359.0
176.90	615.0
178.90	1704.0
180.00	1307.0
180.90	629.0
185.00	1081.0
186.00	7740.0
186.90	2138.0
188.90	536.0
192.00	626.0
193.00	576.0
196.00	1820.0
197.10	349.0
198.00	57160.0
198.90	4156.0
201.40	386.0
202.90	396.0
204.00	1946.0
205.00	3199.0
206.00	14699.0
207.00	2641.0
208.00	590.0
210.80	501.0
216.90	3709.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

218.00	462.0
221.00	2875.0
222.90	759.0
224.00	7956.0
225.00	2141.0
227.00	2881.0
227.90	426.0
229.00	749.0
231.00	325.0
237.00	342.0
241.90	378.0
243.10	512.0
244.00	6912.0
245.10	717.0
245.90	1027.0
255.00	34032.0
256.00	5058.0
258.00	1794.0
264.90	694.0
273.00	982.0
274.00	2575.0
275.00	16372.0
276.00	2040.0
277.00	1131.0
296.00	4205.0
297.00	576.0
302.90	525.0
315.00	456.0
323.00	1567.0
334.00	945.0
352.00	558.0
353.10	339.0
354.00	531.0
364.90	2151.0
372.00	766.0
402.00	389.0
403.00	561.0
421.00	567.0
422.00	467.0
423.00	3166.0
424.00	528.0
441.00	10278.0
442.00	65840.0
443.00	12418.0
444.00	1068.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11