

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090319\
 Data File : BM022500.D
 Acq On : 03 Sep 2019 13:27
 Operator : HP/JU
 Sample : K4619-12
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Sep 04 02:01:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 02 03:54:42 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	17433	20.00	ng	0.00
21) Naphthalene-d8	10.29	136	57441	20.00	ng	0.00
39) Acenaphthene-d10	14.17	164	35340	20.00	ng	0.00
64) Phenanthrene-d10	16.94	188	69805	20.00	ng	0.00
76) Chrysene-d12	21.16	240	76418	20.00	ng	0.00
87) Perylene-d12	23.34	264	84281	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.13	112	122594	123.44	ng	0.00
7) Phenol-d6	6.70	99	148317	111.68	ng	0.00
23) Nitrobenzene-d5	8.68	82	109554	83.54	ng	0.00
42) 2,4,6-Tribromophenol	15.69	330	69897	96.66	ng	0.00
45) 2-Fluorobiphenyl	12.77	172	225366	72.24	ng	0.00
79) Terphenyl-d14	19.58	244	403813	69.74	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.66	163	19322	6.286	ng	Qvalue 100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090319\
Data File : BM022500.D
Acq On : 03 Sep 2019 13:27
Operator : HP/JU
Sample : K4619-12
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Sep 04 02:01:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 02 03:54:42 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.50 min Scan# 768
 Delta R.T. -0.01 min
 Lab File: BM022500.D
 Acq: 03 Sep 2019 13:27

Instrument :
 BNA_M
 ClientSampled :

Tot Ion:152 Resp: 17433
 Ion Ratio Lower Upper
 152 100
 150 144.3 125.4 188.0
 115 64.3 50.3 75.5




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
 Concen: 123.440 ng
 RT: 5.13 min Scan# 364
 Delta R.T. -0.00 min
 Lab File: BM022500.D
 Acq: 03 Sep 2019 13:27

Tgt Ion:112 Resp: 122594
 Ion Ratio Lower Upper
 112 100
 64 57.3 45.4 68.2
 63 43.0 34.5 51.7



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BM0
 Ion 63.00 (62.70 to 63.70): BM0





#7

Phenol-d6

Concen: 111.676 ng

RT: 6.70 min Scan# 632

Delta R.T. -0.01 min

Lab File: BM022500.D

Acq: 03 Sep 2019 13:27

Instrument :

BNA_M

ClientSampled :


Tot Ion: 99 Resp: 148317

Ion Ratio Lower Upper

99 100

42 32.6 26.3 39.5




71 57.1 46.6 69.8



Abundance Ion 99.00 (98.70 to 99.70): BM022500.D (-616) (-)

Ion 42.00 (41.70 to 42.70): BM022500.D (-616) (-)

Ion 71.00 (70.70 to 71.70): BM022500.D (-616) (-)



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.29 min Scan# 1241

Delta R.T. 0.01 min

Lab File: BM022500.D

Acq: 03 Sep 2019 13:27

Tgt Ion: 136 Resp: 57441


Ion Ratio Lower Upper

136 100

137 9.2 8.6 13.0

54 9.7 7.0 10.6

68 4.8 2.9 4.3#






Abundance Ion 136.00 (135.70 to 136.70): BM022500.D (-1223) (-)

Ion 137.00 (136.70 to 137.70): BM022500.D (-1223) (-)

Ion 54.00 (53.70 to 54.70): BM022500.D (-1223) (-)

Ion 68.00 (67.70 to 68.70): BM022500.D (-1223) (-)









#23
 Nitrobenzene-d5
 Concen: 83.538 ng
 RT: 8.68 min Scan# 968
 Delta R.T. -0.00 min
 Lab File: BM022500.D
 Acq: 03 Sep 2019 13:27




Instrument :
 BNA_M
 ClientSampleID :


Tot Ion: 82 Resp: 109554
 Ion Ratio Lower Upper
 82 100
 128 35.6 27.9 41.9
 54 56.8 45.2 67.8



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.17 min Scan# 1901
 Delta R.T. -0.00 min
 Lab File: BM022500.D
 Acq: 03 Sep 2019 13:27

Tgt Ion: 164 Resp: 35340
 Ion Ratio Lower Upper
 164 100
 162 102.5 80.6 121.0
 160 48.4 37.3 55.9






#42
 2,4,6-Tribromophenol
 Concen: 96.656 ng
 RT: 15.69 min Scan# 2159
 Delta R.T. -0.00 min
 Lab File: BM022500.D
 Acq: 03 Sep 2019 13:27

Instrument :
 BNA_M
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
330	100		
332	94.1	77.8	116.8
141	52.5	35.6	53.4






Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45
 2-Fluorobiphenyl
 Concen: 72.243 ng
 RT: 12.77 min Scan# 1663
 Delta R.T. -0.01 min
 Lab File: BM022500.D
 Acq: 03 Sep 2019 13:27

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.3	28.5	42.7
170	24.3	18.9	28.3






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E





#50

Dimethylphthalate

Concen: 6.286 ng

RT: 13.66 min Scan# 1814

Delta R.T. 0.01 min

Lab File: BM022500.D

Acq: 03 Sep 2019 13:27

Instrument :

BNA_M

ClientSampled :


Tot Ion:163 Resp: 19322

Ion Ratio Lower Upper




163 100

194 4.4 3.4 5.2

164 10.0 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 16.94 min Scan# 2373

Delta R.T. 0.01 min

Lab File: BM022500.D

Acq: 03 Sep 2019 13:27


Tgt Ion:188 Resp: 69805

Ion Ratio Lower Upper



188 100


94 6.2 4.4 6.6

80 7.9 5.0 7.6#



Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BM0
Ion 80.00 (79.70 to 80.70): BM0









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.16 min Scan# 3090
Delta R.T. 0.01 min
Lab File: BM022500.D
Acq: 03 Sep 2019 13:27

Instrument :
BNA_M
ClientSampled :

Tot Ion:240 Resp: 76418
Ion Ratio Lower Upper
240 100
120 4.5 3.9 5.9
236 27.0 21.4 32.2




Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E






#79
Terphenyl-d14
Concen: 69.739 ng
RT: 19.58 min Scan# 2821
Delta R.T. -0.01 min
Lab File: BM022500.D
Acq: 03 Sep 2019 13:27

Tgt Ion:244 Resp: 403813
Ion Ratio Lower Upper
244 100
212 9.6 7.0 10.4
122 5.4 3.9 5.9



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E






#87
 Pervlene-d12
 Concen: 20.000 ng
 RT: 23.34 min Scan# 3461
 Delta R.T. -0.00 min
 Lab File: BM022500.D
 Acq: 03 Sep 2019 13:27

Instrument :
 BNA_M
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
264	100		
260	25.8	20.2	30.2
265	25.3	18.2	27.2



Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

