

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090319\
 Data File : BM022497.D
 Acq On : 03 Sep 2019 11:38
 Operator : HP/JU
 Sample : K4619-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP1-COMP

Quant Time: Sep 04 02:00:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 02 03:54:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	16841	20.00	ng	0.00
21) Naphthalene-d8	10.29	136	58036	20.00	ng	0.00
39) Acenaphthene-d10	14.17	164	35367	20.00	ng	0.00
64) Phenanthrene-d10	16.94	188	70651	20.00	ng	0.00
76) Chrysene-d12	21.16	240	80405	20.00	ng	0.00
87) Perylene-d12	23.35	264	96062	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.13	112	89136	92.91	ng	0.00
7) Phenol-d6	6.70	99	107873	84.08	ng	0.00
23) Nitrobenzene-d5	8.69	82	82124	61.98	ng	0.00
42) 2,4,6-Tribromophenol	15.69	330	49740	68.73	ng	0.00
45) 2-Fluorobiphenyl	12.77	172	175718	56.28	ng	0.00
79) Terphenyl-d14	19.58	244	333070	54.67	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.66	163	14661	4.766	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090319\
Data File : BM022497.D
Acq On : 03 Sep 2019 11:38
Operator : HP/JU
Sample : K4619-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP1-COMP

Quant Time: Sep 04 02:00:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 02 03:54:42 2019
Response via : Initial Calibration

