

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090319\
 Data File : BM022512.D
 Acq On : 03 Sep 2019 20:47
 Operator : HP/JU
 Sample : K4563-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 CIW-6S-(0-4)

Quant Time: Sep 04 02:03:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 02 03:54:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	18302	20.00	ng	-0.01
21) Naphthalene-d8	10.28	136	61110	20.00	ng	0.00
39) Acenaphthene-d10	14.17	164	38054	20.00	ng	0.00
64) Phenanthrene-d10	16.94	188	78809	20.00	ng	0.00
76) Chrysene-d12	21.16	240	81674	20.00	ng	0.00
87) Perylene-d12	23.34	264	84793	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.13	112	124812	119.71	ng	0.00
7) Phenol-d6	6.70	99	154759	110.99	ng	0.00
23) Nitrobenzene-d5	8.68	82	116426	83.45	ng	0.00
42) 2,4,6-Tribromophenol	15.68	330	72852	93.56	ng	0.00
45) 2-Fluorobiphenyl	12.77	172	234633	69.85	ng	0.00
79) Terphenyl-d14	19.58	244	433780	70.09	ng	0.00
Target Compounds						
10) Phenol	6.73	94	3804	2.716	ng	Qvalue 82
50) Dimethylphthalate	13.66	163	14482	4.376	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090319\
Data File : BM022512.D
Acq On : 03 Sep 2019 20:47
Operator : HP/JU
Sample : K4563-01
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
CIW-6S-(0-4)

Quant Time: Sep 04 02:03:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 02 03:54:42 2019
Response via : Initial Calibration

