



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090321\
 Data File : BM032036.D
 Acq On : 03 Sep 2021 10:19
 Operator : CG/JU
 Sample : DFTPP016
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP016

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM090321.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Sep 03 16:21:27 2021



AutoFind: Scans 2423, 2424, 2425; Background Corrected with Scan 2417

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.8	30749	PASS
68	69	0.00	2	2.0	611	PASS
69	198	0.00	100	43.6	30605	PASS
70	69	0.00	2	0.7	222	PASS
127	198	10	80	48.3	33883	PASS
197	198	0.00	2	0.7	509	PASS
198	198	100	100	100.0	70208	PASS
199	198	5	9	6.8	4779	PASS
275	198	10	60	28.9	20323	PASS
365	198	1	100	5.3	3691	PASS
441	198	0.01	100	11.1	7827	PASS
442	198	50	100	71.2	49955	PASS
443	442	15	24	19.4	9692	PASS

m/z	Abundance
40.00	297.0
44.00	655.0
72.90	221.0
146.80	172.0
207.00	371.0
281.10	173.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

m/z	Abundance
36.90	226.0
37.90	413.0
39.10	2242.0
39.90	300.0
44.00	1096.0
49.10	341.0
50.10	8908.0
51.10	29200.0
52.00	1765.0
55.00	212.0
55.90	819.0
57.00	1770.0
61.10	523.0
62.00	502.0
63.00	1261.0
65.00	390.0
68.00	591.0
69.00	30208.0
69.90	204.0
73.10	599.0
74.00	3580.0
75.00	5252.0
76.10	1696.0
77.10	31040.0
78.00	2605.0
79.00	2685.0
80.00	1815.0
81.00	2512.0
82.00	661.0
83.00	474.0
83.80	187.0
84.90	216.0
86.00	822.0
86.90	359.0
88.00	284.0
90.90	517.0
92.00	684.0
93.00	4923.0
93.90	324.0
96.10	183.0
98.00	4168.0
99.00	3336.0
99.90	353.0
101.00	1150.0
103.00	539.0
104.00	966.0
104.90	944.0
105.80	346.0
107.00	10896.0
108.00	1596.0
109.10	315.0
110.00	15082.0
110.90	2842.0
112.00	338.0
116.00	624.0
117.00	12806.0
118.00	870.0
122.00	876.0
122.90	1073.0
123.90	504.0
125.00	433.0
126.20	164.0
127.00	30040.0
128.00	2446.0
129.00	13640.0
130.00	1250.0
131.00	306.0
132.10	157.0
133.90	503.0
134.90	1180.0
136.00	472.0
137.00	940.0
137.90	181.0
139.90	177.0
141.00	2363.0
141.90	433.0
142.90	368.0
145.00	165.0
145.90	405.0
147.00	1119.0
148.00	2921.0
148.90	561.0
149.90	172.0
150.90	233.0
152.90	539.0
153.90	317.0
155.00	955.0
156.10	1449.0
156.90	253.0
157.90	410.0
160.00	723.0
160.90	657.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

161.80	243.0
165.00	853.0
165.90	466.0
167.00	4800.0
168.00	3100.0
169.00	405.0
171.90	384.0
172.90	358.0
174.00	739.0
175.00	1553.0
176.10	375.0
176.80	697.0
178.00	230.0
179.00	2729.0
180.00	2142.0
181.00	829.0
185.00	1063.0
186.00	9420.0
187.00	2540.0
187.90	357.0
188.90	838.0
190.90	291.0
191.90	964.0
193.00	1164.0
194.00	276.0
194.80	156.0
196.00	1627.0
197.10	428.0
198.00	57912.0
198.90	3906.0
199.90	254.0
201.30	283.0
202.90	708.0
204.00	2321.0
205.00	4399.0
206.00	15598.0
207.00	2874.0
208.00	846.0
209.00	332.0
209.90	363.0
210.40	377.0
211.00	821.0
215.10	214.0
216.10	367.0
217.00	5164.0
218.00	861.0
221.00	2246.0
222.00	678.0
223.00	1068.0
224.00	9108.0
225.00	2377.0
225.90	296.0
227.00	4415.0
228.00	486.0
229.10	915.0
231.00	333.0
233.90	294.0
234.90	379.0
236.00	333.0
237.00	430.0
239.00	152.0
239.90	197.0
240.90	181.0
242.10	455.0
243.10	594.0
244.00	6619.0
245.00	917.0
246.00	1800.0
247.00	271.0
248.90	295.0
253.00	189.0
254.00	397.0
255.00	35744.0
256.00	5938.0
257.00	651.0
258.00	2479.0
258.90	426.0
260.00	151.0
265.00	1270.0
270.90	220.0
272.00	195.0
273.00	1099.0
274.00	2558.0
275.00	14768.0
276.00	1956.0
277.00	1437.0
278.00	306.0
280.90	307.0
282.90	285.0
285.00	346.0
292.00	164.0
293.00	312.0
296.00	5714.0
296.90	724.0
303.00	605.0
304.00	195.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

314.10	231.0
315.00	656.0
315.90	282.0
321.00	171.0
323.00	992.0
324.00	408.0
326.90	511.0
328.00	170.0
334.00	852.0
335.00	209.0
341.10	210.0
346.00	187.0
351.90	493.0
353.00	348.0
354.00	443.0
355.10	175.0
365.00	2232.0
366.00	255.0
372.00	828.0
372.90	287.0
382.90	175.0
402.00	237.0
403.00	333.0
404.00	193.0
420.90	225.0
421.90	270.0
423.00	1920.0
424.00	563.0
441.00	4689.0
442.00	31224.0
443.00	6013.0
444.10	526.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

m/z	Abundance
37.00	213.0
38.00	669.0
39.10	2431.0
39.90	315.0
43.90	842.0
49.00	359.0
50.00	10789.0
51.10	37968.0
52.00	2029.0
55.10	219.0
56.00	1112.0
57.00	2442.0
61.00	485.0
62.00	728.0
63.00	1493.0
64.00	278.0
65.10	616.0
68.00	787.0
69.00	36864.0
70.00	255.0
73.10	569.0
74.00	4903.0
75.00	6186.0
76.00	2314.0
77.10	42304.0
78.00	2616.0
79.00	3492.0
80.00	2905.0
81.00	3313.0
81.90	827.0
82.90	685.0
84.90	581.0
86.00	1036.0
87.00	434.0
88.10	175.0
90.90	630.0
91.90	740.0
93.00	6577.0
93.90	458.0
96.00	220.0
98.00	4855.0
99.00	3799.0
99.90	352.0
101.00	1879.0
103.00	681.0
104.00	1143.0
104.90	1141.0
106.00	344.0
107.00	14805.0
108.00	2043.0
109.00	382.0
110.00	20560.0
111.00	3444.0
112.10	475.0
112.90	224.0
114.90	261.0
116.00	664.0
117.00	17656.0
118.00	1326.0
120.10	187.0
121.90	1185.0
123.00	1384.0
123.90	737.0
124.90	464.0
126.00	285.0
127.00	41440.0
128.00	3173.0
129.00	17832.0
130.00	1704.0
130.90	205.0
132.10	203.0
133.90	576.0
135.00	1782.0
135.80	372.0
137.00	1051.0
141.00	2804.0
142.00	713.0
143.00	636.0
143.80	153.0
144.90	170.0
145.90	370.0
147.00	1438.0
148.00	4515.0
148.90	593.0
150.00	200.0
151.00	333.0
151.90	186.0
153.00	666.0
153.90	701.0
154.90	1474.0
156.00	1996.0
157.00	183.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

158.00	464.0
159.00	331.0
160.00	1001.0
161.00	1393.0
161.90	328.0
165.00	960.0
165.90	804.0
167.00	6557.0
168.00	3352.0
168.90	492.0
169.90	226.0
170.90	157.0
171.90	613.0
173.00	594.0
173.90	1064.0
175.00	2358.0
176.00	419.0
177.00	1099.0
177.80	421.0
179.00	4033.0
180.00	2575.0
181.00	1258.0
182.00	197.0
183.90	415.0
185.00	1844.0
186.00	13576.0
187.00	4048.0
187.90	358.0
189.00	1137.0
189.90	219.0
191.00	541.0
192.00	1330.0
193.00	1555.0
194.00	442.0
194.80	189.0
195.90	2411.0
197.10	636.0
197.90	85264.0
198.90	5836.0
200.00	505.0
201.20	282.0
203.00	822.0
204.00	3921.0
205.00	6784.0
206.00	24640.0
207.10	4153.0
208.00	934.0
208.90	519.0
210.10	532.0
211.00	1034.0
211.80	218.0
214.90	279.0
216.10	555.0
217.00	8152.0
218.00	1024.0
221.00	3143.0
222.00	1128.0
223.00	1878.0
224.00	14958.0
225.00	3809.0
226.00	426.0
227.00	7368.0
228.00	1296.0
229.00	1389.0
231.00	487.0
232.00	203.0
234.00	570.0
235.00	513.0
236.00	535.0
237.00	776.0
239.00	256.0
240.00	286.0
241.00	304.0
242.00	1023.0
243.00	1070.0
244.10	10630.0
245.00	1506.0
246.00	3270.0
246.90	486.0
249.00	327.0
252.90	395.0
254.10	458.0
255.00	59464.0
256.00	8669.0
257.00	644.0
258.00	4409.0
259.00	644.0
265.00	2081.0
265.90	267.0
272.00	238.0
273.00	1523.0
274.00	4520.0
275.00	24872.0
276.00	2989.0
277.00	2808.0
278.00	590.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

280.90	259.0
283.00	435.0
284.10	183.0
285.00	490.0
293.00	532.0
293.80	214.0
296.00	9869.0
297.00	1514.0
300.90	167.0
301.90	167.0
303.00	959.0
304.00	218.0
307.90	173.0
308.90	265.0
314.10	388.0
314.90	986.0
316.00	601.0
320.90	289.0
321.90	167.0
323.00	2611.0
324.10	474.0
327.00	618.0
328.00	266.0
332.90	289.0
334.00	1669.0
335.00	304.0
341.00	303.0
345.90	658.0
346.90	204.0
352.10	691.0
353.10	523.0
354.10	562.0
355.00	217.0
365.00	4499.0
366.00	806.0
371.00	248.0
372.00	1264.0
373.00	371.0
382.90	297.0
402.00	541.0
403.00	576.0
421.00	484.0
422.00	584.0
423.00	3467.0
424.00	1057.0
441.00	9090.0
442.10	58384.0
443.00	11108.0
444.00	817.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

m/z	Abundance
37.10	155.0
38.00	307.0
39.00	1918.0
39.90	332.0
44.00	757.0
49.10	203.0
50.10	6900.0
51.10	25080.0
51.90	1259.0
55.00	207.0
55.90	735.0
57.00	1533.0
61.00	352.0
62.00	516.0
63.00	1261.0
63.90	187.0
65.00	526.0
68.00	456.0
69.00	24744.0
69.90	208.0
73.00	591.0
74.10	3662.0
75.00	4517.0
76.00	1588.0
77.10	29360.0
78.10	1857.0
79.00	2130.0
80.00	1600.0
81.00	2161.0
82.00	666.0
83.00	405.0
85.00	377.0
86.00	642.0
87.00	372.0
91.00	547.0
91.90	830.0
93.00	5167.0
93.90	376.0
96.00	294.0
97.00	163.0
98.00	3291.0
99.00	2595.0
100.00	195.0
101.00	1065.0
103.00	307.0
103.90	906.0
104.90	749.0
106.10	319.0
107.00	10297.0
108.00	1671.0
108.90	301.0
110.00	14771.0
111.00	2230.0
112.00	454.0
116.10	542.0
117.00	11727.0
118.00	1091.0
118.90	174.0
121.90	668.0
123.00	1045.0
123.90	674.0
125.10	327.0
126.00	207.0
127.00	30168.0
128.00	2571.0
129.00	14161.0
130.00	1268.0
131.00	185.0
133.90	409.0
135.00	1003.0
136.10	389.0
137.00	749.0
137.90	183.0
139.90	179.0
141.00	1831.0
142.00	617.0
143.00	282.0
146.00	386.0
146.90	1069.0
147.90	3004.0
148.90	671.0
151.10	233.0
151.90	237.0
153.00	406.0
153.90	378.0
155.00	896.0
156.00	1423.0
156.80	170.0
157.90	381.0
158.90	275.0
160.00	594.0
161.00	1139.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

162.00	265.0
165.00	836.0
166.00	469.0
167.00	4662.0
168.00	2671.0
168.90	408.0
171.00	182.0
171.90	405.0
172.90	506.0
174.00	742.0
175.00	1656.0
175.90	318.0
177.00	837.0
178.00	249.0
179.00	3094.0
180.00	2326.0
181.00	853.0
182.00	222.0
183.90	220.0
185.00	1477.0
186.00	10620.0
187.00	3064.0
188.00	303.0
189.00	1063.0
191.00	575.0
191.90	1193.0
193.00	1149.0
193.90	293.0
196.00	1850.0
197.10	463.0
197.90	67448.0
198.90	4596.0
200.10	255.0
201.50	278.0
203.00	703.0
204.00	3000.0
205.00	5128.0
206.00	20032.0
207.00	3044.0
208.00	1070.0
209.00	285.0
210.00	484.0
211.00	696.0
215.00	252.0
216.10	411.0
217.00	6214.0
218.00	772.0
221.00	2519.0
221.90	693.0
223.00	1261.0
224.00	11409.0
225.00	2799.0
225.90	263.0
227.00	5771.0
228.00	877.0
229.00	1154.0
230.90	345.0
234.00	343.0
235.00	505.0
235.90	185.0
236.80	568.0
239.00	191.0
240.90	272.0
242.00	736.0
243.10	893.0
244.00	8877.0
245.00	1069.0
246.00	2494.0
247.00	565.0
249.00	240.0
251.80	159.0
252.90	213.0
254.20	412.0
255.00	48712.0
256.00	7126.0
257.00	650.0
258.00	3447.0
258.90	487.0
265.00	1369.0
266.00	249.0
271.00	153.0
272.00	165.0
272.90	1481.0
274.00	4117.0
275.00	21328.0
276.00	2497.0
277.00	2325.0
277.90	524.0
281.10	169.0
282.90	369.0
284.00	154.0
284.90	224.0
292.90	716.0
293.80	155.0
296.00	7810.0
297.00	1246.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016

303.00	1031.0
308.90	153.0
313.90	248.0
314.90	1048.0
316.00	538.0
321.00	276.0
321.90	156.0
323.10	1980.0
324.10	501.0
327.00	641.0
328.00	185.0
332.00	206.0
332.90	193.0
334.00	1598.0
335.00	474.0
341.00	190.0
346.00	640.0
351.90	581.0
353.00	414.0
354.10	593.0
365.00	4343.0
365.90	561.0
371.00	178.0
372.00	1209.0
373.00	272.0
383.00	288.0
391.10	210.0
401.90	498.0
403.00	825.0
421.00	613.0
421.90	433.0
423.00	3479.0
424.00	956.0
440.20	157.0
441.00	9701.0
442.00	60256.0
443.00	11955.0
444.00	1081.0

Instrument :
BNA_M
ClientSampleId :
DFTPP016