

Data Path : Z:\HPCHEM1\BNA_M\Data\BM090415\
 Data File : BM002669.D
 Acq On : 04 Sep 2015 11:50
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.470

Quant Time: Sep 04 13:34:38 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM90315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 04 04:35:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.78	152	8222	0.40	ng/ul	0.00
4) Naphthalene-d8	11.64	136	30333	0.40	ng/ul	0.00
8) Acenaphthene-d10	15.36	164	12261	0.40	ng/ul	0.00
12) Phenanthrene-d10	18.06	188	22614	0.40	ng/ul	-0.02
18) Chrysene-d12	22.16	240	21539	0.40	ng/ul	0.00
22) Perylene-d12	24.77	264	21068	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 1,4-Dioxane-d8	3.82	96	7796	0.94	ng/uL	0.00
6) 2-Methylnaphthalene-d10	13.17	152	14539	0.36	ng/ul	0.00
16) Fluoranthene-d10	20.03	212	20562	0.40	ng/ul	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) 1,4-Dioxane	3.87	88	8796	0.92	ng/ul	99
5) Naphthalene	11.69	128	29569	0.38	ng/ul	98
7) 2-Methylnaphthalene	13.23	142	18825	0.36	ng/ul	98
9) Acenaphthylene	15.09	152	24730	0.38	ng/ul	96
10) Acenaphthene	15.42	153	17144	0.37	ng/ul	94
11) Fluorene	16.39	166	17991	0.37	ng/ul	89
13) Pentachlorophenol	17.74	266	838	0.38	ng/ul	93
14) Phenanthrene	18.11	178	25350	0.37	ng/ul	98
15) Anthracene	18.20	178	25238	0.37	ng/ul	97
17) Fluoranthene	20.07	202	28231	0.40	ng/ul	91
19) Pyrene	20.42	202	29450	0.36	ng/ul	97
20) Benzo(a)anthracene	22.14	228	27173	0.38	ng/ul	94
21) Chrysene	22.19	228	25806	0.37	ng/ul	98
23) Benzo(b)fluoranthene	23.96	252	27036	0.36	ng/ul	99
24) Benzo(k)fluoranthene	24.01	252	27768	0.40	ng/ul	96
25) Benzo(a)pyrene	24.66	252	26421	0.38	ng/ul	96
26) Indeno(1,2,3-cd)pyrene	27.54	276	31060	0.38	ng/ul	99
27) Dibenzo(a,h)anthracene	27.54	278	25226	0.37	ng/ul	96
28) Benzo(g,h,i)perylene	28.40	276	26778	0.38	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM090415\
Data File : BM002669.D
Acq On : 04 Sep 2015 11:50
Operator : UM/IZ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.470

Quant Time: Sep 04 13:34:38 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM90315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 04 04:35:12 2015
Response via : Initial Calibration

