

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM090415\
 Data File : BM002678.D
 Acq On : 04 Sep 2015 17:23
 Operator : UM/IZ
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.471

Quant Time: Sep 04 17:55:23 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM90315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 04 04:35:12 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2 SURR	1,4-Dioxane-d8	0.403	0.343	14.9	94	0.00
3	1,4-Dioxane	0.465	0.386	17.0	95	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
5	Naphthalene	1.037	0.976	5.9	93	0.00
6 SURR	2-Methylnaphthalene-d10	0.537	0.496	7.6	92	0.00
7	2-Methylnaphthalene	0.691	0.641	7.2	92	-0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
9	Acenaphthylene	2.132	1.953	8.4	86	0.00
10 C	Acenaphthene	1.493	1.405	5.9	87	0.00
11	Fluorene	1.588	1.457	8.2	85	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	90	-0.02
13	Pentachlorophenol	0.039	0.033	15.4	92	0.00
14	Phenanthrene	1.197	1.129	5.7	85	0.00
15	Anthracene	1.191	1.106	7.1	84	0.00
16 SURR	Fluoranthene-d10	0.913	0.857	6.1	88	-0.01
17 C	Fluoranthene	1.252	1.158	7.5	89	-0.01
18 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
19	Pyrene	1.512	1.379	8.8	89	0.00
20	Benzo(a)anthracene	1.326	1.224	7.7	92	0.00
21	Chrysene	1.283	1.221	4.8	95	-0.01
22 I	Perylene-d12	1.000	1.000	0.0	99	-0.01
23	Benzo(b)fluoranthene	1.413	1.281	9.3	93	-0.01
24	Benzo(k)fluoranthene	1.326	1.306	1.5	96	-0.01
25 C	Benzo(a)pyrene	1.320	1.245	5.7	94	0.00
26	Indeno(1,2,3-cd)pyrene	1.563	1.468	6.1	93	0.00
27	Dibenzo(a,h)anthracene	1.280	1.189	7.1	92	-0.01
28	Benzo(g,h,i)perylene	1.339	1.267	5.4	94	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0