


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM090420\
 Data File : BM027322.D
 Acq On : 04 Sep 2020 08:21
 Operator : JU/CG
 Sample : DFTPP72
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP72

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090420MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Sep 08 10:18:28 2020



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2449

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.0	71461	PASS
68	69	0.00	2	1.9	2298	PASS
69	198	0.00	100	49.7	122554	PASS
70	69	0.00	2	0.5	619	PASS
127	198	10	80	45.3	111538	PASS
197	198	0.00	2	0.7	1707	PASS
198	198	100	100	100.0	246464	PASS
199	198	5	9	7.0	17248	PASS
275	198	10	60	30.9	76146	PASS
365	198	1	100	5.0	12361	PASS
441	198	0.01	100	12.3	30365	PASS
442	198	50	100	80.0	197090	PASS
443	442	15	24	19.6	38621	PASS

m/z	Abundance
40.00	397.0
43.90	931.0
48.80	181.0
73.10	654.0
91.10	206.0
166.90	166.0
191.00	275.0
193.00	154.0
207.00	1563.0
207.90	267.0
281.00	568.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

m/z	Abundance
37.00	399.0
38.10	835.0
39.00	6217.0
40.00	573.0
41.00	224.0
43.90	1148.0
49.00	651.0
50.00	19592.0
51.10	69992.0
52.00	3479.0
54.90	521.0
56.00	2600.0
57.00	5400.0
57.90	240.0
61.00	1187.0
62.00	1913.0
63.00	3657.0
64.00	712.0
65.00	2227.0
67.10	334.0
68.00	2156.0
69.00	121856.0
70.00	614.0
73.00	1604.0
74.00	14237.0
75.00	18856.0
76.00	6786.0
77.10	116440.0
78.10	7956.0
79.00	9559.0
80.00	7715.0
81.00	9408.0
82.00	2393.0
83.00	2027.0
83.90	316.0
85.00	1721.0
86.00	2675.0
87.00	1544.0
88.00	379.0
89.00	248.0
91.00	2042.0
92.00	2454.0
93.00	15905.0
93.90	887.0
94.90	183.0
96.00	691.0
97.20	217.0
98.00	12606.0
99.00	10411.0
100.00	827.0
101.00	5164.0
102.00	361.0
103.00	1483.0
104.00	3239.0
105.00	2912.0
106.10	1020.0
107.00	35920.0
108.00	4861.0
109.10	1314.0
110.00	56192.0
111.00	9114.0
111.90	1077.0
112.90	352.0
114.90	194.0
116.00	1915.0
117.00	37128.0
117.90	2512.0
118.90	461.0
120.00	563.0
121.90	1789.0
123.00	3485.0
124.00	1570.0
125.00	1239.0
125.90	402.0
127.00	103016.0
128.00	8377.0
129.00	44512.0
130.00	3927.0
131.00	734.0
132.00	543.0
133.00	188.0
134.00	1630.0
135.00	4398.0
136.00	1624.0
137.00	2413.0
138.00	550.0
139.10	340.0
140.00	575.0
141.00	6835.0
142.00	2257.0
143.00	1878.0
144.10	351.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

144.90	430.0
146.10	1332.0
147.00	3586.0
148.00	9358.0
149.10	1621.0
149.90	443.0
151.00	684.0
151.60	476.0
152.90	2115.0
153.90	1613.0
155.00	3764.0
156.00	4941.0
157.00	1000.0
157.90	1321.0
158.90	931.0
160.00	1853.0
161.00	2983.0
162.00	783.0
163.00	386.0
163.80	354.0
164.90	2563.0
166.00	1917.0
167.00	13289.0
168.00	9639.0
169.00	1297.0
169.80	564.0
171.00	480.0
171.90	1264.0
173.00	1469.0
174.00	2801.0
175.00	5183.0
176.00	1200.0
177.00	2691.0
178.00	1024.0
179.00	11036.0
180.00	6763.0
181.00	2902.0
182.00	490.0
184.00	785.0
184.90	4803.0
186.00	34552.0
187.00	10357.0
188.10	1149.0
188.90	2787.0
190.10	392.0
191.00	1220.0
192.00	3620.0
193.00	3523.0
193.90	1167.0
195.00	331.0
196.00	5563.0
197.00	1375.0
197.90	210432.0
199.00	15027.0
199.90	1366.0
200.80	574.0
201.60	498.0
202.10	518.0
203.00	1851.0
204.00	9848.0
205.00	15461.0
206.00	62360.0
207.00	8887.0
208.00	3069.0
209.10	917.0
210.00	1117.0
211.10	2632.0
212.10	426.0
212.90	243.0
214.90	763.0
216.00	1366.0
217.00	19968.0
218.00	2067.0
218.80	314.0
220.00	400.0
221.00	7160.0
221.90	2523.0
223.00	4027.0
224.00	34720.0
225.00	8724.0
226.00	1154.0
227.00	17704.0
228.00	2844.0
229.00	3193.0
230.00	333.0
230.90	1137.0
232.00	268.0
233.00	252.0
233.90	1077.0
234.90	1055.0
236.00	960.0
237.00	2274.0
238.00	247.0
238.90	717.0
239.90	788.0
240.90	979.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

242.00	1653.0
243.00	1995.0
244.00	25632.0
245.00	3489.0
246.00	6086.0
247.00	1329.0
248.00	238.0
249.00	1318.0
250.00	241.0
250.90	282.0
251.80	263.0
253.00	709.0
254.10	926.0
255.00	139584.0
256.00	21160.0
257.00	2453.0
258.00	10294.0
259.00	1566.0
259.90	161.0
260.90	276.0
262.90	157.0
263.90	181.0
265.00	3692.0
266.10	558.0
267.10	166.0
270.10	205.0
270.90	449.0
271.90	328.0
273.00	4255.0
274.00	10226.0
275.00	58728.0
276.00	8484.0
277.00	6615.0
278.00	1318.0
278.90	301.0
281.00	659.0
282.00	235.0
283.00	787.0
283.90	455.0
285.00	790.0
289.00	234.0
289.90	206.0
290.90	201.0
292.00	256.0
293.00	1301.0
293.90	485.0
294.90	387.0
296.00	19464.0
297.00	2402.0
298.00	263.0
300.80	258.0
302.10	364.0
303.00	2197.0
304.00	787.0
307.90	269.0
309.00	179.0
309.90	157.0
314.00	706.0
315.00	2231.0
316.00	1084.0
321.10	507.0
322.10	432.0
323.00	5615.0
324.00	1164.0
326.10	256.0
327.00	1239.0
327.80	555.0
332.00	378.0
333.00	488.0
334.10	3233.0
335.00	943.0
335.90	175.0
341.10	911.0
341.90	274.0
345.90	1397.0
346.80	373.0
351.00	176.0
352.00	1721.0
353.00	1095.0
354.00	1856.0
355.00	533.0
365.00	8237.0
366.00	1005.0
371.10	349.0
372.10	2475.0
372.90	482.0
383.10	673.0
384.00	319.0
389.90	255.0
391.00	390.0
391.90	157.0
401.90	1100.0
403.00	1689.0
404.00	533.0
421.00	1119.0
421.90	1121.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

423.10	7784.0
424.00	2596.0
425.00	459.0
441.10	18536.0
442.10	124200.0
443.00	24024.0
444.00	2026.0
445.00	193.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

m/z	Abundance
36.90	447.0
38.10	1244.0
39.10	6626.0
40.00	618.0
41.10	251.0
44.00	1217.0
45.10	200.0
49.00	879.0
50.00	24160.0
51.10	86576.0
52.00	4619.0
53.10	251.0
55.00	676.0
56.00	3326.0
57.00	7248.0
58.00	324.0
61.00	1499.0
62.10	1613.0
63.00	5262.0
63.90	560.0
65.10	2555.0
66.80	211.0
68.00	3074.0
69.00	145600.0
70.10	710.0
73.00	1685.0
74.00	17200.0
75.00	23880.0
76.00	7782.0
77.00	146304.0
78.00	9377.0
79.00	11580.0
80.00	8590.0
81.00	12064.0
82.00	2679.0
83.00	2209.0
83.80	262.0
85.00	1894.0
86.00	3284.0
87.00	1641.0
88.00	545.0
89.00	272.0
91.00	2876.0
92.00	2911.0
93.00	22632.0
93.90	1047.0
95.00	384.0
96.10	915.0
96.90	318.0
98.00	15715.0
99.00	12283.0
100.00	1068.0
101.00	6102.0
101.90	250.0
103.00	1707.0
104.00	3961.0
105.00	3934.0
106.00	1136.0
107.00	44896.0
108.00	6889.0
109.10	931.0
110.00	72288.0
111.00	11357.0
112.00	1709.0
113.00	439.0
114.90	247.0
116.00	2727.0
117.00	46512.0
118.10	3493.0
118.90	421.0
120.00	1053.0
121.00	363.0
122.00	3895.0
123.00	5568.0
123.90	2244.0
125.00	1723.0
126.20	898.0
127.00	134528.0
128.00	10248.0
129.00	56624.0
129.90	4451.0
131.00	1009.0
132.00	464.0
133.10	237.0
134.00	1833.0
135.00	4651.0
136.00	1642.0
137.00	2513.0
137.80	704.0
139.00	403.0
140.00	1024.0
141.00	8206.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

141.90	2637.0
143.00	1492.0
144.10	568.0
144.90	537.0
146.00	1415.0
147.00	4088.0
148.00	11302.0
149.00	2071.0
149.90	590.0
151.00	964.0
151.90	506.0
153.00	2817.0
154.00	1908.0
155.00	5353.0
156.00	7195.0
157.00	1220.0
157.90	1679.0
159.00	1441.0
160.00	2998.0
161.00	3816.0
161.90	1088.0
162.90	425.0
163.90	559.0
165.00	3457.0
166.00	2967.0
167.00	18688.0
168.00	11062.0
169.00	1956.0
169.90	579.0
171.00	738.0
171.90	2044.0
173.00	2194.0
174.00	3867.0
175.00	8007.0
176.00	1747.0
177.00	3578.0
177.90	934.0
178.90	13363.0
180.00	9068.0
181.00	4391.0
182.20	592.0
182.90	305.0
184.00	1238.0
185.00	6332.0
186.00	49208.0
187.00	14050.0
188.00	1791.0
189.00	3498.0
190.10	490.0
190.90	1936.0
192.00	4308.0
193.00	5748.0
194.00	747.0
195.20	463.0
196.00	8583.0
197.00	1977.0
198.00	298368.0
198.90	20536.0
200.00	2373.0
201.50	1044.0
203.00	2808.0
204.00	13137.0
205.00	22544.0
206.00	88600.0
207.00	13416.0
208.00	3867.0
209.00	1392.0
210.10	1763.0
211.10	3743.0
211.80	670.0
212.90	293.0
215.00	824.0
215.90	2446.0
217.00	26488.0
218.00	3059.0
219.00	340.0
219.90	559.0
221.00	10238.0
221.90	2903.0
223.00	5535.0
224.00	52848.0
225.00	13630.0
226.00	1635.0
227.00	25976.0
228.00	3246.0
229.00	4654.0
230.00	691.0
231.00	1954.0
232.10	333.0
232.90	381.0
233.90	1292.0
234.90	1598.0
236.00	1390.0
237.00	2056.0
237.90	320.0
238.90	1018.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

240.00	720.0
241.00	1323.0
242.00	2597.0
243.00	2912.0
244.00	35168.0
245.00	5060.0
246.00	8954.0
247.00	1712.0
248.00	297.0
248.90	1338.0
250.00	465.0
251.10	326.0
251.90	332.0
253.00	1111.0
254.10	1323.0
255.00	208512.0
256.00	30608.0
257.00	2472.0
258.00	15474.0
259.00	2654.0
260.00	395.0
260.90	277.0
262.90	150.0
263.90	312.0
265.00	6406.0
265.90	573.0
266.90	169.0
270.00	450.0
271.00	620.0
272.00	554.0
273.00	6264.0
274.00	16042.0
275.00	91616.0
276.00	12519.0
277.00	9532.0
277.90	1370.0
279.00	325.0
281.00	563.0
281.90	344.0
283.00	1131.0
284.10	605.0
285.10	1388.0
285.90	289.0
289.00	305.0
290.00	248.0
290.80	185.0
292.00	428.0
293.00	1510.0
294.00	799.0
295.10	551.0
296.00	31552.0
297.00	4590.0
298.10	309.0
300.90	513.0
301.90	438.0
303.00	3350.0
303.90	909.0
304.90	192.0
307.90	416.0
309.00	247.0
310.00	457.0
312.90	329.0
314.00	1394.0
315.00	3465.0
316.00	1560.0
317.00	384.0
320.00	313.0
321.00	1081.0
322.10	288.0
323.10	8175.0
324.10	1642.0
324.90	181.0
326.10	309.0
326.90	2001.0
328.00	939.0
329.00	156.0
332.00	944.0
333.10	729.0
334.00	5821.0
335.00	1990.0
336.00	311.0
339.10	186.0
341.00	1352.0
342.00	266.0
346.00	1941.0
346.90	497.0
350.00	156.0
350.90	246.0
352.10	3099.0
353.00	1888.0
354.00	2841.0
355.00	810.0
365.00	15267.0
366.00	2198.0
366.90	155.0
370.10	223.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

371.00	523.0
372.00	4763.0
373.00	1251.0
374.00	308.0
383.10	1079.0
384.00	422.0
390.00	771.0
390.90	296.0
392.10	353.0
401.00	223.0
402.00	1636.0
403.00	2386.0
404.00	931.0
405.00	271.0
421.00	1949.0
422.10	2118.0
423.10	14322.0
424.10	4471.0
425.10	571.0
440.10	211.0
441.00	37080.0
442.10	230272.0
443.10	46472.0
444.00	3989.0
444.90	163.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

m/z	Abundance
37.10	310.0
38.00	675.0
39.00	4384.0
40.00	348.0
41.10	164.0
44.00	1184.0
48.90	630.0
50.00	15813.0
51.00	57816.0
52.10	3277.0
55.10	437.0
55.90	2070.0
57.10	4174.0
58.00	287.0
61.00	838.0
61.90	1399.0
63.00	3357.0
64.10	511.0
65.00	1576.0
66.00	174.0
67.00	218.0
68.10	1665.0
69.00	100208.0
70.00	535.0
70.90	163.0
73.00	1582.0
74.00	11525.0
75.00	15158.0
76.10	5464.0
77.10	99720.0
78.10	6810.0
79.00	8319.0
80.00	5907.0
81.00	8182.0
81.90	2280.0
82.90	1548.0
83.80	394.0
85.00	1220.0
85.90	2379.0
87.00	1067.0
87.80	427.0
89.00	198.0
90.90	1998.0
91.90	2219.0
93.00	14709.0
94.00	899.0
95.00	333.0
96.10	717.0
97.10	209.0
98.00	11778.0
99.00	8355.0
100.00	777.0
100.90	3590.0
101.90	175.0
103.10	1145.0
104.00	2861.0
105.00	2524.0
106.00	904.0
107.00	33008.0
108.00	4595.0
109.10	967.0
110.00	51400.0
111.00	8108.0
112.00	962.0
113.00	293.0
116.00	1873.0
117.00	36504.0
118.00	2384.0
119.00	540.0
120.00	485.0
121.10	204.0
122.00	2656.0
123.00	3796.0
124.00	1582.0
125.00	1215.0
127.00	97072.0
128.00	6668.0
129.00	40976.0
130.00	3344.0
131.00	863.0
131.90	319.0
133.00	306.0
134.00	1324.0
135.00	3378.0
136.10	1589.0
137.00	2267.0
137.80	525.0
138.80	153.0
140.10	701.0
141.00	6726.0
142.00	1948.0
143.00	1189.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

144.20	280.0
144.80	361.0
146.00	1153.0
147.00	3593.0
148.00	9432.0
149.00	1788.0
150.00	530.0
151.00	662.0
151.80	449.0
153.00	1828.0
154.00	1564.0
155.00	3660.0
156.00	5058.0
157.10	918.0
158.10	1061.0
158.90	883.0
160.00	2309.0
161.00	3355.0
162.00	736.0
163.10	302.0
163.90	350.0
165.00	2934.0
166.00	1646.0
167.00	14155.0
168.00	7933.0
169.10	1297.0
169.90	480.0
171.00	703.0
171.90	1257.0
173.00	1489.0
174.00	3332.0
175.00	5184.0
176.00	1066.0
177.00	2302.0
178.00	1135.0
179.00	11321.0
180.00	7177.0
181.00	3178.0
182.00	559.0
182.80	226.0
184.00	911.0
185.00	5234.0
186.00	36776.0
187.00	10779.0
188.00	1328.0
188.90	2342.0
189.90	644.0
190.90	1326.0
192.00	3485.0
193.00	4482.0
194.00	817.0
195.10	523.0
196.00	6284.0
197.10	1769.0
198.00	230592.0
199.00	16182.0
199.90	1652.0
201.50	1014.0
202.30	225.0
203.00	2277.0
204.00	10839.0
205.00	18400.0
206.00	70832.0
207.00	10003.0
208.00	3399.0
209.00	1148.0
210.00	1281.0
211.00	3038.0
212.00	467.0
213.00	197.0
214.90	1092.0
216.00	1720.0
217.00	21568.0
218.00	3199.0
218.90	215.0
220.10	352.0
221.00	8579.0
222.00	2361.0
223.00	4533.0
224.00	40392.0
225.00	10779.0
226.10	1263.0
227.00	21792.0
228.00	2336.0
229.00	3921.0
229.90	421.0
231.00	1225.0
232.00	245.0
232.90	243.0
234.00	1091.0
234.90	1393.0
236.00	925.0
237.00	2030.0
237.90	223.0
239.00	898.0
240.00	517.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

241.00	1117.0
242.00	2385.0
243.00	2540.0
244.00	30608.0
245.10	4160.0
246.00	7425.0
247.00	1474.0
247.90	285.0
248.90	1139.0
250.90	394.0
251.90	363.0
253.00	1021.0
254.00	1009.0
255.00	178624.0
256.00	25648.0
257.10	1940.0
258.00	12373.0
259.00	1659.0
259.90	360.0
261.00	265.0
262.90	156.0
263.90	252.0
265.00	5351.0
266.00	902.0
267.00	164.0
270.00	241.0
271.00	542.0
271.90	553.0
273.00	5517.0
274.00	15096.0
275.00	78096.0
276.00	10447.0
277.00	8405.0
277.90	1471.0
279.00	318.0
281.10	449.0
281.90	332.0
282.90	901.0
284.00	713.0
285.10	1663.0
286.00	261.0
289.00	414.0
289.90	204.0
292.00	357.0
293.00	1804.0
294.00	440.0
295.10	514.0
296.00	30272.0
297.00	3290.0
298.00	260.0
300.90	358.0
302.10	438.0
303.00	2968.0
304.00	965.0
307.90	189.0
308.90	275.0
310.00	644.0
313.10	315.0
314.00	1480.0
315.00	3301.0
316.00	1897.0
317.00	518.0
319.90	164.0
320.90	819.0
321.90	622.0
323.00	8218.0
324.10	1333.0
324.90	185.0
327.00	1565.0
328.00	991.0
329.00	272.0
332.00	564.0
332.90	806.0
334.00	5665.0
335.00	1460.0
340.90	1148.0
342.10	231.0
346.00	1839.0
346.90	528.0
352.00	2224.0
353.00	1491.0
354.10	2112.0
355.00	630.0
365.00	13581.0
365.90	1943.0
370.00	306.0
371.10	559.0
372.00	4163.0
373.00	992.0
374.10	165.0
383.10	1549.0
384.00	526.0
389.90	673.0
390.90	480.0
392.10	256.0
401.10	308.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72

402.00	1630.0
403.00	2259.0
404.10	1118.0
405.00	279.0
421.00	1905.0
422.00	2139.0
423.00	13925.0
424.00	3967.0
425.00	482.0
441.10	35480.0
442.10	236800.0
443.10	45368.0
444.10	4068.0
445.00	161.0

Instrument :
BNA_M
ClientSampleId :
DFTPP72