




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090424\
Data File : BM047431.D
Acq On : 04 Sep 2024 09:52
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 04 11:08:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 26 18:11:40 2024
Response via : Initial Calibration



TIC: BM047431.D\data.ms

(70) Pentachlorophenol (C)




16.451min (+ 0.012) 119709.78 ng

response 714824

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	64.58
264.00	64.90	62.37
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090424\
Data File : BM047431.D
Acq On : 04 Sep 2024 09:52
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 04 11:08:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 26 18:11:40 2024
Response via : Initial Calibration



TIC: BM047431.D\data.ms

(77) Benzidine**19.069min (+ 0.001) 49134.61 ng****response 2723784**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.97
183.00	11.50	11.25
0.00	0.00	0.00