


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM090816\
 Data File : BM007457.D
 Acq On : 08 Sep 2016 13:20
 Operator : UM/SJ
 Sample : DFTPP57
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP57

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM082316.M
 Title : SVOA CALIBRATION
 Last Update : Fri Sep 09 03:02:55 2016



AutoFind: Scans 2552, 2553, 2554; Background Corrected with Scan 2547

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.9	96725	PASS
68	69	0.00	2	1.7	1916	PASS
69	198	0.00	100	35.0	109808	PASS
70	69	0.00	2	0.6	629	PASS
127	198	10	80	43.1	134960	PASS
197	198	0.00	2	0.5	1533	PASS
198	198	100	100	100.0	313450	PASS
199	198	5	9	6.7	21034	PASS
275	198	10	60	30.2	94589	PASS
365	198	1	100	4.3	13577	PASS
441	443	0.01	100	79.7	39968	PASS
442	198	50	100	84.6	265322	PASS
443	442	15	24	18.9	50144	PASS

m/z	Abundance
37.10	492.0
38.00	1335.0
39.00	7218.0
40.00	1190.0
44.00	1993.0
44.90	310.0
49.00	1133.0
50.10	25656.0
51.10	99048.0
52.00	5630.0
55.00	538.0
56.00	2771.0
57.00	7019.0
61.00	1196.0
62.00	1764.0
63.00	4303.0
64.00	514.0
65.00	2516.0
68.10	2064.0
69.00	109200.0
70.00	546.0
73.00	1533.0
74.00	13230.0
75.00	18400.0
76.00	6020.0
77.10	128672.0
78.10	8317.0
79.00	9414.0
80.00	6894.0
81.00	9181.0
82.00	2637.0
83.00	2238.0
85.00	1848.0
86.00	2887.0
87.10	1742.0
88.00	470.0
91.00	2217.0
92.00	2289.0
93.00	17016.0
94.00	955.0
96.00	907.0
97.20	458.0
98.00	13670.0
99.00	10945.0
100.00	639.0
101.00	4975.0
102.00	419.0
102.90	1879.0
103.90	3419.0
105.00	3376.0
105.90	929.0
107.00	41328.0
108.00	6492.0
109.00	1041.0
110.00	68576.0
111.00	10630.0
111.90	1548.0
113.00	652.0
114.90	305.0
116.00	2280.0
117.00	40240.0
118.00	3065.0
118.90	313.0
120.10	458.0
122.00	2911.0
123.00	3919.0
124.00	2169.0
125.00	1882.0
125.90	492.0
126.20	424.0
127.00	132800.0
128.00	10306.0
129.00	54400.0
130.00	4755.0
131.00	866.0
132.00	437.0
133.10	381.0
134.00	1531.0
135.00	4474.0
136.00	2203.0
137.10	2388.0
137.80	669.0
139.10	546.0
140.10	628.0
141.00	8400.0
142.00	2444.0
143.00	1726.0
143.90	429.0
144.90	402.0
146.10	1160.0
147.00	3939.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

148.00	9049.0
148.90	2068.0
150.00	487.0
150.90	742.0
151.50	416.0
151.90	461.0
153.00	2758.0
154.00	1922.0
155.00	4490.0
156.00	6195.0
157.20	932.0
158.00	1680.0
159.10	1126.0
160.00	2388.0
161.00	3723.0
162.10	1084.0
162.90	359.0
164.00	497.0
165.00	3418.0
166.00	2324.0
167.00	17328.0
168.00	8773.0
169.00	1675.0
169.90	833.0
171.00	703.0
172.00	1630.0
172.90	2014.0
174.00	3422.0
175.00	6243.0
176.00	1463.0
177.00	2924.0
178.00	810.0
179.00	10443.0
180.00	8077.0
181.00	3839.0
181.90	420.0
184.00	854.0
185.10	5578.0
186.00	43184.0
187.00	12957.0
188.00	1665.0
189.00	3321.0
190.00	442.0
191.10	1547.0
192.00	3696.0
193.00	4700.0
194.10	845.0
196.00	7182.0
197.00	1189.0
198.00	279360.0
199.00	18240.0
200.00	1431.0
201.00	516.0
201.60	1451.0
202.20	492.0
203.00	2300.0
204.10	11728.0
205.00	19160.0
206.00	80528.0
207.00	12544.0
208.00	3619.0
209.00	1521.0
210.00	1643.0
211.00	3274.0
212.00	667.0
215.00	888.0
216.00	1695.0
217.00	23264.0
218.00	3307.0
221.10	8254.0
222.00	2829.0
223.00	5093.0
224.00	45328.0
225.10	12306.0
226.10	1018.0
227.00	21520.0
228.00	2764.0
229.00	4857.0
230.00	513.0
231.00	1553.0
233.00	403.0
234.10	1117.0
235.00	1409.0
236.00	928.0
237.00	2162.0
238.90	829.0
239.90	711.0
240.90	1034.0
242.00	2579.0
243.00	2638.0
244.00	31888.0
245.00	4879.0
246.00	7276.0
247.00	1433.0
248.10	359.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

249.00	1507.0
250.00	338.0
251.00	321.0
252.00	454.0
253.00	650.0
254.10	1351.0
255.00	187904.0
256.00	25744.0
257.00	2302.0
258.00	11576.0
259.00	1901.0
260.00	434.0
265.00	5181.0
266.00	544.0
271.00	339.0
272.00	501.0
273.00	5144.0
274.00	14070.0
275.00	77744.0
276.00	10273.0
277.00	7800.0
278.00	1841.0
279.10	311.0
282.00	368.0
283.00	955.0
284.00	783.0
285.00	1688.0
289.10	387.0
289.90	582.0
292.00	375.0
293.00	1827.0
294.00	614.0
295.00	657.0
296.00	24248.0
297.00	3430.0
297.80	316.0
301.90	423.0
303.00	2544.0
304.00	529.0
308.00	334.0
309.00	529.0
314.00	1187.0
315.00	2896.0
316.00	1302.0
321.10	700.0
322.10	305.0
323.00	7086.0
324.00	1719.0
327.00	1514.0
327.90	1138.0
331.90	458.0
333.00	632.0
334.00	4257.0
335.00	1563.0
341.10	752.0
342.10	366.0
345.90	1483.0
352.00	2047.0
353.00	1407.0
354.00	1987.0
355.10	636.0
365.00	9792.0
366.00	1349.0
369.90	416.0
371.10	473.0
372.10	3054.0
373.10	1030.0
383.00	1206.0
390.10	402.0
402.00	996.0
403.00	1805.0
403.90	925.0
421.00	1398.0
422.00	1630.0
423.00	10202.0
424.00	2980.0
425.00	435.0
441.00	27360.0
442.00	178496.0
443.00	35000.0
444.10	3406.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
36.00	310.0
37.10	523.0
38.00	1646.0
39.00	8165.0
40.00	1422.0
40.90	335.0
44.00	1914.0
49.00	1088.0
50.10	29880.0
51.10	115792.0
52.00	6349.0
53.10	311.0
55.00	942.0
56.00	3568.0
57.00	7782.0
57.90	521.0
61.00	2020.0
62.10	1975.0
63.00	5496.0
64.10	654.0
65.00	2350.0
68.00	2225.0
69.00	132416.0
70.00	866.0
73.10	1454.0
74.00	14439.0
75.00	21200.0
76.00	7310.0
77.10	147264.0
78.10	10294.0
79.00	11359.0
80.00	8086.0
81.00	12007.0
82.00	2730.0
83.00	2470.0
83.90	325.0
85.10	2200.0
86.00	3429.0
87.00	1371.0
87.90	792.0
88.90	378.0
91.00	2537.0
92.00	3410.0
93.00	21704.0
94.00	1365.0
94.90	390.0
96.00	963.0
96.90	309.0
98.00	15921.0
99.00	12489.0
100.00	974.0
101.00	6905.0
103.00	1731.0
104.00	4861.0
105.00	4836.0
106.10	1045.0
107.00	51208.0
108.00	7879.0
109.00	1637.0
110.00	82880.0
111.00	14117.0
111.90	1613.0
112.90	692.0
116.00	2872.0
117.00	52008.0
118.00	3698.0
119.90	780.0
120.90	476.0
122.00	3166.0
123.00	5325.0
124.00	2594.0
125.00	2020.0
126.20	803.0
127.00	160576.0
128.00	11790.0
129.00	66680.0
130.00	5087.0
131.00	1071.0
131.90	553.0
134.00	2361.0
135.00	5799.0
136.00	2384.0
137.00	3103.0
137.90	521.0
139.00	460.0
140.00	1068.0
141.00	9723.0
142.00	3287.0
143.00	2162.0
144.00	477.0
144.80	528.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

146.00	1532.0
147.00	5648.0
148.00	12331.0
149.00	2546.0
150.00	540.0
150.90	1128.0
151.90	875.0
153.00	3257.0
154.00	1828.0
155.00	5452.0
156.10	7886.0
157.10	1349.0
158.00	1799.0
159.00	1399.0
160.00	2996.0
161.00	4507.0
162.10	1491.0
163.00	511.0
163.80	459.0
165.00	3956.0
166.00	2956.0
167.00	21816.0
168.00	10616.0
169.00	1475.0
170.00	750.0
171.00	884.0
172.00	1860.0
173.00	2103.0
174.00	4734.0
175.00	7501.0
176.00	1960.0
177.00	3747.0
178.00	1729.0
179.00	16362.0
180.00	10774.0
181.00	4641.0
181.90	816.0
183.10	430.0
184.00	1055.0
185.00	7259.0
186.00	56704.0
187.00	16023.0
188.00	2310.0
189.00	4435.0
190.00	736.0
191.00	2315.0
192.00	4686.0
193.00	6499.0
194.00	1309.0
194.90	531.0
196.00	9669.0
197.10	2031.0
198.00	372032.0
199.00	25200.0
200.00	2050.0
201.50	1712.0
203.00	2769.0
204.00	15539.0
205.10	26032.0
206.00	113712.0
207.00	14395.0
208.10	3967.0
209.00	1505.0
210.10	1655.0
211.00	4618.0
211.80	741.0
213.10	412.0
215.00	1657.0
216.00	3023.0
217.00	31688.0
218.00	4150.0
221.00	11852.0
221.90	3530.0
223.00	7168.0
224.00	61456.0
225.00	15866.0
226.00	1776.0
227.00	30184.0
228.00	4157.0
229.00	6022.0
229.90	894.0
231.00	2172.0
232.00	473.0
233.00	440.0
234.10	1415.0
235.00	2084.0
235.90	1538.0
237.10	2733.0
238.00	441.0
239.00	1263.0
240.00	1016.0
241.00	1572.0
242.00	3192.0
243.00	2999.0
244.00	46864.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

245.10	6397.0
246.00	10403.0
247.00	2015.0
247.90	835.0
249.00	1787.0
249.90	323.0
250.90	546.0
252.10	537.0
253.00	1297.0
254.10	1509.0
255.00	264384.0
256.00	37728.0
257.00	2657.0
258.00	17384.0
258.90	2391.0
261.00	555.0
264.00	446.0
265.00	7444.0
266.00	1032.0
270.00	388.0
271.10	652.0
272.00	676.0
273.00	7529.0
274.00	20888.0
275.00	111496.0
276.00	14806.0
277.00	10525.0
278.00	2360.0
279.10	490.0
280.90	349.0
281.90	337.0
283.00	1285.0
283.90	771.0
285.00	1675.0
286.20	314.0
289.00	577.0
289.90	476.0
292.00	537.0
293.00	2731.0
293.90	638.0
294.90	619.0
296.00	37904.0
297.00	5194.0
297.90	357.0
301.00	405.0
302.10	626.0
303.00	4187.0
304.00	1248.0
307.90	505.0
309.10	307.0
310.00	317.0
314.00	1296.0
315.00	4198.0
316.00	2485.0
317.10	502.0
321.10	972.0
322.00	729.0
323.10	9883.0
324.00	2370.0
326.90	2125.0
327.90	1248.0
331.90	618.0
333.00	1257.0
334.00	7642.0
335.00	1777.0
340.90	1139.0
341.90	540.0
346.00	2610.0
347.00	340.0
352.00	2921.0
353.00	1920.0
354.10	3821.0
355.00	761.0
358.80	301.0
365.00	16800.0
366.00	2608.0
366.90	357.0
369.90	308.0
371.10	553.0
372.00	5284.0
373.00	1420.0
383.00	1405.0
383.90	535.0
389.90	600.0
391.10	406.0
392.10	379.0
402.00	1567.0
402.90	2529.0
404.00	1099.0
405.10	547.0
421.00	2284.0
422.00	2229.0
423.00	17120.0
424.10	4393.0
424.90	430.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

441.00	47296.0
442.10	311168.0
443.10	59744.0
444.10	5957.0
445.10	411.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
38.00	812.0
39.10	5267.0
40.00	1614.0
44.00	1533.0
49.10	600.0
50.00	20584.0
51.10	75336.0
52.10	4013.0
55.00	662.0
56.00	2420.0
57.00	4870.0
58.00	331.0
61.00	1278.0
62.00	1021.0
63.00	3852.0
64.00	496.0
65.00	1853.0
68.00	1460.0
69.00	87808.0
70.00	477.0
73.00	1169.0
74.00	10158.0
75.00	15265.0
76.00	4845.0
77.10	102512.0
78.10	6405.0
79.00	7409.0
80.00	5192.0
81.00	7928.0
82.00	1779.0
83.10	1392.0
83.90	429.0
85.00	1436.0
86.10	2044.0
87.00	951.0
87.90	358.0
91.00	1949.0
92.00	1935.0
93.00	14654.0
94.10	885.0
94.90	406.0
96.10	545.0
98.00	11444.0
99.00	9091.0
100.00	770.0
101.00	4292.0
103.00	1697.0
104.00	2512.0
105.10	2651.0
106.00	703.0
107.00	35048.0
108.00	5243.0
109.00	853.0
110.00	57448.0
111.00	8899.0
112.00	1029.0
113.10	419.0
116.00	1915.0
117.00	35400.0
118.00	2624.0
119.00	400.0
120.00	446.0
122.00	2406.0
123.00	3550.0
123.90	1588.0
125.00	1461.0
126.00	501.0
127.00	111504.0
128.00	8407.0
129.00	47512.0
130.00	3866.0
131.00	514.0
131.90	403.0
134.00	1239.0
135.00	3753.0
136.00	1866.0
137.00	2327.0
138.00	470.0
140.00	672.0
141.00	6977.0
142.10	2002.0
143.00	1471.0
144.10	536.0
144.90	341.0
146.00	1430.0
147.00	3953.0
148.00	8910.0
149.00	1688.0
149.90	468.0
150.80	533.0
151.30	764.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

151.90	405.0
152.90	2087.0
154.00	1418.0
155.00	3383.0
156.00	5621.0
156.90	795.0
157.20	732.0
158.00	1011.0
159.00	1026.0
160.00	2428.0
161.00	3480.0
162.00	1223.0
165.00	2480.0
165.90	2256.0
167.00	15492.0
168.00	8036.0
169.00	861.0
170.00	491.0
170.90	705.0
172.00	1475.0
173.00	2176.0
174.00	3116.0
175.00	5541.0
176.00	1527.0
177.00	2972.0
178.00	936.0
179.00	12026.0
180.00	8714.0
181.00	3578.0
181.90	676.0
182.80	368.0
184.00	882.0
185.00	5380.0
186.00	43544.0
187.00	12319.0
188.00	1371.0
189.00	2784.0
190.00	387.0
191.00	1250.0
192.00	3709.0
193.00	4327.0
194.00	1171.0
196.00	8180.0
197.00	1380.0
198.00	288960.0
199.00	19664.0
200.00	1790.0
200.70	343.0
201.40	1018.0
203.00	2628.0
204.00	11579.0
205.00	18800.0
206.10	85568.0
207.00	13151.0
208.00	3566.0
209.00	1467.0
210.10	1151.0
211.00	3334.0
211.90	674.0
213.10	301.0
215.00	1022.0
216.00	2013.0
217.00	24888.0
218.00	2834.0
219.80	369.0
221.00	10501.0
222.00	2750.0
223.00	5446.0
224.00	49064.0
225.00	12501.0
226.00	1154.0
227.00	23608.0
228.00	3496.0
229.00	4994.0
230.10	552.0
231.00	1973.0
231.90	310.0
233.00	477.0
234.00	1685.0
235.00	1765.0
235.90	1078.0
237.00	1947.0
238.00	404.0
238.90	640.0
239.90	692.0
240.90	1154.0
242.00	2829.0
243.00	2533.0
244.10	39736.0
245.00	4937.0
246.00	8712.0
247.00	1558.0
248.90	1530.0
250.00	321.0
250.90	336.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

252.10	448.0
252.90	876.0
254.00	1357.0
255.00	215040.0
256.00	30032.0
257.00	2227.0
258.00	14239.0
259.00	1748.0
260.90	493.0
263.90	379.0
265.00	6614.0
266.00	678.0
269.90	372.0
270.90	491.0
272.00	386.0
273.00	6762.0
274.00	15921.0
275.00	94528.0
276.00	11817.0
277.00	9361.0
278.00	1658.0
281.10	356.0
283.10	878.0
284.00	690.0
285.00	1826.0
288.90	560.0
292.00	375.0
293.00	2169.0
293.90	472.0
295.10	710.0
296.00	30936.0
297.00	4756.0
301.00	428.0
302.00	304.0
303.00	3210.0
304.00	958.0
308.00	488.0
308.90	316.0
310.00	414.0
314.00	1478.0
315.00	3331.0
316.00	1698.0
320.90	1111.0
322.00	485.0
323.10	9056.0
324.10	1486.0
327.00	2148.0
328.00	1322.0
332.00	753.0
333.00	752.0
334.00	6179.0
335.10	1460.0
341.00	1113.0
341.90	553.0
346.00	1742.0
347.00	370.0
352.00	2969.0
353.00	1758.0
354.10	2988.0
355.10	698.0
365.00	14139.0
366.00	2147.0
371.10	816.0
372.00	4864.0
373.10	1359.0
382.90	1266.0
384.10	439.0
390.00	474.0
391.00	471.0
391.90	331.0
401.10	358.0
402.00	1693.0
403.10	3246.0
404.00	1318.0
421.10	2779.0
422.00	2086.0
423.00	16341.0
424.00	3475.0
425.10	436.0
439.90	303.0
441.00	45248.0
442.00	306304.0
443.10	55688.0
444.10	5770.0
445.00	312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

Scan 2547 (17.568 min): BM007457.D
DFTPP57

m/z	Abundance
36.00	330.0
40.10	1085.0
41.00	307.0
44.00	1931.0
73.00	551.0
191.20	334.0
207.00	2001.0
208.00	341.0
280.90	512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57