

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090820\
 Data File : BM027351.D
 Acq On : 08 Sep 2020 10:23
 Operator : JU/CG
 Sample : PB131455BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK55

Quant Time: Sep 08 11:49:51 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 04 14:46:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	89699	20.00	ng/ul	0.00
18) Naphthalene-d8	10.34	136	357183	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.21	164	273690	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.96	188	665899	20.00	ng/ul	0.00
78) Chrysene-d12	21.16	240	778204	20.00	ng/ul	0.00
86) Perylene-d12	23.33	264	838436	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.13	96	13800	6.85	ng/uL	0.00
5) Phenol-d5	6.75	99	197189	26.67	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.91	67	143068	29.14	ng/ul	0.00
9) 2-Chlorophenol-d4	7.11	132	163590	30.21	ng/ul	0.00
13) 4-Methylphenol-d8	8.27	113	161235	26.74	ng/ul	0.00
19) Nitrobenzene-d5	8.71	128	82517	30.69	ng/ul	0.00
22) 2-Nitrophenol-d4	9.43	143	92650	31.42	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.96	165	178251	28.09	ng/ul	0.00
29) 4-Chloroaniline-d4	10.47	131	229996	30.66	ng/ul	0.00
44) Dimethylphthalate-d6	13.63	166	691615	31.76	ng/ul	0.00
47) Acenaphthylene-d8	13.90	160	786878	31.99	ng/ul	0.00
52) 4-Nitrophenol-d4	14.42	143	88762	23.15	ng/ul	0.00
58) Fluorene-d10	15.21	176	592804	30.39	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.33	200	104438	24.51	ng/ul	0.00
71) Anthracene-d10	17.06	188	1011549	32.12	ng/ul	0.00
79) Pyrene-d10	19.36	212	1219587	35.18	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.19	264	1356568	30.16	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090820\
Data File : BM027351.D
Acq On : 08 Sep 2020 10:23
Operator : JU/CG
Sample : PB131455BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK55

Quant Time: Sep 08 11:49:51 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 04 14:46:56 2020
Response via : Initial Calibration

