


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM090916\
Data File : BM007467.D
Acq On : 09 Sep 2016 12:34
Operator : UM/SJ
Sample : DFTPP58
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP58

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM071416.M
Title : SVOA CALIBRATION
Last Update : Wed Jul 20 04:22:37 2016



AutoFind: Scans 2552, 2553, 2554; Background Corrected with Scan 2546

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.2	83562	PASS
68	69	0.00	2	1.8	1710	PASS
69	198	0.00	100	33.3	95170	PASS
70	69	0.00	2	0.4	341	PASS
127	198	10	80	41.0	117117	PASS
197	198	0.00	2	0.4	1109	PASS
198	198	100	100	100.0	285717	PASS
199	198	5	9	6.7	19074	PASS
275	198	10	60	30.7	87701	PASS
365	198	1	100	4.4	12554	PASS
441	443	0.01	100	78.1	37632	PASS
442	198	50	100	89.4	255381	PASS
443	442	15	24	18.9	48168	PASS

m/z	Abundance
37.10	475.0
38.00	1430.0
39.10	6223.0
39.90	963.0
41.10	329.0
44.00	1171.0
49.10	871.0
50.00	28216.0
51.10	106120.0
52.10	5222.0
55.00	647.0
56.00	3397.0
57.00	7334.0
57.90	318.0
61.00	1555.0
62.00	1885.0
63.00	4480.0
64.00	784.0
65.10	2300.0
68.00	2095.0
69.00	118736.0
70.00	637.0
73.00	1051.0
74.00	13723.0
75.00	21560.0
76.10	7259.0
77.10	136704.0
78.10	8694.0
79.00	9959.0
80.00	7278.0
81.00	11616.0
82.00	3003.0
83.00	2347.0
83.90	501.0
85.00	1841.0
86.00	3489.0
86.90	1094.0
88.00	486.0
91.00	2177.0
92.00	3088.0
93.00	17960.0
94.00	1459.0
96.00	827.0
97.20	435.0
98.00	14409.0
99.00	11943.0
100.00	1138.0
101.00	5518.0
101.90	369.0
103.00	1777.0
104.00	3728.0
105.00	3567.0
106.00	1127.0
107.00	45128.0
108.00	6928.0
109.00	1150.0
110.00	75992.0
111.00	12927.0
112.00	1508.0
113.00	582.0
116.00	2567.0
117.00	45664.0
118.00	3387.0
119.00	448.0
119.90	624.0
122.00	3279.0
123.00	4219.0
124.00	2035.0
125.00	1605.0
126.10	743.0
127.00	139328.0
128.00	10447.0
129.00	57400.0
130.00	4912.0
131.00	1120.0
131.90	500.0
132.90	379.0
134.00	1677.0
135.00	4602.0
136.00	2151.0
137.00	2728.0
137.90	660.0
140.00	852.0
141.00	8745.0
142.00	2996.0
143.00	2124.0
144.00	613.0
145.00	610.0
146.00	1620.0
147.00	4872.0
148.00	10987.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

149.00	2528.0
150.00	753.0
151.00	948.0
152.00	626.0
153.00	2514.0
154.00	1777.0
155.00	5452.0
156.00	7332.0
157.00	1181.0
158.00	1431.0
159.10	978.0
160.00	2730.0
161.00	4047.0
162.00	1444.0
162.90	303.0
163.90	633.0
165.00	3280.0
166.00	2318.0
167.00	18992.0
168.00	9207.0
169.00	1334.0
169.90	485.0
170.90	958.0
172.00	1292.0
173.10	1980.0
174.00	3637.0
175.00	7911.0
176.00	1525.0
177.00	2992.0
178.00	1304.0
179.00	13156.0
180.00	8605.0
181.00	3810.0
182.00	741.0
183.10	462.0
184.00	1206.0
185.10	6478.0
186.00	47384.0
187.00	14875.0
188.00	1412.0
189.00	3711.0
190.00	792.0
191.00	1969.0
192.00	3934.0
193.00	5084.0
194.00	830.0
195.00	494.0
196.00	8811.0
196.80	1152.0
197.10	1414.0
198.00	321472.0
199.00	21968.0
200.00	1933.0
201.60	1299.0
203.00	2816.0
204.00	13760.0
205.00	22520.0
206.00	91960.0
207.00	12967.0
208.00	3397.0
209.00	1004.0
210.00	1298.0
211.00	3394.0
211.90	649.0
212.90	521.0
215.10	1404.0
216.00	2140.0
217.00	26592.0
218.00	2945.0
219.20	357.0
221.00	11820.0
222.00	2997.0
223.00	5415.0
224.00	52792.0
225.10	13257.0
226.10	1457.0
227.00	23304.0
228.00	3370.0
229.00	4617.0
230.10	584.0
231.00	2396.0
232.00	318.0
233.00	432.0
234.00	1668.0
235.00	1675.0
236.00	1231.0
237.10	1507.0
239.00	1041.0
240.10	979.0
240.90	1411.0
242.00	2773.0
243.00	2729.0
244.10	36616.0
245.00	5137.0
246.00	8305.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

246.90	1567.0
248.00	642.0
249.00	1328.0
250.10	303.0
252.00	469.0
253.00	890.0
254.10	1316.0
255.00	216448.0
256.00	30176.0
257.10	2815.0
258.00	12956.0
259.00	2244.0
260.00	531.0
261.10	411.0
263.80	390.0
265.00	5723.0
265.90	783.0
271.10	539.0
272.00	561.0
273.00	6002.0
274.00	16656.0
275.00	91616.0
276.00	13368.0
277.00	9017.0
278.00	1802.0
278.90	318.0
281.00	372.0
282.10	338.0
282.90	1059.0
284.00	838.0
285.10	1298.0
286.00	447.0
291.90	599.0
293.00	2026.0
294.10	422.0
295.10	550.0
296.00	29504.0
297.00	4189.0
301.00	405.0
302.00	399.0
303.10	2411.0
304.10	697.0
308.90	451.0
310.00	339.0
313.90	1424.0
314.90	2902.0
316.00	1425.0
316.90	357.0
321.00	794.0
321.90	589.0
323.00	8260.0
324.00	1077.0
327.00	1297.0
328.00	637.0
331.90	499.0
333.10	504.0
334.00	5285.0
335.00	1311.0
341.10	999.0
342.10	318.0
346.00	1647.0
347.00	397.0
352.10	2485.0
353.00	1507.0
354.00	2314.0
355.10	516.0
365.00	11450.0
365.90	2496.0
370.90	460.0
372.00	4029.0
373.10	1095.0
383.00	1280.0
389.90	422.0
391.00	626.0
402.00	1757.0
402.90	2442.0
404.00	950.0
421.00	1865.0
422.00	1781.0
423.00	13769.0
424.00	3365.0
425.00	562.0
441.00	33400.0
442.00	227072.0
443.00	44576.0
444.10	4360.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
38.10	1062.0
39.10	6207.0
40.00	837.0
40.90	316.0
44.00	1575.0
49.10	708.0
50.10	22544.0
51.10	96008.0
52.00	4119.0
54.90	537.0
56.00	2824.0
57.00	6766.0
57.90	332.0
61.00	1600.0
62.00	1942.0
63.00	4254.0
63.90	630.0
65.00	1897.0
68.00	2013.0
69.00	108448.0
70.00	388.0
73.00	1396.0
74.00	12446.0
75.00	18240.0
76.00	5889.0
77.10	123168.0
78.10	7175.0
79.00	8605.0
80.00	6922.0
81.00	9680.0
82.00	2074.0
83.10	2077.0
85.00	1956.0
86.00	2799.0
86.90	1322.0
88.00	1043.0
91.00	2241.0
92.00	2331.0
93.00	17832.0
94.00	1088.0
95.90	652.0
98.00	14188.0
99.00	10527.0
99.90	881.0
101.10	5534.0
102.10	367.0
103.00	1885.0
104.00	4002.0
105.00	3399.0
106.00	1070.0
107.00	43872.0
108.00	6064.0
109.00	1437.0
110.00	74104.0
111.00	11570.0
112.00	1760.0
112.90	429.0
115.90	2160.0
117.00	42152.0
118.00	2524.0
119.00	349.0
120.00	485.0
122.00	3172.0
123.00	5300.0
124.00	2324.0
124.90	1719.0
126.20	721.0
127.00	136704.0
128.00	10484.0
129.00	54952.0
130.00	5094.0
131.00	838.0
133.00	484.0
134.00	1493.0
135.00	4898.0
136.10	1935.0
137.10	2474.0
137.80	582.0
138.90	362.0
139.90	877.0
141.00	8061.0
142.00	2415.0
143.00	2086.0
144.10	572.0
145.10	432.0
146.00	1743.0
147.00	4818.0
148.00	9676.0
149.00	2201.0
150.00	517.0
151.00	1001.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

152.00	773.0
153.00	2929.0
154.10	1846.0
155.00	4790.0
156.00	5898.0
157.00	1210.0
158.00	1289.0
159.00	1602.0
160.00	3091.0
161.00	3849.0
162.00	933.0
164.10	396.0
165.00	3189.0
166.00	2672.0
167.00	17560.0
168.00	10037.0
169.00	1559.0
170.00	409.0
171.00	919.0
172.00	1344.0
173.00	2069.0
174.00	4342.0
175.00	6688.0
176.00	1856.0
177.00	3328.0
178.00	976.0
179.00	13401.0
180.00	9653.0
181.00	3981.0
182.00	612.0
183.00	439.0
184.00	1178.0
185.00	6843.0
186.00	50504.0
187.00	13586.0
188.10	1588.0
189.00	3955.0
189.90	732.0
191.00	1682.0
192.00	4364.0
193.10	4780.0
193.90	738.0
194.90	442.0
196.00	8971.0
197.00	1914.0
198.00	331136.0
199.00	23016.0
200.00	1912.0
201.40	1111.0
203.00	2703.0
204.00	13414.0
205.00	23472.0
206.00	94304.0
207.10	12115.0
208.00	4143.0
209.00	1494.0
210.00	1647.0
211.00	3725.0
211.90	707.0
215.00	1622.0
216.00	2620.0
217.00	26880.0
218.00	3435.0
219.00	502.0
220.20	687.0
221.00	10417.0
221.90	3347.0
223.00	5612.0
224.00	57640.0
225.10	14268.0
226.00	1798.0
227.00	26776.0
228.00	3988.0
229.00	4112.0
230.10	866.0
231.00	1920.0
232.90	455.0
234.00	1812.0
235.00	1970.0
236.00	1115.0
237.10	2738.0
238.00	487.0
238.90	818.0
239.90	783.0
241.00	1781.0
242.00	3303.0
243.00	3939.0
244.10	43128.0
245.10	5733.0
246.00	9785.0
247.00	1780.0
247.90	428.0
249.00	1724.0
250.00	430.0
252.00	340.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

253.10	834.0
254.10	1041.0
255.00	235776.0
256.00	33360.0
257.00	2675.0
258.00	15445.0
259.00	2279.0
259.80	399.0
260.80	333.0
265.00	6559.0
265.90	770.0
271.00	467.0
272.00	618.0
273.00	8042.0
274.00	18496.0
275.00	104560.0
276.00	13546.0
277.00	10389.0
278.00	1958.0
279.00	365.0
281.00	420.0
282.00	393.0
283.10	857.0
284.00	716.0
285.00	2200.0
289.00	381.0
290.00	306.0
292.00	596.0
293.00	2673.0
293.90	534.0
295.10	637.0
296.00	32120.0
297.00	4597.0
300.90	532.0
302.00	508.0
303.00	4311.0
304.00	1159.0
307.90	342.0
309.90	573.0
312.90	377.0
314.00	1992.0
315.00	4040.0
316.00	2211.0
316.90	318.0
321.10	1137.0
322.00	594.0
323.10	10396.0
324.10	2103.0
327.00	2060.0
328.00	895.0
331.90	623.0
333.00	939.0
334.00	6445.0
335.00	1574.0
341.00	993.0
346.00	1920.0
346.90	407.0
352.00	3252.0
353.00	1849.0
354.00	3050.0
355.10	838.0
365.00	15208.0
366.00	2074.0
371.00	812.0
372.00	4769.0
373.10	1149.0
383.00	1361.0
383.80	439.0
390.00	416.0
401.00	442.0
402.10	1572.0
403.00	2541.0
404.00	1110.0
421.00	2335.0
422.00	2668.0
423.10	16100.0
424.00	3937.0
425.00	506.0
441.10	44888.0
442.10	308352.0
443.10	57656.0
444.00	5907.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
38.00	812.0
39.00	3376.0
40.00	803.0
40.90	341.0
44.00	1666.0
49.10	534.0
50.10	13587.0
51.10	48560.0
52.10	2618.0
55.00	440.0
56.00	1261.0
57.00	4086.0
61.00	775.0
62.00	819.0
63.10	1941.0
64.00	342.0
65.10	1119.0
68.00	1024.0
69.00	58328.0
72.90	555.0
74.00	6031.0
75.00	10478.0
76.00	3113.0
77.10	66320.0
78.10	4904.0
79.00	5094.0
80.00	3447.0
81.00	5103.0
82.00	1190.0
83.00	1115.0
84.90	1129.0
85.90	1361.0
87.00	969.0
91.00	1384.0
92.00	1572.0
93.00	9546.0
94.10	395.0
96.20	530.0
98.00	7006.0
99.00	5209.0
100.00	732.0
101.00	2831.0
103.20	1034.0
104.00	1995.0
105.00	2016.0
106.00	746.0
107.00	26352.0
108.00	3159.0
109.10	523.0
110.00	40392.0
111.00	5797.0
112.10	657.0
113.00	511.0
116.10	1466.0
117.00	23344.0
117.90	1488.0
122.10	1589.0
123.00	2985.0
124.00	1279.0
124.90	1149.0
126.20	556.0
127.00	75320.0
128.00	5580.0
129.00	31704.0
130.00	2334.0
130.90	609.0
132.00	306.0
134.00	869.0
135.00	2890.0
136.10	1130.0
137.10	1599.0
137.90	321.0
140.10	400.0
141.00	4118.0
142.00	1345.0
143.00	895.0
145.90	1073.0
147.00	2125.0
148.00	5692.0
149.00	1447.0
150.00	318.0
150.90	577.0
152.90	1754.0
153.90	1187.0
155.00	3135.0
156.00	4060.0
157.10	822.0
157.90	840.0
159.00	429.0
160.00	1630.0
160.90	1725.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

161.90	817.0
163.00	451.0
165.00	2580.0
166.00	1972.0
167.00	10632.0
168.00	5080.0
168.90	998.0
170.00	448.0
170.90	547.0
172.00	1118.0
173.00	1097.0
174.00	2459.0
175.00	3875.0
176.10	962.0
177.00	2061.0
177.90	423.0
179.00	7470.0
180.00	5510.0
181.00	2801.0
182.10	381.0
184.00	582.0
185.00	3743.0
186.00	29096.0
187.00	8195.0
188.00	684.0
189.00	2081.0
191.00	1215.0
192.00	2675.0
193.00	2491.0
194.10	702.0
194.90	414.0
196.00	5657.0
196.80	794.0
198.00	204544.0
199.00	12239.0
200.00	1529.0
201.30	1147.0
203.00	1898.0
204.00	8167.0
205.10	13036.0
206.00	57544.0
207.00	8877.0
208.00	2090.0
208.90	532.0
210.00	1531.0
211.00	2362.0
211.80	396.0
213.00	360.0
215.00	792.0
216.00	1290.0
217.00	16440.0
218.00	2321.0
219.00	407.0
221.00	7954.0
221.90	1739.0
223.00	4299.0
224.00	35840.0
225.00	8614.0
226.00	1354.0
227.00	16069.0
228.00	2499.0
229.00	3103.0
231.00	1121.0
233.00	320.0
234.00	1001.0
235.00	1025.0
235.90	580.0
237.00	1299.0
237.90	349.0
239.00	630.0
240.00	494.0
241.00	769.0
242.00	2148.0
243.00	1606.0
244.10	27360.0
245.00	3500.0
246.00	6305.0
247.00	1409.0
249.00	1240.0
252.90	454.0
254.20	1308.0
255.00	155456.0
256.00	21168.0
256.90	1725.0
258.00	9151.0
258.90	1271.0
264.00	381.0
265.00	3775.0
265.90	514.0
271.00	411.0
272.00	511.0
273.00	5046.0
274.00	12872.0
275.00	66928.0
276.00	8013.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

277.00	6363.0
277.90	1612.0
281.00	562.0
283.00	894.0
284.10	627.0
285.10	799.0
289.00	361.0
292.00	427.0
293.00	1868.0
294.00	419.0
295.10	680.0
296.00	23752.0
297.00	3142.0
301.00	485.0
301.90	345.0
303.00	2685.0
304.10	706.0
310.00	412.0
313.90	874.0
315.00	2616.0
316.00	1067.0
321.00	529.0
322.10	339.0
323.10	6672.0
324.00	1404.0
327.10	1021.0
328.10	823.0
331.90	705.0
333.10	539.0
334.10	4428.0
335.00	921.0
341.00	935.0
346.00	1588.0
346.80	367.0
352.10	1946.0
353.00	1234.0
354.10	2357.0
355.00	556.0
365.00	11006.0
365.90	1724.0
370.90	431.0
372.10	3939.0
373.10	939.0
383.00	929.0
390.00	328.0
391.00	355.0
402.00	1828.0
403.00	2185.0
404.00	1011.0
421.00	1762.0
422.10	1492.0
423.00	11995.0
424.00	2690.0
441.10	34608.0
442.10	230720.0
443.10	42272.0
444.10	4175.0
445.10	406.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

Scan 2546 (17.562 min): BM007467.D
DFTPP58

m/z	Abundance
40.00	560.0
44.00	1677.0
207.00	1313.0
281.00	311.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58