

Data Path : Z:\HPCHEM1\BNA_M\Data\BM091117\
 Data File : BM011508.D
 Acq On : 11 Sep 2017 22:04
 Operator : SJ/JU
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02013

Quant Time: Sep 12 07:12:32 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 02:59:45 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.487	0.483	0.8	72	0.00
3 S	1,4-Dioxane-d8	0.445	0.449	-0.9	73	0.00
4	Benzaldehyde	1.097	0.988	9.9	61	0.00
5 S	Phenol-d5	1.920	1.754	8.6	68	0.00
6	Phenol	1.906	1.814	4.8	71	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.294	1.307	-1.0	73	0.00
8	Bis(2-Chloroethyl)ether	1.501	1.461	2.7	71	0.00
9 S	2-Chlorophenol-d4	1.447	1.381	4.6	69	0.00
10	2-Chlorophenol	1.412	1.349	4.5	69	0.00
11	2-Methylphenol	1.445	1.358	6.0	70	0.00
12	2,2'-oxybis(1-Chloropropane	2.433	2.671	-9.8	79	0.00
13 S	4-Methylphenol-d8	1.499	1.436	4.2	71	0.00
14	Acetophenone	2.292	2.233	2.6	71	0.00
15 P	N-Nitroso-di-n-propylamine	1.230	1.223	0.6	72	0.00
16	4-Methylphenol	1.582	1.540	2.7	72	0.00
17	Hexachloroethane	0.540	0.533	1.3	73	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
19 S	Nitrobenzene-d5	0.152	0.145	4.6	70	0.00
20	Nitrobenzene	0.357	0.355	0.6	73	0.00
21	Isophorone	0.719	0.696	3.2	71	0.00
22 S	2-Nitrophenol-d4	0.158	0.143	9.5	67	0.00
23 C	2-Nitrophenol	0.166	0.155	6.6	68	0.00
24	2,4-Dimethylphenol	0.355	0.348	2.0	71	0.00
25	Bis(2-Chloroethoxy)methane	0.469	0.450	4.1	71	0.00
26 S	2,4-Dichlorophenol-d3	0.318	0.312	1.9	71	0.00
27 C	2,4-Dichlorophenol	0.306	0.299	2.3	71	0.00
28	Naphthalene	1.037	1.030	0.7	73	0.00
29 S	4-Chloroaniline-d4	0.351	0.381	-8.5	90	0.00
30	4-Chloroaniline	0.333	0.361	-8.4	89	0.00
31 C	Hexachlorobutadiene	0.174	0.172	1.1	74	0.00
32	Caprolactam	0.096	0.092	4.2	72	0.00
33 C	4-Chloro-3-methylphenol	0.325	0.322	0.9	73	0.00
34	2-Methylnaphthalene	0.770	0.780	-1.3	74	0.00
35	1-Methylnaphthalene	0.734	0.752	-2.5	75	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
37	1,2,4,5-Tetrachlorobenzene	0.593	0.568	4.2	75	0.00
38	Hexachlorocyclopentadiene	0.333	0.266	20.1#	67	0.00
39 C	2,4,6-Trichlorophenol	0.404	0.373	7.7	73	0.00
40	2,4,5-Trichlorophenol	0.410	0.394	3.9	75	0.00
41	1,1'-Biphenyl	1.664	1.613	3.1	76	0.00
42	2-Chloronaphthalene	1.252	1.203	3.9	76	0.00
43	2-Nitroaniline	0.389	0.380	2.3	76	0.00
44 S	Dimethylphthalate-d6	1.691	1.648	2.5	76	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM091117\
 Data File : BM011508.D
 Acq On : 11 Sep 2017 22:04
 Operator : SJ/JU
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02013

Quant Time: Sep 12 07:12:32 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 02:59:45 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.619	1.578	2.5	76	0.00
46	2,6-Dinitrotoluene	0.286	0.279	2.4	76	0.00
47 S	Acenaphthylene-d8	2.045	2.058	-0.6	78	0.00
48	Acenaphthylene	2.058	2.071	-0.6	78	0.00
49	3-Nitroaniline	0.353	0.324	8.2	75	0.00
50 C	Acenaphthene	1.391	1.392	-0.1	79	0.00
51	2,4-Dinitrophenol	0.177	0.138	22.0#	71	0.00
52 S	4-Nitrophenol-d4	0.318	0.293	7.9	77	0.00
53	4-Nitrophenol	0.204	0.197	3.4	81	0.00
54	Dibenzofuran	1.945	1.958	-0.7	79	0.00
55	2,4-Dinitrotoluene	0.420	0.418	0.5	77	0.00
56	2,3,4,6-Tetrachlorophenol	0.373	0.365	2.1	77	0.00
57	Diethylphthalate	1.685	1.691	-0.4	78	0.00
58 S	Fluorene-d10	1.465	1.480	-1.0	80	0.00
59	Fluorene	1.630	1.687	-3.5	80	0.00
60	4-Chlorophenyl-phenylether	0.764	0.772	-1.0	79	0.00
61	4-Nitroaniline	0.378	0.365	3.4	78	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.116	0.097	16.4	75	0.00
64	4,6-Dinitro-2-methylphenol	0.119	0.102	14.3	75	0.00
65	N-Nitrosodiphenylamine	0.609	0.605	0.7	80	0.00
66	4-Bromophenyl-phenylether	0.204	0.196	3.9	79	0.00
67	Hexachlorobenzene	0.225	0.214	4.9	79	0.00
68	Atrazine	0.211	0.200	5.2	79	0.00
69 C	Pentachlorophenol	0.144	0.123	14.6	76	0.00
70	Phenanthrene	1.115	1.148	-3.0	82	0.00
71 S	Anthracene-d10	0.985	1.002	-1.7	82	0.00
72	Anthracene	1.144	1.212	-5.9	84	0.00
73	1,2,3,4-Tetrachlorobenzene	0.267	0.248	7.1	75	0.00
74	Pentachlorobenzene	0.270	0.256	5.2	77	0.00
75	Carbazole	0.976	1.044	-7.0	82	0.00
76	Di-n-butylphthalate	1.282	1.357	-5.9	81	0.00
77 C	Fluoranthene	1.147	1.328	-15.8	83	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
79 S	Pyrene-d10	0.936	0.884	5.6	85	0.00
80	Pyrene	1.167	1.147	1.7	85	0.00
81	Butylbenzylphthalate	0.537	0.490	8.8	83	0.00
82	3,3'-Dichlorobenzidine	0.363	0.353	2.8	87	0.00
83	Benzo(a)anthracene	1.101	1.131	-2.7	90	0.00
84	Bis(2-ethylhexyl)phthalate	0.824	0.770	6.6	82	0.00
85	Chrysene	0.998	1.004	-0.6	89	0.00
86 I	Perylene-d12	1.000	1.000	0.0	92	0.00
87	Di-n-octyl phthalate	1.419	1.446	-1.9	85	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM091117\
Data File : BM011508.D
Acq On : 11 Sep 2017 22:04
Operator : SJ/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02013

Quant Time: Sep 12 07:12:32 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090817MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 11 02:59:45 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.203	1.180	1.9	91	0.00
89	Benzo(k)fluoranthene	1.156	1.128	2.4	89	0.00
90 S	Benzo(a)pyrene-d12	0.953	0.939	1.5	91	0.00
91 C	Benzo(a)pyrene	1.136	1.138	-0.2	91	0.00
92	Indeno(1,2,3-cd)pyrene	1.249	1.305	-4.5	94	0.00
93	Dibenzo(a,h)anthracene	1.059	1.095	-3.4	94	0.00
94	Benzo(g,h,i)perylene	1.012	1.051	-3.9	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0