

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM091117\
 Data File : BM011494.D
 Acq On : 11 Sep 2017 13:37
 Operator : SJ/JU
 Sample : MDL-S-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Manual Integrations
 APPROVED

Sohil
 9/12/2017 2:21:54 PM

Quant Time: Sep 11 14:15:33 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 13:34:52 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.97	152	437182	20.00	ng/ul	0.00
18) Naphthalene-d8	10.79	136	1989308	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.60	164	1197575	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.34	188	2680277	20.00	ng/ul	0.00
78) Chrysene-d12	21.51	240	3038789	20.00	ng/ul	0.00
86) Perylene-d12	23.88	264	2861840	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.40	96	45563	4.69	ng/uL	0.00
5) Phenol-d5	7.12	99	1225487	29.19	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.30	67	911806	32.23	ng/ul	0.00
9) 2-Chlorophenol-d4	7.50	132	984997	31.13	ng/ul	0.00
13) 4-Methylphenol-d8	8.67	113	999830	30.51	ng/ul	0.00
19) Nitrobenzene-d5	9.14	128	486803	32.24	ng/ul	0.00
22) 2-Nitrophenol-d4	9.86	143	490969	31.28	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.40	165	946760	29.96	ng/ul	0.00
29) 4-Chloroaniline-d4	10.92	131	1400639	40.16	ng/ul	0.00
44) Dimethylphthalate-d6	14.01	166	3273352	32.33	ng/ul	0.00
47) Acenaphthylene-d8	14.30	160	3991197	32.59	ng/ul	0.00
52) 4-Nitrophenol-d4	14.79	143	492664	25.89	ng/ul	0.00
58) Fluorene-d10	15.60	176	2823498	32.20	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.70	200	403540	25.88	ng/ul	0.00
71) Anthracene-d10	17.44	188	4348566	32.95	ng/ul	0.00
79) Pyrene-d10	19.73	212	4756112	33.43	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.73	264	4530236	33.22	ng/ul	0.00

Target Compounds

					Qvalue	
2) 1,4-Dioxane	3.43	88	17144	1.609	ng/uL#	74
4) Benzaldehyde	7.12	77	52045	2.170	ng/ul	96
6) Phenol	7.15	94	141997	3.408	ng/ul	96
8) Bis(2-Chloroethyl)ether	7.39	93	122834	3.744	ng/ul#	85
10) 2-Chlorophenol	7.54	128	109149	3.537	ng/ul	91
11) 2-Methylphenol	8.41	108	97070	3.072	ng/ul	90
12) 2,2'-oxybis(1-Chloropropan	8.51	45	198025	3.724	ng/ul#	93
14) Acetophenone	8.80	105	175473	3.503	ng/ul#	93
15) N-Nitroso-di-n-propylamine	8.78	70	94913	3.531	ng/ul#	77
16) 4-Methylphenol	8.73	108	116795	3.378	ng/ul	90
17) Hexachloroethane	9.06	117	40874	3.464	ng/ul#	81
20) Nitrobenzene	9.18	77	132451	3.730	ng/ul	95
21) Isophorone	9.70	82	245439	3.433	ng/ul#	97
23) 2-Nitrophenol	9.89	139	58415	3.541	ng/ul#	92
24) 2,4-Dimethylphenol	9.95	107	122280	3.465	ng/ul	95
25) Bis(2-Chloroethoxy)methane	10.19	93	164575	3.530	ng/ul#	97
27) 2,4-Dichlorophenol	10.42	162	104893	3.443	ng/ul#	85
28) Naphthalene	10.83	128	383683	3.719	ng/ul	99
30) 4-Chloroaniline	10.94	127	136501	4.119	ng/ul	96
31) Hexachlorobutadiene	11.12	225	60832	3.518	ng/ul	94
32) Caprolactam	11.75	113	27774m	2.896	ng/ul	
33) 4-Chloro-3-methylphenol	12.05	107	101468	3.141	ng/ul	92
34) 2-Methylnaphthalene	12.44	142	270212	3.527	ng/ul	98

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM091117\
 Data File : BM011494.D
 Acq On : 11 Sep 2017 13:37
 Operator : SJ/JU
 Sample : MDL-S-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Manual Integrations
 APPROVED

Sohil
 9/12/2017 2:21:54 PM

Quant Time: Sep 11 14:15:33 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 13:34:52 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 1-Methylnaphthalene	12.66	142	263581	3.613	ng/ul#	94
37) 1,2,4,5-Tetrachlorobenzene	12.80	216	127846	3.599	ng/ul#	97
38) Hexachlorocyclopentadiene	12.78	237	49456	2.484	ng/ul	98
39) 2,4,6-Trichlorophenol	13.04	196	73748	3.049	ng/ul	98
40) 2,4,5-Trichlorophenol	13.11	196	71572	2.918	ng/ul	95
41) 1,1'-Biphenyl	13.44	154	354646	3.559	ng/ul#	93
42) 2-Chloronaphthalene	13.49	162	259970	3.467	ng/ul	100
43) 2-Nitroaniline	13.69	65	60474	2.597	ng/ul	86
45) Dimethylphthalate	14.06	163	352007	3.631	ng/ul#	98
46) 2,6-Dinitrotoluene	14.17	165	49719	2.904	ng/ul#	86
48) Acenaphthylene	14.33	152	451572	3.664	ng/ul	99
49) 3-Nitroaniline	14.51	138	47418	2.246	ng/ul#	97
50) Acenaphthene	14.67	153	299353	3.595	ng/ul	98
51) 2,4-Dinitrophenol	14.72	184	20024	1.893	ng/ul#	76
53) 4-Nitrophenol	14.80	109	35970	2.941	ng/ul#	34
54) Dibenzofuran	15.00	168	423401	3.635	ng/ul	95
55) 2,4-Dinitrotoluene	14.96	165	77584	3.083	ng/ul#	75
56) 2,3,4,6-Tetrachlorophenol	15.23	232	68970	3.086	ng/ul#	85
57) Diethylphthalate	15.41	149	341820	3.389	ng/ul	97
59) Fluorene	15.65	166	353781	3.624	ng/ul	100
60) 4-Chlorophenyl-phenylether	15.64	204	163056	3.565	ng/ul	91
61) 4-Nitroaniline	15.67	138	55271m	2.444	ng/ul	
64) 4,6-Dinitro-2-methylphenol	15.72	198	50211	3.142	ng/ul#	50
65) N-Nitrosodiphenylamine	15.85	169	288181	3.530	ng/ul	99
66) 4-Bromophenyl-phenylether	16.53	248	93543	3.417	ng/ul	92
67) Hexachlorobenzene	16.65	284	104507	3.468	ng/ul#	89
68) Atrazine	16.80	200	82722	2.924	ng/ul	95
69) Pentachlorophenol	16.99	266	46350	2.395	ng/ul	96
70) Phenanthrene	17.39	178	549594	3.679	ng/ul	99
72) Anthracene	17.48	178	568026	3.705	ng/ul	98
73) 1,2,3,4-Tetrachlorobenzene	13.41	216	127743	3.577	ng/uL	99
74) Pentachlorobenzene	14.92	250	130113	3.592	ng/uL	98
75) Carbazole	17.74	167	454519	3.475	ng/ul	98
76) Di-n-butylphthalate	18.30	149	544264	3.168	ng/ul	99
77) Fluoranthene	19.39	202	589706	3.838	ng/ul#	87
80) Pyrene	19.76	202	673634	3.799	ng/ul#	85
81) Butylbenzylphthalate	20.64	149	228097	2.798	ng/ul#	91
82) 3,3'-Dichlorobenzidine	21.42	252	152786	2.769	ng/ul#	94
83) Benzo(a)anthracene	21.49	228	602604	3.602	ng/ul	100
84) Bis(2-ethylhexyl)phthalate	21.40	149	359214	2.869	ng/ul#	96
85) Chrysene	21.54	228	548429	3.616	ng/ul	100
87) Di-n-octyl phthalate	22.31	149	600128	2.955	ng/ul	100
88) Benzo(b)fluoranthene	23.16	252	576106	3.346	ng/ul#	95
89) Benzo(k)fluoranthene	23.20	252	559455	3.384	ng/ul#	94
91) Benzo(a)pyrene	23.77	252	585804	3.605	ng/ul#	95
92) Indeno(1,2,3-cd)pyrene	26.31	276	613173	3.430	ng/ul#	86
93) Dibenzo(a,h)anthracene	26.32	278	505616	3.336	ng/ul#	94
94) Benzo(g,h,i)perylene	27.06	276	507160	3.503	ng/ul#	90

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM091117\
Data File : BM011494.D
Acq On : 11 Sep 2017 13:37
Operator : SJ/JU
Sample : MDL-S-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MDL-S-05


Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM

Quant Time: Sep 11 14:15:33 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090817MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 11 13:34:52 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						



#2

1,4-Dioxane

Concen: 1.609 ng/uL

RT: 3.43 min Scan# 24

Delta R.T. -0.01 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M





ClientSampleId :

MDL-S-05

Tgt Ion:	88	Resp:	17144
Ion	Ratio	Lower	Upper
88	100		
43	40.1	29.0	43.4
58	85.8	45.9	68.9#

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM



#4

Benzaldehyde

Concen: 2.170 ng/uL




RT: 7.12 min Scan# 652


Delta R.T. 0.01 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Tgt Ion:	77	Resp:	52045
Ion	Ratio	Lower	Upper
77	100		
105	93.3	71.0	106.6
106	88.9	69.5	104.3





#6

Phenol

Concen: 3.408 ng/ul

RT: 7.15 min Scan# 657

Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M





ClientSampled :

MDL-S-05

Tgt Ion:	94	Resp:	141997
Ion	Ratio	Lower	Upper
94	100		
65	30.2	23.0	34.6
66	41.0	30.4	45.6

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM



#8

Bis(2-Chloroethyl)ether

Concen: 3.744 ng/ul




RT: 7.39 min Scan# 698


Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Tgt Ion:	93	Resp:	122834
Ion	Ratio	Lower	Upper
93	100		
63	91.0	59.4	89.2#
95	35.5	26.4	39.6









#10
2-Chlorophenol
Concen: 3.537 ng/ul
RT: 7.54 min Scan# 723
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	109149		
64	57.9	39.8	59.6	
130	29.5	25.8	38.8	




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#11
2-Methylphenol
Concen: 3.072 ng/ul
RT: 8.41 min Scan# 871
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
108	100	97070		
107	83.4	74.1	111.1	









#12
 2,2'-oxybis(1-chloropropane)
 Concen: 3.724 ng/ul
 RT: 8.51 min Scan# 888
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Lower	Upper
45	100		
77	11.8	12.5	18.7#
79	10.3	9.4	14.2




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#14
 Acetophenone
 Concen: 3.503 ng/ul
 RT: 8.80 min Scan# 937
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Lower	Upper
105	100		
77	77.1	63.2	94.8
51	35.3	19.8	29.8#









#15
N-Nitroso-di-n-propylamine
Concen: 3.531 ng/ul
RT: 8.78 min Scan# 934
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion	Ion	Ratio	Lower	Upper
70	100			
42	74.5	40.2	60.4#	
101	9.2	7.7	11.5	
130	19.2	17.8	26.6	




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#16
4-Methylphenol
Concen: 3.378 ng/ul
RT: 8.73 min Scan# 926
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ion	Ratio	Lower	Upper
108	100			
107	106.1	93.6	140.4	





#17

Hexachloroethane

Concen: 3.464 ng/ul

RT: 9.06 min Scan# 982

Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M





ClientSampleId :

MDL-S-05

Tgt Ion	Ratio	Lower	Upper
117	100		
201	101.6	96.1	144.1
199	54.9	59.7	89.5

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM



#20

Nitrobenzene

Concen: 3.730 ng/ul




RT: 9.18 min Scan# 1002


Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Lower	Upper
77	100		
123	43.2	31.8	47.8
65	15.2	10.6	16.0





#21

Isophorone

Concen: 3.433 ng/ul

RT: 9.70 min Scan# 1091

Delta R.T. -0.01 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M





ClientSampleId :

MDL-S-05

Tgt Ion:	82	Resp:	245439
Ion Ratio	Lower	Upper	
82	100		
95	5.1	6.1	9.1#
138	16.7	13.5	20.3

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM



#23

2-Nitrophenol

Concen: 3.541 ng/ul




RT: 9.89 min Scan# 1123


Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Tgt Ion:	139	Resp:	58415
Ion Ratio	Lower	Upper	
139	100		
65	60.3	45.3	67.9
109	27.3	27.7	41.5#





#24

2,4-Dimethylphenol

Concen: 3.465 ng/ul

RT: 9.95 min Scan# 1132

Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M





ClientSampleId :

MDL-S-05

Tgt Ion:	107	Resp:	122280
Ion Ratio	Lower	Upper	
107	100		
121	48.0	38.1	57.1
122	86.8	63.8	95.8

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM



#25

Bis(2-Chloroethoxy)methane

Concen: 3.530 ng/ul




RT: 10.19 min Scan# 1173


Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Tgt Ion:	93	Resp:	164575
Ion Ratio	Lower	Upper	
93	100		
95	31.8	24.7	37.1
123	12.5	8.2	12.2#









#27
 2,4-Dichlorophenol
 Concen: 3.443 ng/ul
 RT: 10.42 min Scan# 1213
 Delta R.T. -0.01 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	104893		
164	79.9	52.0	78.0#	
98	37.4	26.6	39.8	




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#28
 Naphthalene
 Concen: 3.719 ng/ul
 RT: 10.83 min Scan# 1283
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	383683		
129	10.8	8.3	12.5	
127	13.3	10.8	16.2	









#30
4-Chloroaniline
Concen: 4.119 ng/ul
RT: 10.94 min Scan# 1301
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
127	100	136501		
129	33.0	24.6	37.0	




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#31
Hexachlorobutadiene
Concen: 3.518 ng/ul
RT: 11.12 min Scan# 1331
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
225	100	60832		
223	68.4	48.1	72.1	
227	66.2	52.1	78.1	









#32
 Caprolactam
 Concen: 2.896 ng/ul m
 RT: 11.75 min Scan# 1439
 Delta R.T. 0.04 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion:113 Resp: 27774
 Ion Ratio Lower Upper
 113 100
 55 184.5 131.7 197.5
 56 146.7 92.7 139.1#




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#33
 4-Chloro-3-methylphenol
 Concen: 3.141 ng/ul
 RT: 12.05 min Scan# 1490
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion:107 Resp: 101468
 Ion Ratio Lower Upper
 107 100
 144 27.3 21.1 31.7
 142 83.7 60.1 90.1









#34
 2-Methylnaphthalene
 Concen: 3.527 ng/ul
 RT: 12.44 min Scan# 1556
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
142	100	270212		
141	87.9	71.4	107.2	




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#35
 1-Methylnaphthalene
 Concen: 3.613 ng/ul
 RT: 12.66 min Scan# 1593
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
142	100	263581		
141	89.5	76.6	115.0	
116	3.4	3.6	5.4#	









#37
 1,2,4,5-Tetrachlorobenzene
 Concen: 3.599 ng/ul
 RT: 12.80 min Scan# 1618
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
216	100	127846		
214	80.4	64.1	96.1	
179	22.9	14.5	21.7#	
108	18.0	10.4	15.6#	




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#38
 Hexachlorocyclopentadiene
 Concen: 2.484 ng/ul
 RT: 12.78 min Scan# 1614
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
237	100	49456		
235	60.9	49.5	74.3	
272	12.2	8.6	12.8	









#39
 2,4,6-Trichlorophenol
 Concen: 3.049 ng/ul
 RT: 13.04 min Scan# 1658
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	73748		
198	96.5	75.9	113.9	
200	33.9	25.8	38.6	




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#40
 2,4,5-Trichlorophenol
 Concen: 2.918 ng/ul
 RT: 13.11 min Scan# 1670
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	71572		
198	98.7	74.7	112.1	
200	29.7	25.4	38.0	









#41
 1,1'-Biphenyl
 Concen: 3.559 ng/ul
 RT: 13.44 min Scan# 1726
 Delta R.T. -0.01 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp Lower	Resp Upper
154	100		
153	38.8	34.0	51.0
76	14.1	8.4	12.6




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#42
 2-Chloronaphthalene
 Concen: 3.467 ng/ul
 RT: 13.49 min Scan# 1734
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp Lower	Resp Upper
162	100		
164	32.9	26.2	39.4
127	37.1	29.5	44.3









#43
2-Nitroaniline
Concen: 2.597 ng/ul
RT: 13.69 min Scan# 1768
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion	Ratio	Lower	Upper
65	100		
92	60.3	57.0	85.4
138	94.5	87.2	130.8




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#45
Dimethylphthalate
Concen: 3.631 ng/ul
RT: 14.06 min Scan# 1831
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.7	4.0	6.0#
164	10.1	8.4	12.6









#46
2,6-Dinitrotoluene
Concen: 2.904 ng/ul
RT: 14.17 min Scan# 1851
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	49719		
89	66.4	43.8	65.8#	
121	25.6	17.8	26.6	




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#48
Acenaphthylene
Concen: 3.664 ng/ul
RT: 14.33 min Scan# 1877
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	451572		
151	19.7	16.1	24.1	
153	13.4	10.8	16.2	









#49
3-Nitroaniline
Concen: 2.246 ng/ul
RT: 14.51 min Scan# 1908
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion:138 Resp: 47418
Ion Ratio Lower Upper
138 100
108 15.0 7.6 11.4#
92 120.3 94.2 141.2




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#50
Acenaphthene
Concen: 3.595 ng/ul
RT: 14.67 min Scan# 1935
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion:153 Resp: 299353
Ion Ratio Lower Upper
153 100
152 47.1 37.9 56.9
154 90.1 69.8 104.6









#51
2,4-Dinitrophenol
Concen: 1.893 ng/ul
RT: 14.72 min Scan# 1943
Delta R.T. 0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampled :
MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
184	100	20024		
63	78.7	39.7	59.5#	
154	61.9	45.0	67.6	




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#53
4-Nitrophenol
Concen: 2.941 ng/ul
RT: 14.80 min Scan# 1957
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
109	100	35970		
139	160.7	91.2	136.8#	
65	211.8	92.6	139.0#	









#54
Dibenzofuran
Concen: 3.635 ng/ul
RT: 15.00 min Scan# 1992
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion	Ratio	Lower	Upper
168	100		
139	35.9	31.0	46.4




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#55
2,4-Dinitrotoluene
Concen: 3.083 ng/ul
RT: 14.96 min Scan# 1985
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Lower	Upper
165	100		
63	55.5	31.4	47.0#
182	2.8	2.7	4.1









#56
 2,3,4,6-Tetrachlorophenol
 Concen: 3.086 ng/ul
 RT: 15.23 min Scan# 2030
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
232	100	68970		
131	44.7	27.1	40.7#	
130	2.6	0.0	0.0#	
166	30.4	20.0	30.0#	




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#57
 Diethylphthalate
 Concen: 3.389 ng/ul
 RT: 15.41 min Scan# 2061
 Delta R.T. -0.01 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	341820		
177	22.4	19.3	28.9	
150	12.5	10.6	15.8	





#59

Fluorene

Concen: 3.624 ng/ul

RT: 15.65 min Scan# 2102

Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M





ClientSampleId :

MDL-S-05

Tgt Ion:	166	Resp:	353781
Ion Ratio	Lower	Upper	
166	100		
165	97.7	78.4	117.6
167	14.0	11.1	16.7

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM



#60

4-Chlorophenyl-phenylether

Concen: 3.565 ng/ul




RT: 15.64 min Scan# 2100


Delta R.T. -0.01 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Tgt Ion:	204	Resp:	163056
Ion Ratio	Lower	Upper	
204	100		
206	33.4	26.6	40.0
141	60.8	40.7	61.1









#61
4-Nitroaniline
Concen: 2.444 ng/ul m
RT: 15.67 min Scan# 2105
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampled :
MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	55271		
92	47.9	38.2	57.4	
108	56.9	70.0	105.0#	




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#64
4,6-Dinitro-2-methylphenol
Concen: 3.142 ng/ul
RT: 15.72 min Scan# 2113
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
198	100	50211		
182	10.8	3.5	5.3#	
77	49.9	18.2	27.4#	









#65
N-Nitrosodiphenylamine
Concen: 3.530 ng/ul
RT: 15.85 min Scan# 2136
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	288181		
168	67.4	53.8	80.8	
167	35.6	29.5	44.3	




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#66
4-Bromophenyl-phenylether
Concen: 3.417 ng/ul
RT: 16.53 min Scan# 2252
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
248	100	93543		
250	95.9	79.0	118.4	
141	76.7	51.9	77.9	






#67
Hexachlorobenzene
Concen: 3.468 ng/ul
RT: 16.65 min Scan# 2272
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampled :
MDL-S-05




Tgt Ion	Ratio	Resp Lower	Resp Upper
284	100		
142	33.8	21.3	31.9#
249	27.9	26.3	39.5

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM




Abundance Ion 284.00 (283.70 to 284.70): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 2.924 ng/ul
RT: 16.80 min Scan# 2297
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp Lower	Resp Upper
200	100		
173	28.5	20.6	31.0
215	45.9	39.8	59.6



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 2.395 ng/ul
 RT: 16.99 min Scan# 2330
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp Lower	Resp Upper
266	100		
264	59.2	48.2	72.4
268	61.2	52.3	78.5




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#70
 Phenanthrene
 Concen: 3.679 ng/ul
 RT: 17.39 min Scan# 2397
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp Lower	Resp Upper
178	100		
179	15.3	12.3	18.5
176	18.4	15.4	23.0





#72

Anthracene

Concen: 3.705 ng/u1

RT: 17.48 min Scan# 2413

Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M





ClientSampled :

MDL-S-05

Tgt Ion:	178	Resp:	568026
Ion Ratio	Lower	Upper	
178	100		
179	15.6	13.4	20.2
176	18.5	15.4	23.0

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM



#73

1,2,3,4-Tetrachlorobenzene

Concen: 3.577 ng/uL




RT: 13.41 min Scan# 1721


Delta R.T. -0.00 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Tgt Ion:	216	Resp:	127743
Ion Ratio	Lower	Upper	
216	100		
214	80.7	64.1	96.1
218	50.1	39.4	59.0









#74
 Pentachlorobenzene
 Concen: 3.592 ng/uL
 RT: 14.92 min Scan# 1978
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp Lower	Resp Upper
250	100		
248	66.8	51.4	77.2
252	62.7	49.8	74.6




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#75
 Carbazole
 Concen: 3.475 ng/uL
 RT: 17.74 min Scan# 2458
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp Lower	Resp Upper
167	100		
166	20.7	17.0	25.4
139	12.2	9.1	13.7









#76
Di-n-butylphthalate
Concen: 3.168 ng/ul
RT: 18.30 min Scan# 2552
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
150	9.1	7.5	11.3
104	4.2	4.1	6.1




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#77
Fluoranthene
Concen: 3.838 ng/ul
RT: 19.39 min Scan# 2738
Delta R.T. -0.00 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
101	12.5	6.1	9.1#
100	9.6	4.6	7.0#









#80
 Pyrene
 Concen: 3.799 ng/ul
 RT: 19.76 min Scan# 2800
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
101	14.9	6.9	10.3#
100	11.5	5.6	8.4#




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#81
 Butylbenzylphthalate
 Concen: 2.798 ng/ul
 RT: 20.64 min Scan# 2950
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	73.8	53.4	80.0
206	18.6	19.3	28.9#









#82
 3,3'-Dichlorobenzidine
 Concen: 2.769 ng/ul
 RT: 21.42 min Scan# 3083
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	152786		
254	62.2	52.8	79.2	
126	14.5	6.9	10.3	




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#83
 Benzo(a)anthracene
 Concen: 3.602 ng/ul
 RT: 21.49 min Scan# 3095
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
228	100	602604		
229	19.6	16.0	24.0	
226	27.4	21.8	32.8	









#84
 Bis(2-ethylhexyl)phthalate
 Concen: 2.869 ng/ul
 RT: 21.40 min Scan# 3080
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Instrument :
 BNA_M
 ClientSampleId :
 MDL-S-05

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	359214		
167	26.7	23.0	34.6	
279	2.9	4.3	6.5	




Manual Integrations
 APPROVED


Sohil
 9/12/2017 2:21:54 PM



#85
 Chrysene
 Concen: 3.616 ng/ul
 RT: 21.54 min Scan# 3104
 Delta R.T. -0.00 min
 Lab File: BM011494.D
 Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Resp	Lower	Upper
228	100	548429		
226	30.0	24.1	36.1	
229	19.7	16.0	24.0	









#87
Di-n-octyl phthalate
Concen: 2.955 ng/ul
RT: 22.31 min Scan# 3235
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampleId :
MDL-S-05

Tgt Ion:149 Resp: 600128




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#88
Benzo(b)fluoranthene
Concen: 3.346 ng/ul
RT: 23.16 min Scan# 3378
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion:252 Resp: 576106
Ion Ratio Lower Upper
252 100
253 22.7 17.4 26.0
125 11.5 4.9 7.3#





#89

Benzo(k)fluoranthene

Concen: 3.384 ng/ul

RT: 23.20 min Scan# 3386

Delta R.T. -0.01 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M





ClientSampleId :

MDL-S-05

Tgt Ion:	252	Resp:	559455
Ion Ratio	Lower	Upper	
252	100		
253	23.0	17.1	25.7
125	11.8	5.4	8.0#

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM



#91

Benzo(a)pyrene

Concen: 3.605 ng/ul




RT: 23.77 min Scan# 3483


Delta R.T. -0.01 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Tgt Ion:	252	Resp:	585804
Ion Ratio	Lower	Upper	
252	100		
253	21.9	17.2	25.8
125	12.9	5.8	8.6#









#92
Indeno(1,2,3-cd)pyrene
Concen: 3.430 ng/ul
RT: 26.31 min Scan# 3914
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Instrument :
BNA_M
ClientSampled :
MDL-S-05

Tgt Ion	Ratio	Lower	Upper
276	100		
138	28.6	10.9	16.3#
277	23.5	18.9	28.3




Manual Integrations
APPROVED


Sohil
9/12/2017 2:21:54 PM



#93
Dibenzo(a,h)anthracene
Concen: 3.336 ng/ul
RT: 26.32 min Scan# 3916
Delta R.T. -0.01 min
Lab File: BM011494.D
Acq: 11 Sep 2017 13:37

Tgt Ion	Ratio	Lower	Upper
278	100		
139	18.0	9.0	13.4#
279	23.8	19.1	28.7





#94

Benzo(g,h,i)perylene

Concen: 3.503 ng/ul

RT: 27.06 min Scan# 4042

Delta R.T. -0.01 min

Lab File: BM011494.D

Acq: 11 Sep 2017 13:37

Instrument :

BNA_M

ClientSampleId :

MDL-S-05

Tgt	Ion	276	Resp:	507160
	Ratio	100	Lower	Upper
	276	100		
	138	24.4	11.2	16.8#
	277	23.9	19.0	28.6

Manual Integrations
APPROVED

Sohil
9/12/2017 2:21:54 PM

