

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM091417\
 Data File : BM011557.D
 Acq On : 15 Sep 2017 00:53
 Operator : SJ/JU
 Sample : I5191-16
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AE6

Quant Time: Sep 15 02:29:29 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 15 02:27:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.97	152	385024	20.00	ng/ul	0.00
18) Naphthalene-d8	10.78	136	1755590	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.60	164	999501	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.34	188	2180797	20.00	ng/ul	0.00
78) Chrysene-d12	21.50	240	2203007	20.00	ng/ul	0.00
86) Perylene-d12	23.88	264	1976150	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.40	96	13166	1.54	ng/uL	0.00
5) Phenol-d5	7.12	99	213721	5.78	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.29	67	826470	33.17	ng/ul	0.00
9) 2-Chlorophenol-d4	7.50	132	682655	24.50	ng/ul	0.00
13) 4-Methylphenol-d8	8.67	113	413208	14.32	ng/ul	0.00
19) Nitrobenzene-d5	9.13	128	431850	32.40	ng/ul	0.00
22) 2-Nitrophenol-d4	9.86	143	438023	31.62	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.39	165	759485	27.23	ng/ul	0.00
29) 4-Chloroaniline-d4	10.92	131	47107	1.53	ng/ul	0.01
44) Dimethylphthalate-d6	14.00	166	2809195	33.25	ng/ul	0.00
47) Acenaphthylene-d8	14.30	160	3399311	33.26	ng/ul	0.00
52) 4-Nitrophenol-d4	14.79	143	56319	3.55	ng/ul	0.00
58) Fluorene-d10	15.59	176	2483984	33.94	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.70	200	336641	26.53	ng/ul	0.00
71) Anthracene-d10	17.44	188	3919067	36.49	ng/ul	0.00
79) Pyrene-d10	19.73	212	4278634	41.48	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.73	264	3413078	36.24	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM091417\
Data File : BM011557.D
Acq On : 15 Sep 2017 00:53
Operator : SJ/JU
Sample : I5191-16
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AE6

Quant Time: Sep 15 02:29:29 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM090817MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 15 02:27:08 2017
Response via : Initial Calibration

