

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM091417\
 Data File : BM011564.D
 Acq On : 15 Sep 2017 10:24
 Operator : SJ/JU
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02020

Quant Time: Sep 15 12:28:14 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 15 02:27:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.487	0.516	-6.0	88	0.00
3 S	1,4-Dioxane-d8	0.445	0.458	-2.9	86	0.00
4	Benzaldehyde	1.097	1.079	1.6	76	0.00
5 S	Phenol-d5	1.920	1.911	0.5	86	0.00
6	Phenol	1.906	1.918	-0.6	86	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.294	1.385	-7.0	89	0.00
8	Bis(2-Chloroethyl)ether	1.501	1.537	-2.4	87	0.00
9 S	2-Chlorophenol-d4	1.447	1.460	-0.9	84	0.00
10	2-Chlorophenol	1.412	1.421	-0.6	84	0.00
11	2-Methylphenol	1.445	1.425	1.4	85	0.00
12	2,2'-oxybis(1-Chloropropane	2.433	2.758	-13.4	94	0.00
13 S	4-Methylphenol-d8	1.499	1.500	-0.1	85	0.00
14	Acetophenone	2.292	2.298	-0.3	85	0.00
15 P	N-Nitroso-di-n-propylamine	1.230	1.320	-7.3	89	0.00
16	4-Methylphenol	1.582	1.593	-0.7	86	0.00
17	Hexachloroethane	0.540	0.549	-1.7	86	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	84	0.00
19 S	Nitrobenzene-d5	0.152	0.152	0.0	84	0.00
20	Nitrobenzene	0.357	0.386	-8.1	91	0.00
21	Isophorone	0.719	0.748	-4.0	87	0.00
22 S	2-Nitrophenol-d4	0.158	0.159	-0.6	86	0.00
23 C	2-Nitrophenol	0.166	0.168	-1.2	85	0.00
24	2,4-Dimethylphenol	0.355	0.365	-2.8	86	0.00
25	Bis(2-Chloroethoxy)methane	0.469	0.473	-0.9	85	0.00
26 S	2,4-Dichlorophenol-d3	0.318	0.315	0.9	82	0.00
27 C	2,4-Dichlorophenol	0.306	0.304	0.7	83	0.00
28	Naphthalene	1.037	1.037	0.0	84	0.00
29 S	4-Chloroaniline-d4	0.351	0.413	-17.7	112	0.00
30	4-Chloroaniline	0.333	0.390	-17.1	110	0.00
31 C	Hexachlorobutadiene	0.174	0.176	-1.1	86	0.00
32	Caprolactam	0.096	0.083	13.5	75	0.00
33 C	4-Chloro-3-methylphenol	0.325	0.330	-1.5	86	0.00
34	2-Methylnaphthalene	0.770	0.765	0.6	83	0.00
35	1-Methylnaphthalene	0.734	0.733	0.1	84	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
37	1,2,4,5-Tetrachlorobenzene	0.593	0.601	-1.3	84	0.00
38	Hexachlorocyclopentadiene	0.333	0.306	8.1	82	0.00
39 C	2,4,6-Trichlorophenol	0.404	0.394	2.5	82	0.00
40	2,4,5-Trichlorophenol	0.410	0.413	-0.7	84	0.00
41	1,1'-Biphenyl	1.664	1.647	1.0	83	0.00
42	2-Chloronaphthalene	1.252	1.247	0.4	84	0.00
43	2-Nitroaniline	0.389	0.423	-8.7	90	0.00
44 S	Dimethylphthalate-d6	1.691	1.636	3.3	80	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM091417\
 Data File : BM011564.D
 Acq On : 15 Sep 2017 10:24
 Operator : SJ/JU
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02020

Quant Time: Sep 15 12:28:14 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 15 02:27:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.619	1.575	2.7	81	0.00
46	2,6-Dinitrotoluene	0.286	0.286	0.0	83	0.00
47 S	Acenaphthylene-d8	2.045	2.048	-0.1	83	0.00
48	Acenaphthylene	2.058	2.067	-0.4	83	0.00
49	3-Nitroaniline	0.353	0.324	8.2	80	0.00
50 C	Acenaphthene	1.391	1.388	0.2	84	0.00
51	2,4-Dinitrophenol	0.177	0.139	21.5#	76	0.00
52 S	4-Nitrophenol-d4	0.318	0.293	7.9	82	0.00
53	4-Nitrophenol	0.204	0.202	1.0	89	0.00
54	Dibenzofuran	1.945	1.937	0.4	83	0.00
55	2,4-Dinitrotoluene	0.420	0.418	0.5	83	0.00
56	2,3,4,6-Tetrachlorophenol	0.373	0.372	0.3	84	0.00
57	Diethylphthalate	1.685	1.646	2.3	81	0.00
58 S	Fluorene-d10	1.465	1.457	0.5	84	0.00
59	Fluorene	1.630	1.659	-1.8	84	0.00
60	4-Chlorophenyl-phenylether	0.764	0.775	-1.4	85	0.00
61	4-Nitroaniline	0.378	0.343	9.3	78	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.116	0.103	11.2	83	0.00
64	4,6-Dinitro-2-methylphenol	0.119	0.105	11.8	80	0.00
65	N-Nitrosodiphenylamine	0.609	0.599	1.6	83	0.00
66	4-Bromophenyl-phenylether	0.204	0.200	2.0	83	0.00
67	Hexachlorobenzene	0.225	0.218	3.1	83	0.00
68	Atrazine	0.211	0.204	3.3	84	0.00
69 C	Pentachlorophenol	0.144	0.125	13.2	81	0.00
70	Phenanthrene	1.115	1.123	-0.7	83	0.00
71 S	Anthracene-d10	0.985	0.983	0.2	83	0.00
72	Anthracene	1.144	1.175	-2.7	84	0.00
73	1,2,3,4-Tetrachlorobenzene	0.267	0.268	-0.4	85	0.00
74	Pentachlorobenzene	0.270	0.269	0.4	84	0.00
75	Carbazole	0.976	1.027	-5.2	83	0.00
76	Di-n-butylphthalate	1.282	1.318	-2.8	81	0.00
77 C	Fluoranthene	1.147	1.291	-12.6	84	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
79 S	Pyrene-d10	0.936	0.928	0.9	86	0.00
80	Pyrene	1.167	1.192	-2.1	85	0.00
81	Butylbenzylphthalate	0.537	0.513	4.5	83	0.00
82	3,3'-Dichlorobenzidine	0.363	0.347	4.4	83	0.00
83	Benzo(a)anthracene	1.101	1.137	-3.3	87	0.00
84	Bis(2-ethylhexyl)phthalate	0.824	0.781	5.2	80	0.00
85	Chrysene	0.998	1.019	-2.1	87	0.00
86 I	Perylene-d12	1.000	1.000	0.0	82	0.00
87	Di-n-octyl phthalate	1.419	1.572	-10.8	82	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM091417\
Data File : BM011564.D
Acq On : 15 Sep 2017 10:24
Operator : SJ/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02020

Quant Time: Sep 15 12:28:14 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM090817MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 15 02:27:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.203	1.244	-3.4	85	0.00
89	Benzo(k)fluoranthene	1.156	1.170	-1.2	81	0.00
90 S	Benzo(a)pyrene-d12	0.953	0.969	-1.7	83	0.00
91 C	Benzo(a)pyrene	1.136	1.166	-2.6	83	0.00
92	Indeno(1,2,3-cd)pyrene	1.249	1.193	4.5	76	0.00
93	Dibenzo(a,h)anthracene	1.059	1.009	4.7	76	0.00
94	Benzo(a,h,i)perylene	1.012	0.947	6.4	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0