

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM091417\
 Data File : BM011563.D
 Acq On : 15 Sep 2017 09:48
 Operator : SJ/JU
 Sample : PB102331BL
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK31

Quant Time: Sep 15 12:25:25 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 15 02:27:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.97	152	352825	20.00	ng/ul	0.00
18) Naphthalene-d8	10.78	136	1586045	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.60	164	924478	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.34	188	2047556	20.00	ng/ul	0.00
78) Chrysene-d12	21.50	240	2098212	20.00	ng/ul	0.00
86) Perylene-d12	23.88	264	2008792	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.39	96	48114	6.13	ng/uL	0.00
5) Phenol-d5	7.12	99	929903	27.45	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.29	67	715703	31.35	ng/ul	0.00
9) 2-Chlorophenol-d4	7.50	132	740096	28.99	ng/ul	0.00
13) 4-Methylphenol-d8	8.67	113	744928	28.16	ng/ul	0.00
19) Nitrobenzene-d5	9.13	128	365997	30.40	ng/ul	0.00
22) 2-Nitrophenol-d4	9.86	143	373810	29.87	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.39	165	698632	27.73	ng/ul	0.00
29) 4-Chloroaniline-d4	10.91	131	1106641	39.80	ng/ul	0.00
44) Dimethylphthalate-d6	14.00	166	2758738	35.30	ng/ul	0.00
47) Acenaphthylene-d8	14.30	160	3046295	32.23	ng/ul	0.00
52) 4-Nitrophenol-d4	14.79	143	413390	28.14	ng/ul	0.00
58) Fluorene-d10	15.59	176	2305617	34.06	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.70	200	304656	25.57	ng/ul	0.00
71) Anthracene-d10	17.44	188	3737011	37.06	ng/ul	0.00
79) Pyrene-d10	19.73	212	4146242	42.20	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.73	264	3689248	38.54	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA M\DATA\BM091417\
 Data File : BM011563.D
 Acq On : 15 Sep 2017 09:48
 Operator : SJ/JU
 Sample : PB102331BL
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK31

Quant Time: Sep 15 12:25:25 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM090817MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 15 02:27:08 2017
 Response via : Initial Calibration

