

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091520\
 Data File : BM027431.D
 Acq On : 16 Sep 2020 03:18
 Operator : JU/CG
 Sample : PB131620BL
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK20

Quant Time: Sep 16 05:19:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 16 01:07:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	125671	20.00	ng/ul	0.00
18) Naphthalene-d8	10.32	136	422803	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.20	164	269527	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.94	188	579827	20.00	ng/ul	0.00
78) Chrysene-d12	21.14	240	604014	20.00	ng/ul	0.00
86) Perylene-d12	23.30	264	695597	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.13	96	18315	6.48	ng/uL	0.00
5) Phenol-d5	6.75	99	231908	22.39	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.90	67	161850	23.53	ng/ul	0.00
9) 2-Chlorophenol-d4	7.09	132	193930	25.57	ng/ul	0.00
13) 4-Methylphenol-d8	8.26	113	176787	20.93	ng/ul	0.00
19) Nitrobenzene-d5	8.70	128	88535	27.82	ng/ul	0.00
22) 2-Nitrophenol-d4	9.42	143	97444	27.92	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.95	165	180458	24.02	ng/ul	0.00
29) 4-Chloroaniline-d4	10.46	131	213109	24.00	ng/ul	0.00
44) Dimethylphthalate-d6	13.62	166	549752	25.63	ng/ul	0.00
47) Acenaphthylene-d8	13.89	160	671752	27.73	ng/ul	0.00
52) 4-Nitrophenol-d4	14.41	143	54799	14.51	ng/ul	0.00
58) Fluorene-d10	15.19	176	477641	24.87	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.32	200	63027	16.98	ng/ul	0.00
71) Anthracene-d10	17.04	188	773640	28.21	ng/ul	0.00
79) Pyrene-d10	19.34	212	873795	32.47	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.17	264	972658	26.07	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091520\
Data File : BM027431.D
Acq On : 16 Sep 2020 03:18
Operator : JU/CG
Sample : PB131620BL
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK20

Quant Time: Sep 16 05:19:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM090420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 16 01:07:14 2020
Response via : Initial Calibration

