

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036549.D
 Acq On : 15 Sep 2022 17:14
 Operator : CG/JU
 Sample : SSTD01073
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010020

Quant Time: Sep 16 01:50:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 01:29:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	354102	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	1501820	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	944947	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	1930748	20.000	ng/u1	0.00
79) Chrysene-d12	21.550	240	1879115	20.000	ng/u1	0.00
88) Perylene-d12	23.997	264	1838302	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	38076	4.139	ng/uL	0.00
4) Pyridine-d5	3.816	84	230811	8.039	ng/u1	0.00
7) Phenol-d5	7.169	99	284870	9.138	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	186359	9.240	ng/u1	0.00
11) 2-Chlorophenol-d4	7.551	132	219804	9.157	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	212348	8.617	ng/u1	0.00
21) Nitrobenzene-d5	9.186	128	93542	7.798	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	78424	6.436	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	197892	9.194	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	332407	9.429	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	632374	9.991	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	747290	9.876	ng/u1	0.00
54) 4-Nitrophenol-d4	14.821	143	98717	8.083	ng/u1	0.00
60) Fluorene-d10	15.639	176	560564	9.559	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	46643	4.629	ng/u1	0.00
73) Anthracene-d10	17.486	188	877782	10.042	ng/u1	0.00
81) Pyrene-d10	19.762	212	960746	9.790	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.838	264	873154	9.451	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	39893	4.141	ng/uL	98
5) Pyridine	3.840	79	236002	8.271	ng/u1	97
6) Benzaldehyde	7.157	77	164466	11.322	ng/u1	96
8) Phenol	7.198	94	312232	9.430	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.445	93	253095	9.493	ng/u1	100
12) 2-Chlorophenol	7.581	128	237780	9.225	ng/u1	98
13) 2-Methylphenol	8.463	108	230769	9.024	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.563	45	310101	7.746	ng/u1	100
16) Acetophenone	8.851	105	390742	9.381	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.834	70	191670	9.379	ng/u1	99
18) 4-Methylphenol	8.792	108	249853	9.036	ng/u1	96
19) Hexachloroethane	9.116	117	98362	8.947	ng/u1	92
22) Nitrobenzene	9.233	77	256771	8.417	ng/u1	99
23) Isophorone	9.757	82	549654	10.098	ng/u1	100
25) 2-Nitrophenol	9.945	139	100925	7.384	ng/u1	98
26) 2,4-Dimethylphenol	10.004	107	269267	9.602	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.245	93	325309	10.178	ng/u1	99
29) 2,4-Dichlorophenol	10.480	162	211248	9.517	ng/u1	99
30) Naphthalene	10.892	128	813598	9.935	ng/u1	99
32) 4-Chloroaniline	10.992	127	342131	9.426	ng/u1	100
33) Hexachlorobutadiene	11.180	225	135910	9.558	ng/u1	98
34) Caprolactam	11.757	113	65309	9.165	ng/u1	98
35) 4-Chloro-3-methylphenol	12.104	107	234114	9.979	ng/u1	99
36) 2-Methylnaphthalene	12.492	142	532715	10.114	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036549.D
 Acq On : 15 Sep 2022 17:14
 Operator : CG/JU
 Sample : SSTD01073
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD010020


Quant Time: Sep 16 01:50:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 01:29:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.710	142	539870	10.216	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.851	216	261117	9.714	ng/ul	98
40) Hexachlorocyclopentadiene	12.839	237	122067	8.676	ng/ul	98
41) 2,4,6-Trichlorophenol	13.086	196	155297	8.811	ng/ul	98
42) 2,4,5-Trichlorophenol	13.151	196	163274	8.776	ng/ul	98
43) 1,1'-Biphenyl	13.492	154	694885	9.874	ng/ul	98
44) 2-Chloronaphthalene	13.533	162	545304	9.968	ng/ul	99
45) 2-Nitroaniline	13.727	65	116418	7.048	ng/ul	95
47) Dimethylphthalate	14.110	163	671702	10.078	ng/ul	99
48) 2,6-Dinitrotoluene	14.221	165	94872	7.287	ng/ul	98
50) Acenaphthylene	14.374	152	858640	9.992	ng/ul	99
51) 3-Nitroaniline	14.545	138	106870	7.324	ng/ul	99
52) Acenaphthene	14.715	153	600171	9.798	ng/ul	99
53) 2,4-Dinitrophenol	14.745	184	31030	4.793	ng/ul	97
55) 4-Nitrophenol	14.833	109	93710	8.645	ng/ul	98
56) Dibenzofuran	15.045	168	814150	9.760	ng/ul	98
57) 2,4-Dinitrotoluene	14.998	165	138564	7.093	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.268	232	138267	8.740	ng/ul	99
59) Diethylphthalate	15.468	149	673297	9.257	ng/ul	99
61) Fluorene	15.692	166	687712	9.702	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.692	204	331807	9.972	ng/ul	97
63) 4-Nitroaniline	15.698	138	97601	6.842	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.757	198	55194	5.238	ng/ul#	99
67) N-Nitrosodiphenylamine	15.898	169	562617	10.168	ng/ul	99
68) 4-Bromophenyl-phenylether	16.580	248	188455	10.023	ng/ul	97
69) Hexachlorobenzene	16.692	284	224040	10.058	ng/ul	97
70) Atrazine	16.845	200	183351	8.946	ng/ul	98
71) Pentachlorophenol	17.027	266	112158	8.377	ng/ul	97
72) Phenanthrene	17.427	178	1054203	9.953	ng/ul	99
74) Anthracene	17.521	178	1076119	10.016	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.457	216	258993	10.316	ng/ul	99
76) Pentachlorobenzene	14.963	250	280913	10.141	ng/ul	98
77) Carbazole	17.786	167	961574	9.394	ng/ul	100
78) Di-n-butylphthalate	18.356	149	1173384	9.522	ng/ul	100
80) Fluoranthene	19.433	202	1183237	9.828	ng/ul	99
82) Pyrene	19.792	202	1259242	9.864	ng/ul	99
83) Butylbenzylphthalate	20.692	149	427860	7.838	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.468	252	280391	6.986	ng/ul	98
85) Benzo(a)anthracene	21.539	228	1182168	9.837	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.474	149	656990	8.048	ng/ul	100
87) Chrysene	21.592	228	1133915	10.156	ng/ul	99
89) Di-n-octyl phthalate	22.421	149	1018787	6.752	ng/ul	100
90) Benzo(b)fluoranthene	23.250	252	1159799	9.143	ng/ul	99
91) Benzo(k)fluoranthene	23.297	252	1148022	9.556	ng/ul	99
93) Benzo(a)pyrene	23.885	252	1000772	9.570	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.538	276	1166175	11.049	ng/ul	99
95) Dibenzo(a,h)anthracene	26.562	278	1054616	10.984	ng/ul	99
96) Benzo(g,h,i)perylene	27.321	276	1032545	11.445	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 16 01:50:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 16 01:29:28 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.141 ng/uL
RT: 3.428 min Scan# 51
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion: 88 Resp: 39893
Ion Ratio Lower Upper
88 100
43 30.6 23.1 34.7
58 71.7 56.6 85.0



#5
Pyridine
Concen: 8.271 ng/uL
RT: 3.840 min Scan# 128
Delta R.T. 0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion: 79 Resp: 236002
Ion Ratio Lower Upper
79 100
52 73.9 57.2 85.8
51 35.3 26.8 40.2









#6
Benzaldehyde
Concen: 11.322 ng/ul
RT: 7.157 min Scan# 692
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion: 77 Resp: 164466
Ion Ratio Lower Upper
77 100
105 97.3 75.1 112.7
106 94.7 73.0 109.6



#8
Phenol
Concen: 9.430 ng/ul
RT: 7.198 min Scan# 699
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion: 94 Resp: 312232
Ion Ratio Lower Upper
94 100
65 27.5 22.2 33.4
66 35.3 28.9 43.3









#10
Bis(2-Chloroethyl)ether
Concen: 9.493 ng/ul
RT: 7.445 min Scan# 741
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion: 93 Resp: 253095
Ion Ratio Lower Upper
93 100
63 74.0 59.1 88.7
95 32.2 26.1 39.1



#12
2-Chlorophenol
Concen: 9.225 ng/ul
RT: 7.581 min Scan# 764
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:128 Resp: 237780
Ion Ratio Lower Upper
128 100
64 46.2 37.8 56.8
130 31.5 26.0 39.0









#13
2-Methylphenol
Concen: 9.024 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:108 Resp: 230769
Ion Ratio Lower Upper
108 100
107 86.9 70.7 106.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 7.746 ng/ul
RT: 8.563 min Scan# 931
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion: 45 Resp: 310101
Ion Ratio Lower Upper
45 100
77 16.3 13.1 19.7
79 12.0 9.5 14.3









#16
Acetophenone
Concen: 9.381 ng/ul
RT: 8.851 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:105 Resp: 390742
Ion Ratio Lower Upper
105 100
77 79.2 61.6 92.4
51 25.2 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 9.379 ng/ul
RT: 8.834 min Scan# 977
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion: 70 Resp: 191670
Ion Ratio Lower Upper
70 100
42 52.8 43.0 64.4
101 8.9 7.7 11.5
130 19.4 15.1 22.7





#18

4-Methylphenol

Concen: 9.036 ng/ul

RT: 8.792 min Scan# 91

Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Instrument :

BNA_M

ClientSampleId :





SSTD010020

Tgt Ion:108 Resp: 249853

Ion Ratio Lower Upper

108 100

107 107.9 89.4 134.0



#19

Hexachloroethane

Concen: 8.947 ng/ul

RT: 9.116 min Scan# 1025

Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14




Tgt Ion:117 Resp: 98362


Ion Ratio Lower Upper

117 100

201 96.3 68.7 103.1

199 56.6 43.0 64.6









#22
Nitrobenzene
Concen: 8.417 ng/ul
RT: 9.233 min Scan# 1045
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion: 77 Resp: 256771
Ion Ratio Lower Upper
77 100
123 42.7 34.9 52.3
65 13.5 10.7 16.1



#23
Isophorone
Concen: 10.098 ng/ul
RT: 9.757 min Scan# 1134
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion: 82 Resp: 549654
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.2
138 17.1 13.5 20.3









#25
2-Nitrophenol
Concen: 7.384 ng/ul
RT: 9.945 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:139 Resp: 100925
Ion Ratio Lower Upper
139 100
65 53.2 40.9 61.3
109 29.2 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 9.602 ng/ul
RT: 10.004 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:107 Resp: 269267
Ion Ratio Lower Upper
107 100
121 50.1 39.4 59.2
122 88.8 69.1 103.7





#27

Bis(2-Chloroethoxy)methane

Concen: 10.178 ng/ul

RT: 10.245 min Scan# 1217

Delta R.T. -0.006 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Instrument :

BNA_M

ClientSampleId :

SSTD010020





Tgt Ion: 93 Resp: 325309

Ion Ratio Lower Upper

93 100

95 31.7 26.2 39.4

123 11.4 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 9.517 ng/ul

RT: 10.480 min Scan# 1257

Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14




Tgt Ion:162 Resp: 211248


Ion Ratio Lower Upper

162 100

164 65.5 52.1 78.1

98 37.8 29.1 43.7









#30
Naphthalene
Concen: 9.935 ng/ul
RT: 10.892 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:128 Resp: 813598
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 12.3 10.3 15.5



#32
4-Chloroaniline
Concen: 9.426 ng/ul
RT: 10.992 min Scan# 1344
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:127 Resp: 342131
Ion Ratio Lower Upper
127 100
129 31.6 25.4 38.2





#33

Hexachlorobutadiene

Concen: 9.558 ng/ul

RT: 11.180 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Instrument :

BNA_M

ClientSampleId :

SSTD010020





Tgt Ion:225 Resp: 135910

Ion Ratio Lower Upper

225 100

223 62.2 49.8 74.8

227 61.2 51.0 76.6



#34

Caprolactam

Concen: 9.165 ng/ul

RT: 11.757 min Scan# 1474

Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14




Tgt Ion:113 Resp: 65309


Ion Ratio Lower Upper

113 100

55 160.0 127.5 191.3

56 122.1 94.6 142.0









#35
4-Chloro-3-methylphenol
Concen: 9.979 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:107 Resp: 234114
Ion Ratio Lower Upper
107 100
144 25.9 21.5 32.3
142 80.4 63.9 95.9



#36
2-Methylnaphthalene
Concen: 10.114 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:142 Resp: 532715
Ion Ratio Lower Upper
142 100
141 87.9 70.7 106.1









#37
1-Methylnaphthalene
Concen: 10.216 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:142 Resp: 539870
Ion Ratio Lower Upper
142 100
141 92.9 73.4 110.2
116 3.9 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.714 ng/ul
RT: 12.851 min Scan# 1660
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:216 Resp: 261117
Ion Ratio Lower Upper
216 100
214 78.6 62.4 93.6
179 21.4 16.7 25.1
108 21.8 15.4 23.0









#40
Hexachlorocyclopentadiene
Concen: 8.676 ng/ul
RT: 12.839 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:237 Resp: 122067
Ion Ratio Lower Upper
237 100
235 62.3 50.9 76.3
272 13.6 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 8.811 ng/ul
RT: 13.086 min Scan# 1700
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:196 Resp: 155297
Ion Ratio Lower Upper
196 100
198 95.2 77.3 115.9
200 30.1 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 8.776 ng/ul

RT: 13.151 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Instrument :

BNA_M

ClientSampleId :

SSTD010020





Tgt Ion:196 Resp: 163274

Ion Ratio Lower Upper

196 100

198 97.3 75.6 113.4

200 30.6 24.7 37.1



#43

1,1'-Biphenyl

Concen: 9.874 ng/ul

RT: 13.492 min Scan# 1769

Delta R.T. -0.006 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14




Tgt Ion:154 Resp: 694885


Ion Ratio Lower Upper

154 100

153 40.5 31.4 47.2

76 13.5 10.4 15.6





#44

2-Chloronaphthalene

Concen: 9.968 ng/ul

RT: 13.533 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Instrument :

BNA_M

ClientSampleId :

SSTD010020





Tgt Ion:162 Resp: 545304

Ion Ratio Lower Upper

162 100

164 32.6 26.2 39.2

127 40.0 32.6 49.0



#45

2-Nitroaniline

Concen: 7.048 ng/ul

RT: 13.727 min Scan# 1809

Delta R.T. -0.006 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14




Tgt Ion: 65 Resp: 116418


Ion Ratio Lower Upper

65 100

92 68.3 54.4 81.6

138 100.0 86.1 129.1









#47
Dimethylphthalate
Concen: 10.078 ng/ul
RT: 14.110 min Scan# 1874
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:163 Resp: 671702
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.6
164 10.1 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 7.287 ng/ul
RT: 14.221 min Scan# 1893
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:165 Resp: 94872
Ion Ratio Lower Upper
165 100
89 67.6 52.8 79.2
121 25.6 19.4 29.0









#50
Acenaphthylene
Concen: 9.992 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:152 Resp: 858640
Ion Ratio Lower Upper
152 100
151 19.9 15.7 23.5
153 12.8 10.5 15.7



#51
3-Nitroaniline
Concen: 7.324 ng/ul
RT: 14.545 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:138 Resp: 106870
Ion Ratio Lower Upper
138 100
108 18.7 12.5 18.7
92 120.1 95.4 143.2





#52

Acenaphthene

Concen: 9.798 ng/ul

RT: 14.715 min Scan# 1977

Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Instrument :

BNA_M

ClientSampleId :

SSTD010020





Tgt Ion:153 Resp: 600171

Ion Ratio Lower Upper

153 100

152 49.0 38.2 57.2

154 88.8 71.0 106.6



#53

2,4-Dinitrophenol

Concen: 4.793 ng/ul

RT: 14.745 min Scan# 1982

Delta R.T. -0.006 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14




Tgt Ion:184 Resp: 31030


Ion Ratio Lower Upper

184 100

63 77.4 58.8 88.2

154 69.2 55.0 82.4









#55
4-Nitrophenol
Concen: 8.645 ng/ul
RT: 14.833 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:109 Resp: 93710
Ion Ratio Lower Upper
109 100
139 126.2 102.2 153.4
65 129.4 105.7 158.5



#56
Dibenzofuran
Concen: 9.760 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:168 Resp: 814150
Ion Ratio Lower Upper
168 100
139 38.4 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 7.093 ng/ul
RT: 14.998 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:165 Resp: 138564
Ion Ratio Lower Upper
165 100
63 50.6 39.7 59.5
182 2.5 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.740 ng/ul
RT: 15.268 min Scan# 2071
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:232 Resp: 138267
Ion Ratio Lower Upper
232 100
131 47.7 39.1 58.7
130 2.3 1.8 2.6
166 32.1 25.4 38.0









#59
Diethylphthalate
Concen: 9.257 ng/ul
RT: 15.468 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:149 Resp: 673297
Ion Ratio Lower Upper
149 100
177 21.9 17.3 25.9
150 12.3 10.1 15.1



#61
Fluorene
Concen: 9.702 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:166 Resp: 687712
Ion Ratio Lower Upper
166 100
165 97.6 78.5 117.7
167 13.6 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 9.972 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:204 Resp: 331807
Ion Ratio Lower Upper
204 100
206 32.1 25.7 38.5
141 58.7 49.4 74.0



#63
4-Nitroaniline
Concen: 6.842 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:138 Resp: 97601
Ion Ratio Lower Upper
138 100
92 52.1 42.1 63.1
108 85.2 66.2 99.2









#66
4,6-Dinitro-2-methylphenol
Concen: 5.238 ng/ul
RT: 15.757 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:198 Resp: 55194
Ion Ratio Lower Upper
198 100
182 6.4 4.2 6.2#
77 30.3 24.0 36.0



#67
N-Nitrosodiphenylamine
Concen: 10.168 ng/ul
RT: 15.898 min Scan# 2178
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:169 Resp: 562617
Ion Ratio Lower Upper
169 100
168 67.1 54.6 81.8
167 36.8 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 10.023 ng/ul
RT: 16.580 min Scan# 2294
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:248 Resp: 188455
Ion Ratio Lower Upper
248 100
250 96.9 78.9 118.3
141 78.3 65.7 98.5



#69
Hexachlorobenzene
Concen: 10.058 ng/ul
RT: 16.692 min Scan# 2313
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:284 Resp: 224040
Ion Ratio Lower Upper
284 100
142 35.4 30.8 46.2
249 30.3 24.2 36.2









#70
Atrazine
Concen: 8.946 ng/ul
RT: 16.845 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:200 Resp: 183351
Ion Ratio Lower Upper
200 100
173 26.6 21.8 32.6
215 47.6 39.4 59.2



#71
Pentachlorophenol
Concen: 8.377 ng/ul
RT: 17.027 min Scan# 2370
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:266 Resp: 112158
Ion Ratio Lower Upper
266 100
264 61.3 49.7 74.5
268 61.6 51.8 77.6









#72
Phenanthrene
Concen: 9.953 ng/u1
RT: 17.427 min Scan# 2438
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:178 Resp: 1054203
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 19.3 15.7 23.5



#74
Anthracene
Concen: 10.016 ng/u1
RT: 17.521 min Scan# 2454
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:178 Resp: 1076119
Ion Ratio Lower Upper
178 100
179 15.7 12.3 18.5
176 18.5 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 10.316 ng/uL
RT: 13.457 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:216 Resp: 258993
Ion Ratio Lower Upper
216 100
214 78.5 62.6 94.0
218 49.1 38.2 57.2



#76
Pentachlorobenzene
Concen: 10.141 ng/uL
RT: 14.963 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:250 Resp: 280913
Ion Ratio Lower Upper
250 100
248 63.9 50.2 75.2
252 65.0 50.5 75.7









#77
 Carbazole
 Concen: 9.394 ng/ul
 RT: 17.786 min Scan# 2499
 Delta R.T. -0.000 min
 Lab File: BM036549.D
 Acq: 15 Sep 2022 17:14




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010020


Tgt Ion:167 Resp: 961574
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.2 25.8
 139 13.2 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 9.522 ng/ul
 RT: 18.356 min Scan# 2596
 Delta R.T. -0.006 min
 Lab File: BM036549.D
 Acq: 15 Sep 2022 17:14

Tgt Ion:149 Resp: 1173384
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.2 10.8
 104 5.6 4.3 6.5









#80
Fluoranthene
Concen: 9.828 ng/ul
RT: 19.433 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:202 Resp: 1183237
Ion Ratio Lower Upper
202 100
101 11.9 9.8 14.6
100 8.8 7.4 11.2



#82
Pyrene
Concen: 9.864 ng/ul
RT: 19.792 min Scan# 2840
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:202 Resp: 1259242
Ion Ratio Lower Upper
202 100
101 14.3 10.9 16.3
100 11.4 8.7 13.1









#83
Butylbenzylphthalate
Concen: 7.838 ng/ul
RT: 20.692 min Scan# 2994
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:149 Resp: 427860
Ion Ratio Lower Upper
149 100
91 77.0 55.2 82.8
206 18.5 16.2 24.2



#84
3,3'-Dichlorobenzidine
Concen: 6.986 ng/ul
RT: 21.468 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:252 Resp: 280391
Ion Ratio Lower Upper
252 100
254 65.2 53.0 79.4
126 11.8 10.4 15.6









#85
Benzo(a)anthracene
Concen: 9.837 ng/ul
RT: 21.539 min Scan# 3137
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:228 Resp: 1182168
Ion Ratio Lower Upper
228 100
229 19.6 15.8 23.8
226 27.1 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 8.048 ng/ul
RT: 21.474 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:149 Resp: 656990
Ion Ratio Lower Upper
149 100
167 27.7 22.0 33.0
279 4.1 3.3 4.9





#87

Chrysene

Concen: 10.156 ng/ul

RT: 21.592 min Scan# 31

Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Instrument :

BNA_M

ClientSampleId :

SSTD010020





Tgt Ion:228 Resp: 1133915

Ion Ratio Lower Upper

228 100

226 29.8 24.4 36.6

229 19.6 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 6.752 ng/ul




RT: 22.421 min Scan# 3287


Delta R.T. -0.000 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Tgt Ion:149 Resp: 1018787









#90
Benzo(b)fluoranthene
Concen: 9.143 ng/ul
RT: 23.250 min Scan# 3428
Delta R.T. -0.000 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14




Instrument :
BNA_M
ClientSampleId :
SSTD010020


Tgt Ion:252 Resp: 1159799
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 10.5 9.1 13.7



#91
Benzo(k)fluoranthene
Concen: 9.556 ng/ul
RT: 23.297 min Scan# 3436
Delta R.T. -0.006 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:252 Resp: 1148022
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 11.2 8.2 12.4





#93

Benzo(a)pyrene

Concen: 9.570 ng/ul

RT: 23.885 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14

Instrument :

BNA_M

ClientSampleId :

SSTD010020





Tgt Ion:252 Resp: 1000772

Ion Ratio Lower Upper

252 100

253 21.8 17.5 26.3

125 11.7 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 11.049 ng/ul

RT: 26.538 min Scan# 3987

Delta R.T. -0.012 min

Lab File: BM036549.D

Acq: 15 Sep 2022 17:14




Tgt Ion:276 Resp: 1166175


Ion Ratio Lower Upper

276 100

138 24.4 20.2 30.4

277 24.6 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 10.984 ng/ul
RT: 26.562 min Scan# 3991
Delta R.T. -0.012 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Instrument :
BNA_M
ClientSampleId :
SSTD010020

Tgt Ion:278 Resp: 1054616
Ion Ratio Lower Upper
278 100
139 17.9 14.5 21.7
279 24.2 19.7 29.5



#96
Benzo(g,h,i)perylene
Concen: 11.445 ng/ul
RT: 27.321 min Scan# 4120
Delta R.T. -0.012 min
Lab File: BM036549.D
Acq: 15 Sep 2022 17:14

Tgt Ion:276 Resp: 1032545
Ion Ratio Lower Upper
276 100
138 23.2 18.7 28.1
277 22.8 19.2 28.8

