

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036550.D
 Acq On : 15 Sep 2022 17:50
 Operator : CG/JU
 Sample : SSTD02074
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020021

Quant Time: Sep 16 01:51:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 01:29:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	336372	20.000	ng/u1	0.00
20) Naphthalene-d8	10.845	136	1422198	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	898704	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	1848925	20.000	ng/u1	0.00
79) Chrysene-d12	21.556	240	1918023	20.000	ng/u1	0.00
88) Perylene-d12	23.997	264	1921660	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	71504	8.182	ng/uL	0.00
4) Pyridine-d5	3.816	84	495345	18.163	ng/u1	0.00
7) Phenol-d5	7.175	99	548000	18.505	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	352200	18.384	ng/u1	0.00
11) 2-Chlorophenol-d4	7.551	132	422783	18.541	ng/u1	0.00
15) 4-Methylphenol-d8	8.734	113	407590	17.412	ng/u1	0.00
21) Nitrobenzene-d5	9.192	128	181532	15.981	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	156174	13.534	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.457	165	389186	19.094	ng/u1	0.00
31) 4-Chloroaniline-d4	10.975	131	623810	18.686	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	1182382	19.642	ng/u1	0.00
49) Acenaphthylene-d8	14.351	160	1399809	19.452	ng/u1	0.00
54) 4-Nitrophenol-d4	14.827	143	207436	17.859	ng/u1	0.00
60) Fluorene-d10	15.639	176	1065891	19.112	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	105900	10.975	ng/u1	0.00
73) Anthracene-d10	17.486	188	1650935	19.723	ng/u1	0.00
81) Pyrene-d10	19.768	212	1825935	18.228	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.838	264	1843202	19.086	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	77988	8.523	ng/uL	100
5) Pyridine	3.834	79	497587	18.357	ng/u1	100
6) Benzaldehyde	7.157	77	315945	22.896	ng/u1	100
8) Phenol	7.204	94	594856	18.913	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.445	93	476743	18.825	ng/u1	100
12) 2-Chlorophenol	7.587	128	457876	18.701	ng/u1	100
13) 2-Methylphenol	8.463	108	436893	17.984	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.563	45	582201	15.309	ng/u1	100
16) Acetophenone	8.857	105	730234	18.456	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.839	70	360880	18.590	ng/u1	100
18) 4-Methylphenol	8.792	108	475826	18.115	ng/u1	100
19) Hexachloroethane	9.116	117	184379	17.656	ng/u1	100
22) Nitrobenzene	9.234	77	498063	17.241	ng/u1	100
23) Isophorone	9.763	82	1004651	19.490	ng/u1	100
25) 2-Nitrophenol	9.951	139	199923	15.446	ng/u1	100
26) 2,4-Dimethylphenol	10.004	107	516556	19.451	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.251	93	609371	20.134	ng/u1	100
29) 2,4-Dichlorophenol	10.480	162	413859	19.689	ng/u1	100
30) Naphthalene	10.892	128	1534629	19.789	ng/u1	100
32) 4-Chloroaniline	10.998	127	647364	18.834	ng/u1	100
33) Hexachlorobutadiene	11.180	225	253417	18.820	ng/u1	100
34) Caprolactam	11.757	113	62786	9.305	ng/u1	100
35) 4-Chloro-3-methylphenol	12.110	107	451419	20.319	ng/u1	100
36) 2-Methylnaphthalene	12.492	142	1008792	20.226	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036550.D
 Acq On : 15 Sep 2022 17:50
 Operator : CG/JU
 Sample : SSTD02074
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020021

Quant Time: Sep 16 01:51:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 01:29:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.710	142	1014449	20.272	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.857	216	498292	19.491	ng/ul	100
40) Hexachlorocyclopentadiene	12.839	237	228429	17.071	ng/ul	100
41) 2,4,6-Trichlorophenol	13.092	196	299417	17.861	ng/ul	100
42) 2,4,5-Trichlorophenol	13.157	196	327168	18.490	ng/ul	100
43) 1,1'-Biphenyl	13.498	154	1302685	19.463	ng/ul	100
44) 2-Chloronaphthalene	13.533	162	1022505	19.654	ng/ul	100
45) 2-Nitroaniline	13.733	65	243464	15.498	ng/ul	100
47) Dimethylphthalate	14.110	163	1263163	19.926	ng/ul	100
48) 2,6-Dinitrotoluene	14.227	165	197875	15.981	ng/ul	100
50) Acenaphthylene	14.374	152	1622654	19.854	ng/ul	100
51) 3-Nitroaniline	14.551	138	225227	16.229	ng/ul	100
52) Acenaphthene	14.716	153	1130033	19.398	ng/ul	100
53) 2,4-Dinitrophenol	14.751	184	62953	10.225	ng/ul	100
55) 4-Nitrophenol	14.839	109	184464	17.893	ng/ul	100
56) Dibenzofuran	15.051	168	1527664	19.256	ng/ul	100
57) 2,4-Dinitrotoluene	15.004	165	294438	15.848	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.268	232	270463	17.976	ng/ul	100
59) Diethylphthalate	15.468	149	1255954	18.156	ng/ul	100
61) Fluorene	15.698	166	1306083	19.374	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.692	204	613424	19.385	ng/ul	100
63) 4-Nitroaniline	15.704	138	232733	17.154	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.763	198	125468	12.433	ng/ul	100
67) N-Nitrosodiphenylamine	15.904	169	1063412	20.069	ng/ul	100
68) 4-Bromophenyl-phenylether	16.580	248	351839	19.541	ng/ul	100
69) Hexachlorobenzene	16.692	284	417573	19.576	ng/ul	100
70) Atrazine	16.851	200	351890	17.928	ng/ul	100
71) Pentachlorophenol	17.033	266	221591	17.284	ng/ul	100
72) Phenanthrene	17.433	178	2015010	19.866	ng/ul	100
74) Anthracene	17.521	178	2042914	19.857	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.457	216	492680	20.493	ng/ul	100
76) Pentachlorobenzene	14.968	250	534780	20.160	ng/ul	100
77) Carbazole	17.786	167	1833780	18.708	ng/ul	100
78) Di-n-butylphthalate	18.362	149	2081678	17.640	ng/ul	100
80) Fluoranthene	19.433	202	2262068	18.408	ng/ul	100
82) Pyrene	19.798	202	2423714	18.600	ng/ul	100
83) Butylbenzylphthalate	20.697	149	1078103	19.348	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.468	252	714477	17.441	ng/ul	100
85) Benzo(a)anthracene	21.539	228	2375120	19.363	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.480	149	1310973	15.734	ng/ul	100
87) Chrysene	21.592	228	2259256	19.824	ng/ul	100
89) Di-n-octyl phthalate	22.421	149	2066454	13.102	ng/ul	100
90) Benzo(b)fluoranthene	23.250	252	2428153	18.312	ng/ul	100
91) Benzo(k)fluoranthene	23.303	252	2354389	18.747	ng/ul	100
93) Benzo(a)pyrene	23.891	252	2105461	19.261	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.550	276	2610465	23.661	ng/ul	100
95) Dibenzo(a,h)anthracene	26.574	278	2323655	23.151	ng/ul	100
96) Benzo(g,h,i)perylene	27.332	276	2263866	24.004	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.523 ng/uL
RT: 3.428 min Scan# 58
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion: 88 Resp: 77988
Ion Ratio Lower Upper
88 100
43 28.9 23.1 34.7
58 70.8 56.6 85.0



#5
Pyridine
Concen: 18.357 ng/uL
RT: 3.834 min Scan# 127
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion: 79 Resp: 497587
Ion Ratio Lower Upper
79 100
52 71.5 57.2 85.8
51 33.5 26.8 40.2









#6
Benzaldehyde
Concen: 22.896 ng/ul
RT: 7.157 min Scan# 692
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion: 77 Resp: 315945
Ion Ratio Lower Upper
77 100
105 93.9 75.1 112.7
106 91.3 73.0 109.6



#8
Phenol
Concen: 18.913 ng/ul
RT: 7.204 min Scan# 700
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion: 94 Resp: 594856
Ion Ratio Lower Upper
94 100
65 27.8 22.2 33.4
66 36.1 28.9 43.3









#10
Bis(2-Chloroethyl)ether
Concen: 18.825 ng/ul
RT: 7.445 min Scan# 741
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion: 93 Resp: 476743
Ion Ratio Lower Upper
93 100
63 73.9 59.1 88.7
95 32.6 26.1 39.1



#12
2-Chlorophenol
Concen: 18.701 ng/ul
RT: 7.587 min Scan# 765
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:128 Resp: 457876
Ion Ratio Lower Upper
128 100
64 47.3 37.8 56.8
130 32.5 26.0 39.0





#13

2-Methylphenol

Concen: 17.984 ng/ul

RT: 8.463 min Scan# 914

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :





SSTD020021

Tgt Ion:108 Resp: 436893

Ion Ratio Lower Upper

108 100

107 88.4 70.7 106.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 15.309 ng/ul

RT: 8.563 min Scan# 931

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50




Tgt Ion: 45 Resp: 582201


Ion Ratio Lower Upper

45 100

77 16.4 13.1 19.7

79 11.9 9.5 14.3









#16
Acetophenone
Concen: 18.456 ng/ul
RT: 8.857 min Scan# 981
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:105 Resp: 730234
Ion Ratio Lower Upper
105 100
77 77.0 61.6 92.4
51 25.6 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 18.590 ng/ul
RT: 8.839 min Scan# 978
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion: 70 Resp: 360880
Ion Ratio Lower Upper
70 100
42 53.7 43.0 64.4
101 9.6 7.7 11.5
130 18.9 15.1 22.7









#18
4-Methylphenol
Concen: 18.115 ng/ul
RT: 8.792 min Scan# 970
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:108 Resp: 475826
Ion Ratio Lower Upper
108 100
107 111.7 89.4 134.0



#19
Hexachloroethane
Concen: 17.656 ng/ul
RT: 9.116 min Scan# 1025
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:117 Resp: 184379
Ion Ratio Lower Upper
117 100
201 85.9 68.7 103.1
199 53.8 43.0 64.6









#22
Nitrobenzene
Concen: 17.241 ng/ul
RT: 9.234 min Scan# 1045
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion: 77 Resp: 498063
Ion Ratio Lower Upper
77 100
123 43.6 34.9 52.3
65 13.4 10.7 16.1



#23
Isophorone
Concen: 19.490 ng/ul
RT: 9.763 min Scan# 1135
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion: 82 Resp: 1004651
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 16.9 13.5 20.3









#25
2-Nitrophenol
Concen: 15.446 ng/ul
RT: 9.951 min Scan# 1167
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:139 Resp: 199923
Ion Ratio Lower Upper
139 100
65 51.1 40.9 61.3
109 29.9 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 19.451 ng/ul
RT: 10.004 min Scan# 1176
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:107 Resp: 516556
Ion Ratio Lower Upper
107 100
121 49.3 39.4 59.2
122 86.4 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.134 ng/ul
 RT: 10.251 min Scan# 1218
 Delta R.T. 0.000 min
 Lab File: BM036550.D
 Acq: 15 Sep 2022 17:50




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020021


Tgt Ion: 93 Resp: 609371
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.2 39.4
 123 11.5 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 19.689 ng/ul
 RT: 10.480 min Scan# 1257
 Delta R.T. 0.000 min
 Lab File: BM036550.D
 Acq: 15 Sep 2022 17:50

Tgt Ion: 162 Resp: 413859
 Ion Ratio Lower Upper
 162 100
 164 65.1 52.1 78.1
 98 36.4 29.1 43.7









#30
Naphthalene
Concen: 19.789 ng/ul
RT: 10.892 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:128 Resp: 1534629
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.9 10.3 15.5



#32
4-Chloroaniline
Concen: 18.834 ng/ul
RT: 10.998 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:127 Resp: 647364
Ion Ratio Lower Upper
127 100
129 31.8 25.4 38.2





#33

Hexachlorobutadiene

Concen: 18.820 ng/ul

RT: 11.180 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:225 Resp: 253417

Ion Ratio Lower Upper

225 100

223 62.3 49.8 74.8

227 63.8 51.0 76.6



#34

Caprolactam

Concen: 9.305 ng/ul

RT: 11.757 min Scan# 1474

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50




Tgt Ion:113 Resp: 62786


Ion Ratio Lower Upper

113 100

55 159.4 127.5 191.3

56 118.3 94.6 142.0









#35
4-Chloro-3-methylphenol
Concen: 20.319 ng/ul
RT: 12.110 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:107 Resp: 451419
Ion Ratio Lower Upper
107 100
144 26.9 21.5 32.3
142 79.9 63.9 95.9



#36
2-Methylnaphthalene
Concen: 20.226 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:142 Resp: 1008792
Ion Ratio Lower Upper
142 100
141 88.4 70.7 106.1





#37

1-Methylnaphthalene

Concen: 20.272 ng/ul

RT: 12.710 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:142 Resp: 1014449

Ion Ratio Lower Upper

142 100

141 91.8 73.4 110.2

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.491 ng/ul

RT: 12.857 min Scan# 1661

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Tgt Ion:216 Resp: 498292




Ion Ratio Lower Upper


216 100

214 78.0 62.4 93.6

179 20.9 16.7 25.1

108 19.2 15.4 23.0









#40
Hexachlorocyclopentadiene
Concen: 17.071 ng/ul
RT: 12.839 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:237 Resp: 228429
Ion Ratio Lower Upper
237 100
235 63.6 50.9 76.3
272 13.6 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 17.861 ng/ul
RT: 13.092 min Scan# 1701
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:196 Resp: 299417
Ion Ratio Lower Upper
196 100
198 96.6 77.3 115.9
200 31.5 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 18.490 ng/ul
RT: 13.157 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:196 Resp: 327168
Ion Ratio Lower Upper
196 100
198 94.5 75.6 113.4
200 30.9 24.7 37.1



#43
1,1'-Biphenyl
Concen: 19.463 ng/ul
RT: 13.498 min Scan# 1770
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:154 Resp: 1302685
Ion Ratio Lower Upper
154 100
153 39.3 31.4 47.2
76 13.0 10.4 15.6









#44
2-Chloronaphthalene
Concen: 19.654 ng/ul
RT: 13.533 min Scan# 1776
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:162 Resp: 1022505
Ion Ratio Lower Upper
162 100
164 32.7 26.2 39.2
127 40.8 32.6 49.0



#45
2-Nitroaniline
Concen: 15.498 ng/ul
RT: 13.733 min Scan# 1810
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion: 65 Resp: 243464
Ion Ratio Lower Upper
65 100
92 68.0 54.4 81.6
138 107.6 86.1 129.1









#47
Dimethylphthalate
Concen: 19.926 ng/ul
RT: 14.110 min Scan# 1874
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:163 Resp: 1263163
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.6
164 10.4 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 15.981 ng/ul
RT: 14.227 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:165 Resp: 197875
Ion Ratio Lower Upper
165 100
89 66.0 52.8 79.2
121 24.2 19.4 29.0









#50
Acenaphthylene
Concen: 19.854 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:152 Resp: 1622654
Ion Ratio Lower Upper
152 100
151 19.6 15.7 23.5
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 16.229 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:138 Resp: 225227
Ion Ratio Lower Upper
138 100
108 15.6 12.5 18.7
92 119.3 95.4 143.2









#52
Acenaphthene
Concen: 19.398 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:153 Resp: 1130033
Ion Ratio Lower Upper
153 100
152 47.7 38.2 57.2
154 88.8 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 10.225 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:184 Resp: 62953
Ion Ratio Lower Upper
184 100
63 73.5 58.8 88.2
154 68.7 55.0 82.4









#55
4-Nitrophenol
Concen: 17.893 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:109 Resp: 184464
Ion Ratio Lower Upper
109 100
139 127.8 102.2 153.4
65 132.1 105.7 158.5



#56
Dibenzofuran
Concen: 19.256 ng/ul
RT: 15.051 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:168 Resp: 1527664
Ion Ratio Lower Upper
168 100
139 37.1 29.7 44.5





#57

2,4-Dinitrotoluene

Concen: 15.848 ng/ul

RT: 15.004 min Scan# 2026

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:165 Resp: 294438

Ion Ratio Lower Upper

165 100

63 49.6 39.7 59.5

182 2.6 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 17.976 ng/ul

RT: 15.268 min Scan# 2071

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Tgt Ion:232 Resp: 270463




Ion Ratio Lower Upper


232 100

131 48.9 39.1 58.7

130 2.2 1.8 2.6

166 31.7 25.4 38.0









#59
Diethylphthalate
Concen: 18.156 ng/ul
RT: 15.468 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:149 Resp: 1255954
Ion Ratio Lower Upper
149 100
177 21.6 17.3 25.9
150 12.6 10.1 15.1



#61
Fluorene
Concen: 19.374 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:166 Resp: 1306083
Ion Ratio Lower Upper
166 100
165 98.1 78.5 117.7
167 13.0 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 19.385 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:204 Resp: 613424
Ion Ratio Lower Upper
204 100
206 32.1 25.7 38.5
141 61.7 49.4 74.0



#63
4-Nitroaniline
Concen: 17.154 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:138 Resp: 232733
Ion Ratio Lower Upper
138 100
92 52.6 42.1 63.1
108 82.7 66.2 99.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 12.433 ng/ul
 RT: 15.763 min Scan# 2155
 Delta R.T. 0.000 min
 Lab File: BM036550.D
 Acq: 15 Sep 2022 17:50




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020021


Tgt Ion:198 Resp: 125468
 Ion Ratio Lower Upper
 198 100
 182 5.2 4.2 6.2
 77 30.0 24.0 36.0



#67
 N-Nitrosodiphenylamine
 Concen: 20.069 ng/ul
 RT: 15.904 min Scan# 2179
 Delta R.T. 0.000 min
 Lab File: BM036550.D
 Acq: 15 Sep 2022 17:50

Tgt Ion:169 Resp: 1063412
 Ion Ratio Lower Upper
 169 100
 168 68.2 54.6 81.8
 167 37.0 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 19.541 ng/ul
RT: 16.580 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:248 Resp: 351839
Ion Ratio Lower Upper
248 100
250 98.6 78.9 118.3
141 82.1 65.7 98.5



#69
Hexachlorobenzene
Concen: 19.576 ng/ul
RT: 16.692 min Scan# 2313
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:284 Resp: 417573
Ion Ratio Lower Upper
284 100
142 38.5 30.8 46.2
249 30.2 24.2 36.2





#70

Atrazine

Concen: 17.928 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:200 Resp: 351890

Ion Ratio Lower Upper

200 100

173 27.2 21.8 32.6

215 49.3 39.4 59.2



#71

Pentachlorophenol

Concen: 17.284 ng/ul

RT: 17.033 min Scan# 2371

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50




Tgt Ion:266 Resp: 221591


Ion Ratio Lower Upper

266 100

264 62.1 49.7 74.5

268 64.7 51.8 77.6









#72
Phenanthrene
Concen: 19.866 ng/ul
RT: 17.433 min Scan# 2439
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:178 Resp: 2015010
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 19.6 15.7 23.5



#74
Anthracene
Concen: 19.857 ng/ul
RT: 17.521 min Scan# 2454
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:178 Resp: 2042914
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.8 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.493 ng/uL
RT: 13.457 min Scan# 1763
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50




Instrument :
BNA_M
ClientSampleId :
SSTD020021


Tgt Ion:216 Resp: 492680
Ion Ratio Lower Upper
216 100
214 78.3 62.6 94.0
218 47.7 38.2 57.2



#76
Pentachlorobenzene
Concen: 20.160 ng/uL
RT: 14.968 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BM036550.D
Acq: 15 Sep 2022 17:50

Tgt Ion:250 Resp: 534780
Ion Ratio Lower Upper
250 100
248 62.7 50.2 75.2
252 63.1 50.5 75.7









#77
 Carbazole
 Concen: 18.708 ng/ul
 RT: 17.786 min Scan# 2499
 Delta R.T. 0.000 min
 Lab File: BM036550.D
 Acq: 15 Sep 2022 17:50




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020021


Tgt Ion:167 Resp: 1833780
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.2 25.8
 139 13.2 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 17.640 ng/ul
 RT: 18.362 min Scan# 2597
 Delta R.T. 0.000 min
 Lab File: BM036550.D
 Acq: 15 Sep 2022 17:50

Tgt Ion:149 Resp: 2081678
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.2 10.8
 104 5.4 4.3 6.5









#80
 Fluoranthene
 Concen: 18.408 ng/u1
 RT: 19.433 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BM036550.D
 Acq: 15 Sep 2022 17:50




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020021


Tgt Ion:202 Resp: 2262068
 Ion Ratio Lower Upper
 202 100
 101 12.2 9.8 14.6
 100 9.3 7.4 11.2



#82
 Pyrene
 Concen: 18.600 ng/u1
 RT: 19.798 min Scan# 2841
 Delta R.T. 0.000 min
 Lab File: BM036550.D
 Acq: 15 Sep 2022 17:50

Tgt Ion:202 Resp: 2423714
 Ion Ratio Lower Upper
 202 100
 101 13.6 10.9 16.3
 100 10.9 8.7 13.1





#83

Butylbenzylphthalate

Concen: 19.348 ng/ul

RT: 20.697 min Scan# 2994

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:149 Resp: 1078103

Ion Ratio Lower Upper

149 100

91 69.0 55.2 82.8

206 20.2 16.2 24.2



#84

3,3'-Dichlorobenzidine

Concen: 17.441 ng/ul

RT: 21.468 min Scan# 3125

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50




Tgt Ion:252 Resp: 714477


Ion Ratio Lower Upper

252 100

254 66.2 53.0 79.4

126 13.0 10.4 15.6





#85

Benzo(a)anthracene

Concen: 19.363 ng/ul

RT: 21.539 min Scan# 3137

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:228 Resp: 2375120

Ion Ratio Lower Upper

228 100

229 19.8 15.8 23.8

226 27.4 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 15.734 ng/ul

RT: 21.480 min Scan# 3127

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50




Tgt Ion:149 Resp: 1310973


Ion Ratio Lower Upper

149 100

167 27.5 22.0 33.0

279 4.1 3.3 4.9





#87

Chrysene

Concen: 19.824 ng/ul

RT: 21.592 min Scan# 3146

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:228 Resp: 2259256

Ion Ratio Lower Upper

228 100

226 30.5 24.4 36.6

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 13.102 ng/ul




RT: 22.421 min Scan# 3287


Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Tgt Ion:149 Resp: 2066454





#90

Benzo(b)fluoranthene

Concen: 18.312 ng/ul

RT: 23.250 min Scan# 3428

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:252 Resp: 2428153

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 11.4 9.1 13.7



#91

Benzo(k)fluoranthene

Concen: 18.747 ng/ul

RT: 23.303 min Scan# 3437

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50




Tgt Ion:252 Resp: 2354389


Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8

125 10.3 8.2 12.4





#93

Benzo(a)pyrene

Concen: 19.261 ng/ul

RT: 23.891 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:252 Resp: 2105461

Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 11.9 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 23.661 ng/ul

RT: 26.550 min Scan# 3989

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50




Tgt Ion:276 Resp: 2610465


Ion Ratio Lower Upper

276 100

138 25.3 20.2 30.4

277 24.6 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 23.151 ng/ul

RT: 26.574 min Scan# 3993

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Instrument :

BNA_M

ClientSampleId :

SSTD020021





Tgt Ion:278 Resp: 2323655

Ion Ratio Lower Upper

278 100

139 18.1 14.5 21.7

279 24.6 19.7 29.5



#96

Benzo(g,h,i)perylene

Concen: 24.004 ng/ul

RT: 27.332 min Scan# 4122

Delta R.T. 0.000 min

Lab File: BM036550.D

Acq: 15 Sep 2022 17:50

Tgt Ion:276 Resp: 2263866

Ion Ratio Lower Upper

276 100

138 23.4 18.7 28.1

277 24.0 19.2 28.8

