

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036551.D
 Acq On : 15 Sep 2022 18:27
 Operator : CG/JU
 Sample : SSTD04075
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040022

Quant Time: Sep 16 01:52:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 01:29:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	352445	20.000	ng/u1	0.00
20) Naphthalene-d8	10.845	136	1487086	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.657	164	924514	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	1895378	20.000	ng/u1	0.00
79) Chrysene-d12	21.556	240	2072449	20.000	ng/u1	0.00
88) Perylene-d12	24.003	264	2060321	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	155349	16.966	ng/uL	0.00
4) Pyridine-d5	3.810	84	1144151	40.040	ng/u1	0.00
7) Phenol-d5	7.175	99	1294934	41.734	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.357	67	788710	39.291	ng/u1	0.00
11) 2-Chlorophenol-d4	7.557	132	986738	41.300	ng/u1	0.00
15) 4-Methylphenol-d8	8.734	113	971043	39.591	ng/u1	0.00
21) Nitrobenzene-d5	9.198	128	451961	38.052	ng/u1	0.00
24) 2-Nitrophenol-d4	9.922	143	416441	34.515	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.457	165	925216	43.412	ng/u1	0.00
31) 4-Chloroaniline-d4	10.975	131	1423261	40.774	ng/u1	0.00
46) Dimethylphthalate-d6	14.069	166	2670820	43.130	ng/u1	0.00
49) Acenaphthylene-d8	14.351	160	3193593	43.140	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	511467	42.805	ng/u1	0.00
60) Fluorene-d10	15.645	176	2391693	41.687	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	310755	31.417	ng/u1	0.00
73) Anthracene-d10	17.492	188	3726434	43.427	ng/u1	0.00
81) Pyrene-d10	19.768	212	4192906	38.739	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.844	264	4430413	42.788	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	168716	17.597	ng/uL	99
5) Pyridine	3.834	79	1166814	41.082	ng/u1	98
6) Benzaldehyde	7.163	77	693816	47.988	ng/u1	95
8) Phenol	7.204	94	1392925	42.267	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.451	93	1083808	40.843	ng/u1	100
12) 2-Chlorophenol	7.587	128	1063946	41.473	ng/u1	99
13) 2-Methylphenol	8.469	108	1024328	40.243	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.563	45	1292631	32.440	ng/u1	99
16) Acetophenone	8.857	105	1676889	40.448	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.851	70	841316	41.361	ng/u1	97
18) 4-Methylphenol	8.804	108	1113831	40.471	ng/u1	96
19) Hexachloroethane	9.122	117	417245	38.132	ng/u1	94
22) Nitrobenzene	9.239	77	1178336	39.009	ng/u1	98
23) Isophorone	9.769	82	2310700	42.871	ng/u1	99
25) 2-Nitrophenol	9.951	139	511393	37.787	ng/u1	99
26) 2,4-Dimethylphenol	10.010	107	1193356	42.975	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.251	93	1373216	43.391	ng/u1	100
29) 2,4-Dichlorophenol	10.486	162	969821	44.126	ng/u1	99
30) Naphthalene	10.898	128	3433796	42.346	ng/u1	100
32) 4-Chloroaniline	10.998	127	1470161	40.906	ng/u1	100
33) Hexachlorobutadiene	11.186	225	568708	40.393	ng/u1	100
34) Caprolactam	11.763	113	144381	20.463	ng/u1	97
35) 4-Chloro-3-methylphenol	12.110	107	1063738	45.792	ng/u1	98
36) 2-Methylnaphthalene	12.498	142	2308769	44.269	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036551.D
 Acq On : 15 Sep 2022 18:27
 Operator : CG/JU
 Sample : SSTD04075
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD040022


Quant Time: Sep 16 01:52:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 01:29:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.716	142	2316187	44.265	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.857	216	1135782	43.186	ng/ul	99
40) Hexachlorocyclopentadiene	12.845	237	583691	42.403	ng/ul	99
41) 2,4,6-Trichlorophenol	13.092	196	712427	41.312	ng/ul	98
42) 2,4,5-Trichlorophenol	13.163	196	768128	42.200	ng/ul	98
43) 1,1'-Biphenyl	13.498	154	2918549	42.388	ng/ul	99
44) 2-Chloronaphthalene	13.539	162	2299627	42.967	ng/ul	99
45) 2-Nitroaniline	13.733	65	626159	38.745	ng/ul	99
47) Dimethylphthalate	14.116	163	2832293	43.432	ng/ul	99
48) 2,6-Dinitrotoluene	14.227	165	515472	40.469	ng/ul	98
50) Acenaphthylene	14.380	152	3632078	43.199	ng/ul	99
51) 3-Nitroaniline	14.557	138	571136	40.006	ng/ul	95
52) Acenaphthene	14.722	153	2529560	42.211	ng/ul	100
53) 2,4-Dinitrophenol	14.757	184	191798	30.284	ng/ul	96
55) 4-Nitrophenol	14.851	109	435149	41.030	ng/ul	97
56) Dibenzofuran	15.051	168	3351476	41.065	ng/ul	98
57) 2,4-Dinitrotoluene	15.010	165	745342	38.998	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.269	232	640634	41.390	ng/ul	97
59) Diethylphthalate	15.474	149	2851381	40.069	ng/ul	100
61) Fluorene	15.698	166	2941728	42.418	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.692	204	1395621	42.872	ng/ul	99
63) 4-Nitroaniline	15.710	138	583285	41.792	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.768	198	361251	34.920	ng/ul#	94
67) N-Nitrosodiphenylamine	15.904	169	2386023	43.926	ng/ul	99
68) 4-Bromophenyl-phenylether	16.586	248	798315	43.251	ng/ul	96
69) Hexachlorobenzene	16.692	284	934309	42.727	ng/ul	98
70) Atrazine	16.851	200	818221	40.666	ng/ul	98
71) Pentachlorophenol	17.033	266	544796	41.451	ng/ul	99
72) Phenanthrene	17.433	178	4459050	42.885	ng/ul	99
74) Anthracene	17.521	178	4562382	43.259	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.463	216	1113253	45.171	ng/ul	99
76) Pentachlorobenzene	14.969	250	1181691	43.454	ng/ul	99
77) Carbazole	17.786	167	4070173	40.505	ng/ul	99
78) Di-n-butylphthalate	18.362	149	4907089	40.563	ng/ul	99
80) Fluoranthene	19.439	202	5161696	38.874	ng/ul	99
82) Pyrene	19.798	202	5450284	38.709	ng/ul	97
83) Butylbenzylphthalate	20.698	149	2156834	35.824	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.468	252	1901271	42.952	ng/ul	98
85) Benzo(a)anthracene	21.539	228	5538191	41.786	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.474	149	3441673	38.227	ng/ul	99
87) Chrysene	21.598	228	5225043	42.431	ng/ul	99
89) Di-n-octyl phthalate	22.421	149	5527084	32.685	ng/ul	100
90) Benzo(b)fluoranthene	23.256	252	5596622	39.367	ng/ul	100
91) Benzo(k)fluoranthene	23.309	252	5692990	42.280	ng/ul	99
93) Benzo(a)pyrene	23.897	252	5005883	42.712	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.562	276	6064511	51.268	ng/ul	99
95) Dibenzo(a,h)anthracene	26.585	278	5503975	51.147	ng/ul	99
96) Benzo(g,h,i)perylene	27.350	276	5261758	52.036	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 16 01:52:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 16 01:29:28 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 17.597 ng/uL
RT: 3.428 min Scan# 58
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion: 88 Resp: 168716
Ion Ratio Lower Upper
88 100
43 28.0 23.1 34.7
58 71.6 56.6 85.0



#5
Pyridine
Concen: 41.082 ng/uL
RT: 3.834 min Scan# 127
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion: 79 Resp: 1166814
Ion Ratio Lower Upper
79 100
52 69.3 57.2 85.8
51 32.8 26.8 40.2









#6
Benzaldehyde
Concen: 47.988 ng/ul
RT: 7.163 min Scan# 693
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion: 77 Resp: 693816
Ion Ratio Lower Upper
77 100
105 98.1 75.1 112.7
106 96.6 73.0 109.6



#8
Phenol
Concen: 42.267 ng/ul
RT: 7.204 min Scan# 700
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion: 94 Resp: 1392925
Ion Ratio Lower Upper
94 100
65 28.0 22.2 33.4
66 36.0 28.9 43.3









#10
Bis(2-Chloroethyl)ether
Concen: 40.843 ng/ul
RT: 7.451 min Scan# 741
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion: 93 Resp: 1083808
Ion Ratio Lower Upper
93 100
63 74.0 59.1 88.7
95 32.5 26.1 39.1



#12
2-Chlorophenol
Concen: 41.473 ng/ul
RT: 7.587 min Scan# 765
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:128 Resp: 1063946
Ion Ratio Lower Upper
128 100
64 46.8 37.8 56.8
130 32.0 26.0 39.0





#13

2-Methylphenol

Concen: 40.243 ng/ul

RT: 8.469 min Scan# 914

Delta R.T. 0.006 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :





SSTD040022

Tgt Ion:108 Resp: 1024328

Ion Ratio Lower Upper

108 100

107 87.0 70.7 106.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 32.440 ng/ul

RT: 8.563 min Scan# 931

Delta R.T. 0.000 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27




Tgt Ion: 45 Resp: 1292631


Ion Ratio Lower Upper

45 100

77 16.8 13.1 19.7

79 12.4 9.5 14.3









#16
Acetophenone
Concen: 40.448 ng/ul
RT: 8.857 min Scan# 981
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:105 Resp: 1676889
Ion Ratio Lower Upper
105 100
77 77.0 61.6 92.4
51 25.2 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 41.361 ng/ul
RT: 8.851 min Scan# 980
Delta R.T. 0.012 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion: 70 Resp: 841316
Ion Ratio Lower Upper
70 100
42 52.0 43.0 64.4
101 9.7 7.7 11.5
130 21.3 15.1 22.7





#18

4-Methylphenol

Concen: 40.471 ng/ul

RT: 8.804 min Scan# 91

Delta R.T. 0.012 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :





SSTD040022

Tgt Ion:108 Resp: 1113831

Ion Ratio Lower Upper

108 100

107 107.5 89.4 134.0



#19

Hexachloroethane

Concen: 38.132 ng/ul

RT: 9.122 min Scan# 1026

Delta R.T. 0.006 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27




Tgt Ion:117 Resp: 417245


Ion Ratio Lower Upper

117 100

201 92.5 68.7 103.1

199 57.1 43.0 64.6









#22
Nitrobenzene
Concen: 39.009 ng/ul
RT: 9.239 min Scan# 1046
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion: 77 Resp: 1178336
Ion Ratio Lower Upper
77 100
123 44.9 34.9 52.3
65 13.4 10.7 16.1



#23
Isophorone
Concen: 42.871 ng/ul
RT: 9.769 min Scan# 1136
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion: 82 Resp: 2310700
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.2
138 17.3 13.5 20.3









#25
2-Nitrophenol
Concen: 37.787 ng/ul
RT: 9.951 min Scan# 1167
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:139 Resp: 511393
Ion Ratio Lower Upper
139 100
65 51.2 40.9 61.3
109 30.6 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 42.975 ng/ul
RT: 10.010 min Scan# 1177
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:107 Resp: 1193356
Ion Ratio Lower Upper
107 100
121 50.8 39.4 59.2
122 88.2 69.1 103.7





#27

Bis(2-Chloroethoxy)methane

Concen: 43.391 ng/ul

RT: 10.251 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :





BNA_M

ClientSampleId :

SSTD040022

Tgt Ion: 93 Resp: 1373216

Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.2	39.4
123	11.5	9.2	13.8



#29

2,4-Dichlorophenol

Concen: 44.126 ng/ul

RT: 10.486 min Scan# 1258




Delta R.T. 0.006 min


Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Tgt Ion: 162 Resp: 969821

Ion	Ratio	Lower	Upper
162	100		
164	64.3	52.1	78.1
98	35.9	29.1	43.7





#30

Naphthalene

Concen: 42.346 ng/ul

RT: 10.898 min Scan# 1327

Delta R.T. 0.006 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD040022





Tgt Ion:128 Resp: 3433796

Ion Ratio Lower Upper

128 100

129 10.8 8.7 13.1

127 13.0 10.3 15.5



#32

4-Chloroaniline

Concen: 40.906 ng/ul

RT: 10.998 min Scan# 1345

Delta R.T. 0.000 min

Lab File: BM036551.D




Acq: 15 Sep 2022 18:27


Tgt Ion:127 Resp: 1470161

Ion Ratio Lower Upper

127 100

129 31.8 25.4 38.2









#33
Hexachlorobutadiene
Concen: 40.393 ng/ul
RT: 11.186 min Scan# 111
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:225 Resp: 568708
Ion Ratio Lower Upper
225 100
223 62.4 49.8 74.8
227 64.2 51.0 76.6



#34
Caprolactam
Concen: 20.463 ng/ul
RT: 11.763 min Scan# 1475
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:113 Resp: 144381
Ion Ratio Lower Upper
113 100
55 162.9 127.5 191.3
56 120.9 94.6 142.0









#35
 4-Chloro-3-methylphenol
 Concen: 45.792 ng/ul
 RT: 12.110 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BM036551.D
 Acq: 15 Sep 2022 18:27




Instrument :
 BNA_M
 ClientSampleId :
 SST040022


Tgt Ion:107 Resp: 1063738
 Ion Ratio Lower Upper
 107 100
 144 25.1 21.5 32.3
 142 79.0 63.9 95.9



#36
 2-Methylnaphthalene
 Concen: 44.269 ng/ul
 RT: 12.498 min Scan# 1600
 Delta R.T. 0.006 min
 Lab File: BM036551.D
 Acq: 15 Sep 2022 18:27

Tgt Ion:142 Resp: 2308769
 Ion Ratio Lower Upper
 142 100
 141 89.3 70.7 106.1





#37

1-Methylnaphthalene

Concen: 44.265 ng/ul

RT: 12.716 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD040022





Tgt Ion:142 Resp: 2316187

Ion Ratio Lower Upper

142 100

141 91.1 73.4 110.2

116 3.8 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 43.186 ng/ul

RT: 12.857 min Scan# 1661

Delta R.T. 0.000 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Tgt Ion:216 Resp: 1135782




Ion Ratio Lower Upper


216 100

214 79.2 62.4 93.6

179 21.2 16.7 25.1

108 19.9 15.4 23.0









#40
Hexachlorocyclopentadiene
Concen: 42.403 ng/ul
RT: 12.845 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:237 Resp: 583691
Ion Ratio Lower Upper
237 100
235 63.2 50.9 76.3
272 13.5 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 41.312 ng/ul
RT: 13.092 min Scan# 1701
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:196 Resp: 712427
Ion Ratio Lower Upper
196 100
198 94.7 77.3 115.9
200 30.5 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 42.200 ng/ul

RT: 13.163 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD040022





Tgt Ion:196 Resp: 768128

Ion Ratio Lower Upper

196 100

198 96.7 75.6 113.4

200 31.3 24.7 37.1



#43

1,1'-Biphenyl

Concen: 42.388 ng/ul

RT: 13.498 min Scan# 1770

Delta R.T. 0.000 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27




Tgt Ion:154 Resp: 2918549


Ion Ratio Lower Upper

154 100

153 40.4 31.4 47.2

76 13.0 10.4 15.6





#44

2-Chloronaphthalene

Concen: 42.967 ng/ul

RT: 13.539 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD040022





Tgt Ion:162 Resp: 2299627

Ion Ratio Lower Upper

162 100

164 32.6 26.2 39.2

127 39.7 32.6 49.0



#45

2-Nitroaniline

Concen: 38.745 ng/ul

RT: 13.733 min Scan# 1810

Delta R.T. 0.000 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27




Tgt Ion: 65 Resp: 626159


Ion Ratio Lower Upper

65 100

92 69.9 54.4 81.6

138 106.7 86.1 129.1









#47
Dimethylphthalate
Concen: 43.432 ng/ul
RT: 14.116 min Scan# 1875
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:163 Resp: 2832293
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.0 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 40.469 ng/ul
RT: 14.227 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:165 Resp: 515472
Ion Ratio Lower Upper
165 100
89 67.3 52.8 79.2
121 23.5 19.4 29.0









#50
Acenaphthylene
Concen: 43.199 ng/ul
RT: 14.380 min Scan# 1919
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:152 Resp: 3632078
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 13.2 10.5 15.7



#51
3-Nitroaniline
Concen: 40.006 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:138 Resp: 571136
Ion Ratio Lower Upper
138 100
108 12.9 12.5 18.7
92 114.3 95.4 143.2









#52
Acenaphthene
Concen: 42.211 ng/ul
RT: 14.722 min Scan# 1978
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:153 Resp: 2529560
Ion Ratio Lower Upper
153 100
152 48.2 38.2 57.2
154 88.6 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 30.284 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:184 Resp: 191798
Ion Ratio Lower Upper
184 100
63 71.2 58.8 88.2
154 64.8 55.0 82.4









#55
4-Nitrophenol
Concen: 41.030 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. 0.012 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:109 Resp: 435149
Ion Ratio Lower Upper
109 100
139 131.9 102.2 153.4
65 129.2 105.7 158.5



#56
Dibenzofuran
Concen: 41.065 ng/ul
RT: 15.051 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:168 Resp: 3351476
Ion Ratio Lower Upper
168 100
139 38.4 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 38.998 ng/ul
RT: 15.010 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:165 Resp: 745342
Ion Ratio Lower Upper
165 100
63 47.0 39.7 59.5
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.390 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:232 Resp: 640634
Ion Ratio Lower Upper
232 100
131 51.5 39.1 58.7
130 2.6 1.8 2.6
166 32.7 25.4 38.0









#59
Diethylphthalate
Concen: 40.069 ng/ul
RT: 15.474 min Scan# 2106
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:149 Resp: 2851381
Ion Ratio Lower Upper
149 100
177 21.6 17.3 25.9
150 12.4 10.1 15.1



#61
Fluorene
Concen: 42.418 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:166 Resp: 2941728
Ion Ratio Lower Upper
166 100
165 97.5 78.5 117.7
167 13.8 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 42.872 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:204 Resp: 1395621
Ion Ratio Lower Upper
204 100
206 33.0 25.7 38.5
141 60.7 49.4 74.0



#63
4-Nitroaniline
Concen: 41.792 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:138 Resp: 583285
Ion Ratio Lower Upper
138 100
92 52.2 42.1 63.1
108 80.5 66.2 99.2









#66
4,6-Dinitro-2-methylphenol
Concen: 34.920 ng/ul
RT: 15.768 min Scan# 2155
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:198 Resp: 361251
Ion Ratio Lower Upper
198 100
182 4.0 4.2 6.2#
77 26.6 24.0 36.0



#67
N-Nitrosodiphenylamine
Concen: 43.926 ng/ul
RT: 15.904 min Scan# 2179
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:169 Resp: 2386023
Ion Ratio Lower Upper
169 100
168 66.9 54.6 81.8
167 36.6 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 43.251 ng/ul
RT: 16.586 min Scan# 2295
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:248 Resp: 798315
Ion Ratio Lower Upper
248 100
250 97.4 78.9 118.3
141 76.2 65.7 98.5



#69
Hexachlorobenzene
Concen: 42.727 ng/ul
RT: 16.692 min Scan# 2313
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:284 Resp: 934309
Ion Ratio Lower Upper
284 100
142 39.8 30.8 46.2
249 31.0 24.2 36.2





#70

Atrazine

Concen: 40.666 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD040022





Tgt Ion:200 Resp: 818221

Ion Ratio Lower Upper

200 100

173 27.4 21.8 32.6

215 47.7 39.4 59.2



#71

Pentachlorophenol

Concen: 41.451 ng/ul

RT: 17.033 min Scan# 2371

Delta R.T. 0.000 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27




Tgt Ion:266 Resp: 544796


Ion Ratio Lower Upper

266 100

264 62.4 49.7 74.5

268 63.7 51.8 77.6









#72
Phenanthrene
Concen: 42.885 ng/ul
RT: 17.433 min Scan# 2439
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:178 Resp: 4459050
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 19.9 15.7 23.5



#74
Anthracene
Concen: 43.259 ng/ul
RT: 17.521 min Scan# 2454
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:178 Resp: 4562382
Ion Ratio Lower Upper
178 100
179 16.0 12.3 18.5
176 19.1 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 45.171 ng/uL

RT: 13.463 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD040022





Tgt Ion:216 Resp: 1113253

Ion Ratio Lower Upper

216 100

214 77.7 62.6 94.0

218 48.5 38.2 57.2



#76

Pentachlorobenzene

Concen: 43.454 ng/uL

RT: 14.969 min Scan# 2020

Delta R.T. 0.000 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27




Tgt Ion:250 Resp: 1181691


Ion Ratio Lower Upper

250 100

248 63.8 50.2 75.2

252 64.2 50.5 75.7









#77
 Carbazole
 Concen: 40.505 ng/ul
 RT: 17.786 min Scan# 2499
 Delta R.T. 0.000 min
 Lab File: BM036551.D
 Acq: 15 Sep 2022 18:27




Instrument :
 BNA_M
 ClientSampleId :
 SST040022


Tgt Ion:167 Resp: 4070173
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.2 25.8
 139 13.7 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 40.563 ng/ul
 RT: 18.362 min Scan# 2597
 Delta R.T. 0.000 min
 Lab File: BM036551.D
 Acq: 15 Sep 2022 18:27

Tgt Ion:149 Resp: 4907089
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.2 10.8
 104 5.2 4.3 6.5









#80
Fluoranthene
Concen: 38.874 ng/ul
RT: 19.439 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:202 Resp: 5161696
Ion Ratio Lower Upper
202 100
101 11.8 9.8 14.6
100 8.9 7.4 11.2



#82
Pyrene
Concen: 38.709 ng/ul
RT: 19.798 min Scan# 2841
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:202 Resp: 5450284
Ion Ratio Lower Upper
202 100
101 14.6 10.9 16.3
100 11.9 8.7 13.1









#83
 Butylbenzylphthalate
 Concen: 35.824 ng/ul
 RT: 20.698 min Scan# 2994
 Delta R.T. 0.000 min
 Lab File: BM036551.D
 Acq: 15 Sep 2022 18:27




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040022


Tgt Ion:149 Resp: 2156834
 Ion Ratio Lower Upper
 149 100
 91 72.9 55.2 82.8
 206 20.4 16.2 24.2



#84
 3,3'-Dichlorobenzidine
 Concen: 42.952 ng/ul
 RT: 21.468 min Scan# 3125
 Delta R.T. 0.000 min
 Lab File: BM036551.D
 Acq: 15 Sep 2022 18:27

Tgt Ion:252 Resp: 1901271
 Ion Ratio Lower Upper
 252 100
 254 64.2 53.0 79.4
 126 12.8 10.4 15.6









#85
Benzo(a)anthracene
Concen: 41.786 ng/ul
RT: 21.539 min Scan# 3137
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:228 Resp: 5538191
Ion Ratio Lower Upper
228 100
229 20.2 15.8 23.8
226 28.2 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.227 ng/ul
RT: 21.474 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:149 Resp: 3441673
Ion Ratio Lower Upper
149 100
167 27.9 22.0 33.0
279 3.9 3.3 4.9









#87
Chrysene
Concen: 42.431 ng/ul
RT: 21.598 min Scan# 3146
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:228 Resp: 5225043
Ion Ratio Lower Upper
228 100
226 30.8 24.4 36.6
229 19.8 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 32.685 ng/ul
RT: 22.421 min Scan# 3287
Delta R.T. 0.000 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:149 Resp: 5527084









#90
Benzo(b)fluoranthene
Concen: 39.367 ng/ul
RT: 23.256 min Scan# 3428
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD040022


Tgt Ion:252 Resp: 5596622
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 11.2 9.1 13.7



#91
Benzo(k)fluoranthene
Concen: 42.280 ng/ul
RT: 23.309 min Scan# 3438
Delta R.T. 0.006 min
Lab File: BM036551.D
Acq: 15 Sep 2022 18:27

Tgt Ion:252 Resp: 5692990
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.8
125 10.9 8.2 12.4





#93

Benzo(a)pyrene

Concen: 42.712 ng/ul

RT: 23.897 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD040022





Tgt Ion:252 Resp: 5005883

Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3

125 12.3 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 51.268 ng/ul

RT: 26.562 min Scan# 3991

Delta R.T. 0.012 min

Lab File: BM036551.D

Acq: 15 Sep 2022 18:27




Tgt Ion:276 Resp: 6064511


Ion Ratio Lower Upper

276 100

138 25.8 20.2 30.4

277 24.8 19.7 29.5









#95
 Dibenzo(a,h)anthracene
 Concen: 51.147 ng/ul
 RT: 26.585 min Scan# 3995
 Delta R.T. 0.012 min
 Lab File: BM036551.D
 Acq: 15 Sep 2022 18:27

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040022

Tgt Ion:278 Resp: 5503975
 Ion Ratio Lower Upper
 278 100
 139 17.6 14.5 21.7
 279 24.1 19.7 29.5



#96
 Benzo(g,h,i)perylene
 Concen: 52.036 ng/ul
 RT: 27.350 min Scan# 4125
 Delta R.T. 0.018 min
 Lab File: BM036551.D
 Acq: 15 Sep 2022 18:27

Tgt Ion:276 Resp: 5261758
 Ion Ratio Lower Upper
 276 100
 138 22.8 18.7 28.1
 277 23.4 19.2 28.8

