

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036554.D
 Acq On : 15 Sep 2022 20:15
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV025

Quant Time: Sep 16 02:46:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	253100	20.000	ng/u1	0.00
20) Naphthalene-d8	10.845	136	1095663	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	693967	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	1459683	20.000	ng/u1	0.00
79) Chrysene-d12	21.550	240	1653116	20.000	ng/u1	0.00
88) Perylene-d12	23.997	264	1722876	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	48117	7.127	ng/uL	0.00
4) Pyridine-d5	3.816	84	324331	16.472	ng/u1	0.00
7) Phenol-d5	7.175	99	357132	15.790	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	262051	19.078	ng/u1	0.00
11) 2-Chlorophenol-d4	7.551	132	293232	18.007	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	291594	17.448	ng/u1	0.00
21) Nitrobenzene-d5	9.192	128	134006	18.070	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	114932	17.302	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	268555	17.445	ng/u1	0.00
31) 4-Chloroaniline-d4	10.974	131	437490	17.093	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	867715	18.539	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1053971	18.984	ng/u1	0.00
54) 4-Nitrophenol-d4	14.827	143	147516	16.250	ng/u1	0.00
60) Fluorene-d10	15.639	176	756090	18.026	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	80398	13.921	ng/u1	0.00
73) Anthracene-d10	17.486	188	1166569	17.479	ng/u1	0.00
81) Pyrene-d10	19.768	212	1382290	16.997	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.838	264	1466012	17.417	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	49157	6.759	ng/uL	97
5) Pyridine	3.834	79	299714	15.150	ng/u1	98
6) Benzaldehyde	7.157	77	265187	24.212	ng/u1	98
8) Phenol	7.198	94	391754	16.138	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.445	93	317358	16.803	ng/u1	98
12) 2-Chlorophenol	7.586	128	304905	17.309	ng/u1	98
13) 2-Methylphenol	8.463	108	292317	16.383	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.563	45	369382	16.501	ng/u1	98
16) Acetophenone	8.851	105	508770	17.551	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.839	70	241315	17.252	ng/u1	97
18) 4-Methylphenol	8.792	108	319744	16.572	ng/u1	99
19) Hexachloroethane	9.122	117	120017	16.821	ng/u1	93
22) Nitrobenzene	9.233	77	337670	17.149	ng/u1	99
23) Isophorone	9.763	82	648000	16.141	ng/u1	98
25) 2-Nitrophenol	9.951	139	139165	16.962	ng/u1	99
26) 2,4-Dimethylphenol	10.004	107	344078	17.003	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.251	93	415948	17.383	ng/u1	99
29) 2,4-Dichlorophenol	10.480	162	281268	17.379	ng/u1	99
30) Naphthalene	10.892	128	1017002	17.039	ng/u1	100
32) 4-Chloroaniline	10.998	127	445920	16.875	ng/u1	99
33) Hexachlorobutadiene	11.180	225	168674	17.041	ng/u1	98
34) Caprolactam	11.751	113	84717	16.137	ng/u1	99
35) 4-Chloro-3-methylphenol	12.104	107	305423	17.096	ng/u1	98
36) 2-Methylnaphthalene	12.492	142	692908	17.424	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036554.D
 Acq On : 15 Sep 2022 20:15
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV025

Quant Time: Sep 16 02:46:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.710	142	667783	16.703	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.857	216	339903	17.305	ng/ul	98
40) Hexachlorocyclopentadiene	12.839	237	163933	16.015	ng/ul	99
41) 2,4,6-Trichlorophenol	13.086	196	202594	16.845	ng/ul	97
42) 2,4,5-Trichlorophenol	13.157	196	226206	17.543	ng/ul	98
43) 1,1'-Biphenyl	13.492	154	919984	17.984	ng/ul	100
44) 2-Chloronaphthalene	13.533	162	693073	17.265	ng/ul	99
45) 2-Nitroaniline	13.733	65	170062	17.212	ng/ul	99
47) Dimethylphthalate	14.110	163	858446	17.291	ng/ul	99
48) 2,6-Dinitrotoluene	14.221	165	163451	20.249	ng/ul	96
50) Acenaphthylene	14.374	152	1124089	17.710	ng/ul	100
51) 3-Nitroaniline	14.551	138	185447	19.156	ng/ul	96
52) Acenaphthene	14.715	153	767005	17.270	ng/ul	100
53) 2,4-Dinitrophenol	14.751	184	46440	12.966	ng/ul	95
55) 4-Nitrophenol	14.839	109	127926	16.485	ng/ul	97
56) Dibenzofuran	15.051	168	1076432	18.048	ng/ul	99
57) 2,4-Dinitrotoluene	15.004	165	208137	17.737	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.268	232	192557	17.840	ng/ul	99
59) Diethylphthalate	15.468	149	839740	16.896	ng/ul	99
61) Fluorene	15.692	166	895938	17.637	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.692	204	416893	17.160	ng/ul	99
63) 4-Nitroaniline	15.704	138	198226	21.075	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.762	198	88957	13.650	ng/ul	97
67) N-Nitrosodiphenylamine	15.898	169	738615	17.180	ng/ul	100
68) 4-Bromophenyl-phenylether	16.580	248	239580	16.652	ng/ul	98
69) Hexachlorobenzene	16.692	284	282065	16.576	ng/ul	99
70) Atrazine	16.845	200	246323	16.237	ng/ul	97
71) Pentachlorophenol	17.033	266	145321	14.377	ng/ul	98
72) Phenanthrene	17.427	178	1387207	17.254	ng/ul	100
74) Anthracene	17.521	178	1409035	17.242	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.457	216	330332	16.546	ng/ul	99
76) Pentachlorobenzene	14.968	250	335239	15.525	ng/ul	99
77) Carbazole	17.786	167	1340886	17.722	ng/ul	100
78) Di-n-butylphthalate	18.356	149	1376308	16.049	ng/ul	100
80) Fluoranthene	19.433	202	1587350	15.928	ng/ul	100
82) Pyrene	19.797	202	1725574	16.358	ng/ul	99
83) Butylbenzylphthalate	20.697	149	617886	15.458	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.468	252	662047	20.496	ng/ul	98
85) Benzo(a)anthracene	21.538	228	1778443	17.170	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.474	149	1069918	18.323	ng/ul	99
87) Chrysene	21.591	228	1709265	17.386	ng/ul	99
89) Di-n-octyl phthalate	22.421	149	1602195	15.425	ng/ul	100
90) Benzo(b)fluoranthene	23.250	252	1858733	17.013	ng/ul	99
91) Benzo(k)fluoranthene	23.297	252	1850122	16.947	ng/ul	99
93) Benzo(a)pyrene	23.891	252	1642080	17.097	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.544	276	2022580	18.033	ng/ul	99
95) Dibenzo(a,h)anthracene	26.568	278	1798085	17.677	ng/ul	98
96) Benzo(g,h,i)perylene	27.326	276	1727800	17.770	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091522\
 Data File : BM036554.D
 Acq On : 15 Sep 2022 20:15
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV025

Quant Time: Sep 16 02:46:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.759 ng/uL
RT: 3.428 min Scan# 58
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion: 88 Resp: 49157
Ion Ratio Lower Upper
88 100
43 26.1 23.1 34.7
58 69.2 56.6 85.0



#5
Pyridine
Concen: 15.150 ng/uL
RT: 3.834 min Scan# 127
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion: 79 Resp: 299714
Ion Ratio Lower Upper
79 100
52 69.6 57.2 85.8
51 33.6 26.8 40.2









#6
Benzaldehyde
Concen: 24.212 ng/ul
RT: 7.157 min Scan# 692
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion: 77 Resp: 265187
Ion Ratio Lower Upper
77 100
105 96.8 75.1 112.7
106 92.3 73.0 109.6



#8
Phenol
Concen: 16.138 ng/ul
RT: 7.198 min Scan# 699
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion: 94 Resp: 391754
Ion Ratio Lower Upper
94 100
65 28.0 22.2 33.4
66 36.1 28.9 43.3









#10
Bis(2-Chloroethyl)ether
Concen: 16.803 ng/ul
RT: 7.445 min Scan# 741
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion: 93 Resp: 317358
Ion Ratio Lower Upper
93 100
63 72.2 59.1 88.7
95 32.1 26.1 39.1



#12
2-Chlorophenol
Concen: 17.309 ng/ul
RT: 7.586 min Scan# 765
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion: 128 Resp: 304905
Ion Ratio Lower Upper
128 100
64 45.5 37.8 56.8
130 32.6 26.0 39.0





#13

2-Methylphenol

Concen: 16.383 ng/ul

RT: 8.463 min Scan# 914

Delta R.T. -0.000 min

Lab File: BM036554.D

Acq: 15 Sep 2022 20:15

Instrument :

BNA_M

ClientSampleId :





SICV025

Tgt Ion:108 Resp: 292317

Ion Ratio Lower Upper

108 100

107 87.8 70.7 106.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 16.501 ng/ul

RT: 8.563 min Scan# 931

Delta R.T. -0.000 min

Lab File: BM036554.D

Acq: 15 Sep 2022 20:15




Tgt Ion: 45 Resp: 369382


Ion Ratio Lower Upper

45 100

77 15.9 13.1 19.7

79 13.2 9.5 14.3









#16
Acetophenone
Concen: 17.551 ng/ul
RT: 8.851 min Scan# 981
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:105 Resp: 508770
Ion Ratio Lower Upper
105 100
77 77.2 61.6 92.4
51 25.3 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 17.252 ng/ul
RT: 8.839 min Scan# 978
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion: 70 Resp: 241315
Ion Ratio Lower Upper
70 100
42 51.3 43.0 64.4
101 9.5 7.7 11.5
130 20.5 15.1 22.7









#18
4-Methylphenol
Concen: 16.572 ng/ul
RT: 8.792 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:108 Resp: 319744
Ion Ratio Lower Upper
108 100
107 110.5 89.4 134.0



#19
Hexachloroethane
Concen: 16.821 ng/ul
RT: 9.122 min Scan# 1026
Delta R.T. 0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:117 Resp: 120017
Ion Ratio Lower Upper
117 100
201 94.3 68.7 103.1
199 56.6 43.0 64.6









#22
Nitrobenzene
Concen: 17.149 ng/ul
RT: 9.233 min Scan# 1045
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion: 77 Resp: 337670
Ion Ratio Lower Upper
77 100
123 43.1 34.9 52.3
65 12.5 10.7 16.1



#23
Isophorone
Concen: 16.141 ng/ul
RT: 9.763 min Scan# 1135
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion: 82 Resp: 648000
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.2
138 18.0 13.5 20.3









#25
2-Nitrophenol
Concen: 16.962 ng/ul
RT: 9.951 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:139 Resp: 139165
Ion Ratio Lower Upper
139 100
65 51.7 40.9 61.3
109 29.6 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 17.003 ng/ul
RT: 10.004 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:107 Resp: 344078
Ion Ratio Lower Upper
107 100
121 49.8 39.4 59.2
122 87.3 69.1 103.7





#27

Bis(2-Chloroethoxy)methane

Concen: 17.383 ng/ul

RT: 10.251 min Scan# 1218

Delta R.T. -0.000 min

Lab File: BM036554.D

Acq: 15 Sep 2022 20:15

Instrument :

BNA_M

ClientSampleId :

SICV025





Tgt Ion: 93 Resp: 415948

Ion Ratio Lower Upper

93 100

95 32.4 26.2 39.4

123 12.0 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 17.379 ng/ul

RT: 10.480 min Scan# 1257

Delta R.T. -0.000 min

Lab File: BM036554.D

Acq: 15 Sep 2022 20:15




Tgt Ion: 162 Resp: 281268


Ion Ratio Lower Upper

162 100

164 66.2 52.1 78.1

98 35.7 29.1 43.7









#30
Naphthalene
Concen: 17.039 ng/ul
RT: 10.892 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:128 Resp: 1017002
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.8 10.3 15.5



#32
4-Chloroaniline
Concen: 16.875 ng/ul
RT: 10.998 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:127 Resp: 445920
Ion Ratio Lower Upper
127 100
129 32.4 25.4 38.2









#33
Hexachlorobutadiene
Concen: 17.041 ng/ul
RT: 11.180 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:225 Resp: 168674
Ion Ratio Lower Upper
225 100
223 61.9 49.8 74.8
227 67.0 51.0 76.6



#34
Caprolactam
Concen: 16.137 ng/ul
RT: 11.751 min Scan# 1473
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:113 Resp: 84717
Ion Ratio Lower Upper
113 100
55 159.2 127.5 191.3
56 121.2 94.6 142.0









#35
4-Chloro-3-methylphenol
Concen: 17.096 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:107 Resp: 305423
Ion Ratio Lower Upper
107 100
144 25.5 21.5 32.3
142 81.0 63.9 95.9



#36
2-Methylnaphthalene
Concen: 17.424 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:142 Resp: 692908
Ion Ratio Lower Upper
142 100
141 87.9 70.7 106.1









#37
1-Methylnaphthalene
Concen: 16.703 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:142 Resp: 667783
Ion Ratio Lower Upper
142 100
141 92.9 73.4 110.2
116 3.9 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.305 ng/ul
RT: 12.857 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:216 Resp: 339903
Ion Ratio Lower Upper
216 100
214 79.7 62.4 93.6
179 20.4 16.7 25.1
108 18.8 15.4 23.0









#40
Hexachlorocyclopentadiene
Concen: 16.015 ng/ul
RT: 12.839 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:237 Resp: 163933
Ion Ratio Lower Upper
237 100
235 64.1 50.9 76.3
272 12.8 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 16.845 ng/ul
RT: 13.086 min Scan# 1700
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:196 Resp: 202594
Ion Ratio Lower Upper
196 100
198 93.0 77.3 115.9
200 30.8 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 17.543 ng/ul
RT: 13.157 min Scan# 1712
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:196 Resp: 226206
Ion Ratio Lower Upper
196 100
198 92.4 75.6 113.4
200 31.2 24.7 37.1



#43
1,1'-Biphenyl
Concen: 17.984 ng/ul
RT: 13.492 min Scan# 1769
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:154 Resp: 919984
Ion Ratio Lower Upper
154 100
153 39.6 31.4 47.2
76 12.9 10.4 15.6









#44
2-Chloronaphthalene
Concen: 17.265 ng/ul
RT: 13.533 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:162 Resp: 693073
Ion Ratio Lower Upper
162 100
164 32.4 26.2 39.2
127 40.4 32.6 49.0



#45
2-Nitroaniline
Concen: 17.212 ng/ul
RT: 13.733 min Scan# 1810
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion: 65 Resp: 170062
Ion Ratio Lower Upper
65 100
92 69.0 54.4 81.6
138 107.4 86.1 129.1









#47
Dimethylphthalate
Concen: 17.291 ng/ul
RT: 14.110 min Scan# 1874
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:163 Resp: 858446
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.6
164 9.9 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.249 ng/ul
RT: 14.221 min Scan# 1893
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:165 Resp: 163451
Ion Ratio Lower Upper
165 100
89 68.8 52.8 79.2
121 22.4 19.4 29.0









#50
Acenaphthylene
Concen: 17.710 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:152 Resp: 1124089
Ion Ratio Lower Upper
152 100
151 19.6 15.7 23.5
153 13.3 10.5 15.7



#51
3-Nitroaniline
Concen: 19.156 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:138 Resp: 185447
Ion Ratio Lower Upper
138 100
108 12.6 12.5 18.7
92 115.6 95.4 143.2









#52
Acenaphthene
Concen: 17.270 ng/ul
RT: 14.715 min Scan# 1977
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:153 Resp: 767005
Ion Ratio Lower Upper
153 100
152 48.1 38.2 57.2
154 88.9 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 12.966 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:184 Resp: 46440
Ion Ratio Lower Upper
184 100
63 74.8 58.8 88.2
154 76.0 55.0 82.4









#55
4-Nitrophenol
Concen: 16.485 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:109 Resp: 127926
Ion Ratio Lower Upper
109 100
139 131.8 102.2 153.4
65 130.1 105.7 158.5



#56
Dibenzofuran
Concen: 18.048 ng/ul
RT: 15.051 min Scan# 2034
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:168 Resp: 1076432
Ion Ratio Lower Upper
168 100
139 37.9 29.7 44.5





#57

2,4-Dinitrotoluene

Concen: 17.737 ng/ul

RT: 15.004 min Scan# 2026

Delta R.T. -0.000 min

Lab File: BM036554.D

Acq: 15 Sep 2022 20:15

Instrument :

BNA_M

ClientSampleId :

SICV025





Tgt Ion:165 Resp: 208137

Ion Ratio Lower Upper

165 100

63 45.5 39.7 59.5

182 2.7 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 17.840 ng/ul

RT: 15.268 min Scan# 2071

Delta R.T. -0.000 min

Lab File: BM036554.D

Acq: 15 Sep 2022 20:15

Tgt Ion:232 Resp: 192557




Ion Ratio Lower Upper


232 100

131 49.0 39.1 58.7

130 2.6 1.8 2.6

166 33.1 25.4 38.0









#59
Diethylphthalate
Concen: 16.896 ng/ul
RT: 15.468 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:149 Resp: 839740
Ion Ratio Lower Upper
149 100
177 21.1 17.3 25.9
150 12.6 10.1 15.1



#61
Fluorene
Concen: 17.637 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:166 Resp: 895938
Ion Ratio Lower Upper
166 100
165 96.6 78.5 117.7
167 13.7 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 17.160 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:204 Resp: 416893
Ion Ratio Lower Upper
204 100
206 32.8 25.7 38.5
141 61.1 49.4 74.0



#63
4-Nitroaniline
Concen: 21.075 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:138 Resp: 198226
Ion Ratio Lower Upper
138 100
92 52.0 42.1 63.1
108 79.9 66.2 99.2









#66
4,6-Dinitro-2-methylphenol
Concen: 13.650 ng/ul
RT: 15.762 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:198 Resp: 88957
Ion Ratio Lower Upper
198 100
182 5.6 4.2 6.2
77 28.5 24.0 36.0



#67
N-Nitrosodiphenylamine
Concen: 17.180 ng/ul
RT: 15.898 min Scan# 2178
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:169 Resp: 738615
Ion Ratio Lower Upper
169 100
168 68.0 54.6 81.8
167 36.9 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 16.652 ng/ul
RT: 16.580 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:248 Resp: 239580
Ion Ratio Lower Upper
248 100
250 98.1 78.9 118.3
141 79.5 65.7 98.5



#69
Hexachlorobenzene
Concen: 16.576 ng/ul
RT: 16.692 min Scan# 2313
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:284 Resp: 282065
Ion Ratio Lower Upper
284 100
142 37.4 30.8 46.2
249 30.4 24.2 36.2





#70

Atrazine

Concen: 16.237 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM036554.D

Acq: 15 Sep 2022 20:15

Instrument :

BNA_M

ClientSampleId :

SICV025





Tgt Ion:200 Resp: 246323

Ion Ratio Lower Upper

200 100

173 26.2 21.8 32.6

215 46.8 39.4 59.2



#71

Pentachlorophenol

Concen: 14.377 ng/ul

RT: 17.033 min Scan# 2371

Delta R.T. -0.000 min

Lab File: BM036554.D

Acq: 15 Sep 2022 20:15




Tgt Ion:266 Resp: 145321


Ion Ratio Lower Upper

266 100

264 64.2 49.7 74.5

268 65.3 51.8 77.6









#72
Phenanthrene
Concen: 17.254 ng/ul
RT: 17.427 min Scan# 2438
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:178 Resp: 1387207
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 19.7 15.7 23.5



#74
Anthracene
Concen: 17.242 ng/ul
RT: 17.521 min Scan# 2454
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:178 Resp: 1409035
Ion Ratio Lower Upper
178 100
179 15.7 12.3 18.5
176 18.6 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 16.546 ng/uL
RT: 13.457 min Scan# 1763
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:216 Resp: 330332
Ion Ratio Lower Upper
216 100
214 77.7 62.6 94.0
218 48.0 38.2 57.2



#76
Pentachlorobenzene
Concen: 15.525 ng/uL
RT: 14.968 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:250 Resp: 335239
Ion Ratio Lower Upper
250 100
248 62.8 50.2 75.2
252 64.9 50.5 75.7









#77
 Carbazole
 Concen: 17.722 ng/ul
 RT: 17.786 min Scan# 2499
 Delta R.T. -0.000 min
 Lab File: BM036554.D
 Acq: 15 Sep 2022 20:15




Instrument :
 BNA_M
 ClientSampleId :
 SICV025


Tgt Ion:167 Resp: 1340886
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.2 25.8
 139 13.1 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 16.049 ng/ul
 RT: 18.356 min Scan# 2596
 Delta R.T. -0.006 min
 Lab File: BM036554.D
 Acq: 15 Sep 2022 20:15

Tgt Ion:149 Resp: 1376308
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.2 10.8
 104 5.5 4.3 6.5









#80
Fluoranthene
Concen: 15.928 ng/ul
RT: 19.433 min Scan# 2779
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:202 Resp: 1587350
Ion Ratio Lower Upper
202 100
101 12.4 9.8 14.6
100 9.2 7.4 11.2



#82
Pyrene
Concen: 16.358 ng/ul
RT: 19.797 min Scan# 2841
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:202 Resp: 1725574
Ion Ratio Lower Upper
202 100
101 13.3 10.9 16.3
100 10.7 8.7 13.1









#83
Butylbenzylphthalate
Concen: 15.458 ng/ul
RT: 20.697 min Scan# 2994
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:149 Resp: 617886
Ion Ratio Lower Upper
149 100
91 71.8 55.2 82.8
206 20.9 16.2 24.2



#84
3,3'-Dichlorobenzidine
Concen: 20.496 ng/ul
RT: 21.468 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:252 Resp: 662047
Ion Ratio Lower Upper
252 100
254 64.9 53.0 79.4
126 11.9 10.4 15.6









#85
Benzo(a)anthracene
Concen: 17.170 ng/ul
RT: 21.538 min Scan# 3137
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:228 Resp: 1778443
Ion Ratio Lower Upper
228 100
229 20.0 15.8 23.8
226 27.7 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.323 ng/ul
RT: 21.474 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:149 Resp: 1069918
Ion Ratio Lower Upper
149 100
167 27.8 22.0 33.0
279 4.4 3.3 4.9









#87
Chrysene
Concen: 17.386 ng/ul
RT: 21.591 min Scan# 3146
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:228 Resp: 1709265
Ion Ratio Lower Upper
228 100
226 29.9 24.4 36.6
229 19.5 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 15.425 ng/ul
RT: 22.421 min Scan# 3287
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:149 Resp: 1602195









#90
Benzo(b)fluoranthene
Concen: 17.013 ng/ul
RT: 23.250 min Scan# 3428
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:252 Resp: 1858733
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 10.8 9.1 13.7



#91
Benzo(k)fluoranthene
Concen: 16.947 ng/ul
RT: 23.297 min Scan# 3436
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:252 Resp: 1850122
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.8
125 11.2 8.2 12.4









#93
Benzo(a)pyrene
Concen: 17.097 ng/ul
RT: 23.891 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15




Instrument :
BNA_M
ClientSampleId :
SICV025


Tgt Ion:252 Resp: 1642080
Ion Ratio Lower Upper
252 100
253 21.4 17.5 26.3
125 11.3 9.5 14.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.033 ng/ul
RT: 26.544 min Scan# 3988
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:276 Resp: 2022580
Ion Ratio Lower Upper
276 100
138 24.5 20.2 30.4
277 24.3 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 17.677 ng/ul
RT: 26.568 min Scan# 3992
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Instrument :
BNA_M
ClientSampleId :
SICV025

Tgt Ion:278 Resp: 1798085
Ion Ratio Lower Upper
278 100
139 17.3 14.5 21.7
279 23.6 19.7 29.5



#96
Benzo(g,h,i)perylene
Concen: 17.770 ng/ul
RT: 27.326 min Scan# 4121
Delta R.T. -0.006 min
Lab File: BM036554.D
Acq: 15 Sep 2022 20:15

Tgt Ion:276 Resp: 1727800
Ion Ratio Lower Upper
276 100
138 23.8 18.7 28.1
277 23.3 19.2 28.8

