


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091620\  
 Data File : BM027437.D  
 Acq On : 16 Sep 2020 11:04  
 Operator : JU/CG  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\8270-BM090220.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Sep 16 13:41:16 2020



AutoFind: Scans 2452, 2453, 2454; Background Corrected with Scan 2446

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 31.7      | 13778   | PASS             |
| 68          | 69           | 0.00         | 2            | 2.0       | 425     | PASS             |
| 69          | 198          | 0.00         | 100          | 50.0      | 21723   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 46.1      | 20037   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.5       | 203     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 43474   | PASS             |
| 199         | 198          | 5            | 9            | 7.2       | 3126    | PASS             |
| 275         | 198          | 10           | 60           | 31.6      | 13739   | PASS             |
| 365         | 198          | 1            | 100          | 6.0       | 2613    | PASS             |
| 441         | 198          | 0.01         | 100          | 13.1      | 5716    | PASS             |
| 442         | 198          | 50           | 100          | 85.5      | 37192   | PASS             |
| 443         | 442          | 15           | 24           | 18.6      | 6912    | PASS             |