

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
 Data File : BM036573.D
 Acq On : 19 Sep 2022 10:19
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Sep 19 22:28:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	344330	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	1514563	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	958235	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	1921183	20.000	ng/u1	0.00
79) Chrysene-d12	21.550	240	1871595	20.000	ng/u1	0.00
88) Perylene-d12	23.991	264	1797073	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	69865	7.606	ng/uL	0.00
4) Pyridine-d5	3.810	84	508631	18.987	ng/u1	0.00
7) Phenol-d5	7.175	99	569868	18.520	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	360415	19.287	ng/u1	0.00
11) 2-Chlorophenol-d4	7.551	132	441482	19.928	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	442996	19.484	ng/u1	0.00
21) Nitrobenzene-d5	9.186	128	209816	20.467	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	192297	20.942	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	418741	19.678	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	670564	18.953	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	1248765	19.322	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1471607	19.197	ng/u1	0.00
54) 4-Nitrophenol-d4	14.827	143	226172	18.043	ng/u1	0.00
60) Fluorene-d10	15.633	176	1124149	19.410	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	146666	19.296	ng/u1	0.00
73) Anthracene-d10	17.486	188	1689092	19.229	ng/u1	0.00
81) Pyrene-d10	19.762	212	1847412	20.064	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	1732031	19.728	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	75349	7.615	ng/uL	98
5) Pyridine	3.834	79	504131	18.732	ng/u1	96
6) Benzaldehyde	7.157	77	312744	20.989	ng/u1	94
8) Phenol	7.198	94	623955	18.893	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.445	93	499040	19.421	ng/u1	100
12) 2-Chlorophenol	7.581	128	478702	19.976	ng/u1	99
13) 2-Methylphenol	8.463	108	461648	19.018	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.563	45	597538	19.620	ng/u1	98
16) Acetophenone	8.851	105	776332	19.685	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.839	70	387767	20.377	ng/u1	97
18) 4-Methylphenol	8.792	108	506851	19.309	ng/u1	97
19) Hexachloroethane	9.116	117	189950	19.569	ng/u1	91
22) Nitrobenzene	9.233	77	553071	20.320	ng/u1	99
23) Isophorone	9.757	82	1066166	19.212	ng/u1	99
25) 2-Nitrophenol	9.945	139	234839	20.706	ng/u1	98
26) 2,4-Dimethylphenol	10.004	107	555241	19.849	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.245	93	655132	19.806	ng/u1	99
29) 2,4-Dichlorophenol	10.480	162	444271	19.859	ng/u1	99
30) Naphthalene	10.886	128	1617736	19.607	ng/u1	99
32) 4-Chloroaniline	10.992	127	689986	18.889	ng/u1	100
33) Hexachlorobutadiene	11.180	225	263233	19.239	ng/u1	98
34) Caprolactam	11.751	113	126963	17.496	ng/u1	93
35) 4-Chloro-3-methylphenol	12.104	107	487781	19.752	ng/u1	99
36) 2-Methylnaphthalene	12.492	142	1078264	19.615	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
 Data File : BM036573.D
 Acq On : 19 Sep 2022 10:19
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Sep 19 22:28:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.704	142	1082062	19.580	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.851	216	531894	19.611	ng/ul	99
40) Hexachlorocyclopentadiene	12.839	237	274926	19.451	ng/ul	96
41) 2,4,6-Trichlorophenol	13.086	196	326958	19.689	ng/ul	99
42) 2,4,5-Trichlorophenol	13.157	196	357309	20.069	ng/ul	100
43) 1,1'-Biphenyl	13.492	154	1410788	19.972	ng/ul	100
44) 2-Chloronaphthalene	13.533	162	1089942	19.664	ng/ul	98
45) 2-Nitroaniline	13.727	65	286963	21.034	ng/ul	96
47) Dimethylphthalate	14.110	163	1326838	19.354	ng/ul	99
48) 2,6-Dinitrotoluene	14.221	165	233932	20.988	ng/ul	98
50) Acenaphthylene	14.374	152	1710424	19.515	ng/ul	99
51) 3-Nitroaniline	14.545	138	256446	19.184	ng/ul	99
52) Acenaphthene	14.710	153	1195773	19.499	ng/ul	99
53) 2,4-Dinitrophenol	14.745	184	94773	19.163	ng/ul	97
55) 4-Nitrophenol	14.839	109	200016	18.666	ng/ul	98
56) Dibenzofuran	15.045	168	1616007	19.622	ng/ul	96
57) 2,4-Dinitrotoluene	15.004	165	337662	20.839	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.262	232	288842	19.380	ng/ul#	95
59) Diethylphthalate	15.468	149	1315332	19.166	ng/ul	99
61) Fluorene	15.692	166	1367279	19.493	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.686	204	649470	19.361	ng/ul	98
63) 4-Nitroaniline	15.704	138	242352	18.660	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.757	198	165869	19.338	ng/ul	99
67) N-Nitrosodiphenylamine	15.898	169	1123590	19.857	ng/ul	99
68) 4-Bromophenyl-phenylether	16.580	248	331872	17.526	ng/ul	99
69) Hexachlorobenzene	16.686	284	420005	18.753	ng/ul	95
70) Atrazine	16.845	200	358418	17.950	ng/ul	99
71) Pentachlorophenol	17.033	266	229489	17.250	ng/ul	99
72) Phenanthrene	17.427	178	2081736	19.672	ng/ul	100
74) Anthracene	17.521	178	2105578	19.576	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.457	216	518491	19.732	ng/ul	99
76) Pentachlorobenzene	14.962	250	558741	19.660	ng/ul	99
77) Carbazole	17.786	167	1872692	18.805	ng/ul	100
78) Di-n-butylphthalate	18.356	149	2081262	18.439	ng/ul	100
80) Fluoranthene	19.433	202	2269750	20.117	ng/ul	99
82) Pyrene	19.792	202	2427099	20.322	ng/ul	98
83) Butylbenzylphthalate	20.692	149	875727	19.350	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.462	252	705087	19.281	ng/ul	97
85) Benzo(a)anthracene	21.533	228	2328921	19.860	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.468	149	1333490	20.171	ng/ul	100
87) Chrysene	21.586	228	2191047	19.685	ng/ul	99
89) Di-n-octyl phthalate	22.415	149	2134797	19.704	ng/ul	100
90) Benzo(b)fluoranthene	23.244	252	2313478	20.301	ng/ul	100
91) Benzo(k)fluoranthene	23.297	252	2201864	19.337	ng/ul	98
93) Benzo(a)pyrene	23.885	252	1968510	19.649	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.538	276	2112942	18.061	ng/ul	99
95) Dibenzo(a,h)anthracene	26.562	278	1938920	18.275	ng/ul	98
96) Benzo(g,h,i)perylene	27.320	276	1783369	17.584	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
 Data File : BM036573.D
 Acq On : 19 Sep 2022 10:19
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Sep 19 22:28:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.615 ng/uL
RT: 3.428 min Scan# 58
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
88	100		
43	26.7	23.1	34.7
58	69.8	56.6	85.0



#5
Pyridine
Concen: 18.732 ng/uL
RT: 3.834 min Scan# 127
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion	Ratio	Lower	Upper
79	100		
52	67.8	57.2	85.8
51	32.0	26.8	40.2









#6
Benzaldehyde
Concen: 20.989 ng/ul
RT: 7.157 min Scan# 692
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 312744
Ion Ratio Lower Upper
77 100
105 99.3 75.1 112.7
106 96.5 73.0 109.6



#8
Phenol
Concen: 18.893 ng/ul
RT: 7.198 min Scan# 699
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion: 94 Resp: 623955
Ion Ratio Lower Upper
94 100
65 26.9 22.2 33.4
66 36.3 28.9 43.3









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.421 ng/ul
 RT: 7.445 min Scan# 741
 Delta R.T. -0.000 min
 Lab File: BM036573.D
 Acq: 19 Sep 2022 10:19

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion: 93 Resp: 499040




Ion	Ratio	Lower	Upper
93	100		
63	74.1	59.1	88.7
95	33.0	26.1	39.1




#12
 2-Chlorophenol
 Concen: 19.976 ng/ul
 RT: 7.581 min Scan# 764
 Delta R.T. -0.006 min
 Lab File: BM036573.D
 Acq: 19 Sep 2022 10:19

Tgt Ion: 128 Resp: 478702

Ion	Ratio	Lower	Upper
128	100		
64	46.7	37.8	56.8
130	32.7	26.0	39.0





#13

2-Methylphenol

Concen: 19.018 ng/ul

RT: 8.463 min Scan# 914

Delta R.T. -0.000 min


Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:108 Resp: 461648

Ion Ratio Lower Upper

108 100

107 89.2 70.7 106.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.620 ng/ul

RT: 8.563 min Scan# 931

Delta R.T. -0.000 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19




Tgt Ion: 45 Resp: 597538


Ion Ratio Lower Upper

45 100

77 16.1 13.1 19.7

79 13.0 9.5 14.3









#16
Acetophenone
Concen: 19.685 ng/ul
RT: 8.851 min Scan# 980
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:105 Resp: 776332
Ion Ratio Lower Upper
105 100
77 77.0 61.6 92.4
51 25.9 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 20.377 ng/ul
RT: 8.839 min Scan# 978
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion: 70 Resp: 387767
Ion Ratio Lower Upper
70 100
42 51.9 43.0 64.4
101 9.8 7.7 11.5
130 20.4 15.1 22.7





#18

4-Methylphenol

Concen: 19.309 ng/ul

RT: 8.792 min Scan# 970

Delta R.T. -0.000 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Instrument :

BNA_M





ClientSampleId :

Tgt Ion:108 Resp: 506851

Ion Ratio Lower Upper

108 100

107 108.9 89.4 134.0



#19

Hexachloroethane

Concen: 19.569 ng/ul

RT: 9.116 min Scan# 1025

Delta R.T. -0.000 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19




Tgt Ion:117 Resp: 189950


Ion Ratio Lower Upper

117 100

201 95.8 68.7 103.1

199 58.0 43.0 64.6









#22
Nitrobenzene
Concen: 20.320 ng/ul
RT: 9.233 min Scan# 1045
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 553071
Ion Ratio Lower Upper
77 100
123 44.6 34.9 52.3
65 13.2 10.7 16.1



#23
Isophorone
Concen: 19.212 ng/ul
RT: 9.757 min Scan# 1134
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion: 82 Resp: 1066166
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 17.2 13.5 20.3









#25
2-Nitrophenol
Concen: 20.706 ng/ul
RT: 9.945 min Scan# 1166
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:139 Resp: 234839
Ion Ratio Lower Upper
139 100
65 52.6 40.9 61.3
109 30.7 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 19.849 ng/ul
RT: 10.004 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:107 Resp: 555241
Ion Ratio Lower Upper
107 100
121 50.0 39.4 59.2
122 86.8 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.806 ng/ul
 RT: 10.245 min Scan# 1217
 Delta R.T. -0.006 min
 Lab File: BM036573.D
 Acq: 19 Sep 2022 10:19

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion: 93 Resp: 655132




Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.2	39.4
123	11.1	9.2	13.8




#29
 2,4-Dichlorophenol
 Concen: 19.859 ng/ul
 RT: 10.480 min Scan# 1257
 Delta R.T. -0.000 min
 Lab File: BM036573.D
 Acq: 19 Sep 2022 10:19

Tgt Ion: 162 Resp: 444271

Ion	Ratio	Lower	Upper
162	100		
164	64.5	52.1	78.1
98	36.9	29.1	43.7









#30
Naphthalene
Concen: 19.607 ng/ul
RT: 10.886 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 1617736
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.4 10.3 15.5



#32
4-Chloroaniline
Concen: 18.889 ng/ul
RT: 10.992 min Scan# 1344
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:127 Resp: 689986
Ion Ratio Lower Upper
127 100
129 31.7 25.4 38.2





#33

Hexachlorobutadiene

Concen: 19.239 ng/ul

RT: 11.180 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:225 Resp: 263233

Ion Ratio Lower Upper

225 100

223 62.5 49.8 74.8

227 66.1 51.0 76.6



#34

Caprolactam

Concen: 17.496 ng/ul

RT: 11.751 min Scan# 1473

Delta R.T. -0.006 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19




Tgt Ion:113 Resp: 126963


Ion Ratio Lower Upper

113 100

55 168.8 127.5 191.3

56 124.9 94.6 142.0









#35
4-Chloro-3-methylphenol
Concen: 19.752 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:107 Resp: 487781
Ion Ratio Lower Upper
107 100
144 24.8 21.5 32.3
142 79.9 63.9 95.9



#36
2-Methylnaphthalene
Concen: 19.615 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:142 Resp: 1078264
Ion Ratio Lower Upper
142 100
141 87.0 70.7 106.1









#37
1-Methylnaphthalene
Concen: 19.580 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:142 Resp: 1082062
Ion Ratio Lower Upper
142 100
141 92.0 73.4 110.2
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.611 ng/ul
RT: 12.851 min Scan# 1660
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:216 Resp: 531894
Ion Ratio Lower Upper
216 100
214 77.9 62.4 93.6
179 21.2 16.7 25.1
108 20.3 15.4 23.0









#40
Hexachlorocyclopentadiene
Concen: 19.451 ng/ul
RT: 12.839 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:237 Resp: 274926
Ion Ratio Lower Upper
237 100
235 60.3 50.9 76.3
272 13.5 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 19.689 ng/ul
RT: 13.086 min Scan# 1700
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19


Tgt Ion:196 Resp: 326958
Ion Ratio Lower Upper
196 100
198 95.6 77.3 115.9
200 30.7 25.2 37.8








#42
2,4,5-Trichlorophenol
Concen: 20.069 ng/ul
RT: 13.157 min Scan# 1712
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:196 Resp: 357309
Ion Ratio Lower Upper
196 100
198 94.5 75.6 113.4
200 31.3 24.7 37.1




#43
1,1'-Biphenyl
Concen: 19.972 ng/ul
RT: 13.492 min Scan# 1769
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19



Tgt Ion:154 Resp: 1410788
Ion Ratio Lower Upper
154 100
153 39.6 31.4 47.2
76 12.9 10.4 15.6





#44

2-Chloronaphthalene

Concen: 19.664 ng/ul

RT: 13.533 min Scan# 1

Delta R.T. -0.000 min


Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:162 Resp: 1089942

Ion Ratio Lower Upper

162 100

164 32.3 26.2 39.2

127 39.4 32.6 49.0



#45

2-Nitroaniline

Concen: 21.034 ng/ul

RT: 13.727 min Scan# 1809

Delta R.T. -0.006 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19




Tgt Ion: 65 Resp: 286963


Ion Ratio Lower Upper

65 100

92 70.3 54.4 81.6

138 103.1 86.1 129.1









#47
 Dimethylphthalate
 Concen: 19.354 ng/ul
 RT: 14.110 min Scan# 1874
 Delta R.T. -0.000 min
 Lab File: BM036573.D
 Acq: 19 Sep 2022 10:19




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:163 Resp: 1326838
 Ion Ratio Lower Upper
 163 100
 194 4.0 3.0 4.6
 164 9.8 8.3 12.5



#48
 2,6-Dinitrotoluene
 Concen: 20.988 ng/ul
 RT: 14.221 min Scan# 1893
 Delta R.T. -0.006 min
 Lab File: BM036573.D
 Acq: 19 Sep 2022 10:19

Tgt Ion:165 Resp: 233932
 Ion Ratio Lower Upper
 165 100
 89 67.7 52.8 79.2
 121 23.2 19.4 29.0









#50
Acenaphthylene
Concen: 19.515 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:152 Resp: 1710424
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 13.0 10.5 15.7



#51
3-Nitroaniline
Concen: 19.184 ng/ul
RT: 14.545 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19


Tgt Ion:138 Resp: 256446
Ion Ratio Lower Upper
138 100
108 12.7 12.5 18.7
92 119.3 95.4 143.2








#52
Acenaphthene
Concen: 19.499 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:153 Resp: 1195773
Ion Ratio Lower Upper
153 100
152 48.9 38.2 57.2
154 88.3 71.0 106.6




#53
2,4-Dinitrophenol
Concen: 19.163 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19



Tgt Ion:184 Resp: 94773
Ion Ratio Lower Upper
184 100
63 79.0 58.8 88.2
154 68.7 55.0 82.4









#55
4-Nitrophenol
Concen: 18.666 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:109 Resp: 200016
Ion Ratio Lower Upper
109 100
139 128.9 102.2 153.4
65 128.6 105.7 158.5



#56
Dibenzofuran
Concen: 19.622 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:168 Resp: 1616007
Ion Ratio Lower Upper
168 100
139 39.3 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 20.839 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:165 Resp: 337662
Ion Ratio Lower Upper
165 100
63 47.3 39.7 59.5
182 2.5 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.380 ng/ul
RT: 15.262 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:232 Resp: 288842
Ion Ratio Lower Upper
232 100
131 52.8 39.1 58.7
130 2.8 1.8 2.6#
166 33.5 25.4 38.0









#59
Diethylphthalate
Concen: 19.166 ng/ul
RT: 15.468 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 1315332
Ion Ratio Lower Upper
149 100
177 21.4 17.3 25.9
150 12.2 10.1 15.1



#61
Fluorene
Concen: 19.493 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:166 Resp: 1367279
Ion Ratio Lower Upper
166 100
165 97.7 78.5 117.7
167 13.4 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 19.361 ng/ul
RT: 15.686 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:204 Resp: 649470
Ion Ratio Lower Upper
204 100
206 32.5 25.7 38.5
141 63.6 49.4 74.0



#63
4-Nitroaniline
Concen: 18.660 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:138 Resp: 242352
Ion Ratio Lower Upper
138 100
92 51.4 42.1 63.1
108 80.8 66.2 99.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.338 ng/ul
 RT: 15.757 min Scan# 2154
 Delta R.T. -0.006 min
 Lab File: BM036573.D
 Acq: 19 Sep 2022 10:19




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:198 Resp: 165869
 Ion Ratio Lower Upper
 198 100
 182 5.0 4.2 6.2
 77 30.6 24.0 36.0



#67
 N-Nitrosodiphenylamine
 Concen: 19.857 ng/ul
 RT: 15.898 min Scan# 2178
 Delta R.T. -0.006 min
 Lab File: BM036573.D
 Acq: 19 Sep 2022 10:19

Tgt Ion:169 Resp: 1123590
 Ion Ratio Lower Upper
 169 100
 168 67.0 54.6 81.8
 167 37.1 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 17.526 ng/ul
RT: 16.580 min Scan# 211
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:248 Resp: 331872
Ion Ratio Lower Upper
248 100
250 98.1 78.9 118.3
141 79.7 65.7 98.5



#69
Hexachlorobenzene
Concen: 18.753 ng/ul
RT: 16.686 min Scan# 2312
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19


Tgt Ion:284 Resp: 420005
Ion Ratio Lower Upper
284 100
142 42.5 30.8 46.2
249 32.3 24.2 36.2








#70
Atrazine
Concen: 17.950 ng/ul
RT: 16.845 min Scan# 2339
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Instrument :
BNA_M
ClientSampleId :







Tgt Ion:200 Resp: 358418
Ion Ratio Lower Upper
200 100
173 27.4 21.8 32.6
215 47.9 39.4 59.2



#71
Pentachlorophenol
Concen: 17.250 ng/ul
RT: 17.033 min Scan# 2371
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19


Tgt Ion:266 Resp: 229489
Ion Ratio Lower Upper
266 100
264 61.8 49.7 74.5
268 63.9 51.8 77.6








#72
Phenanthrene
Concen: 19.672 ng/ul
RT: 17.427 min Scan# 2438
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Instrument :
BNA_M
ClientSampleId :







Tgt Ion:178 Resp: 2081736
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 19.6 15.7 23.5



#74
Anthracene
Concen: 19.576 ng/ul
RT: 17.521 min Scan# 2454
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:178 Resp: 2105578
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.6 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.732 ng/uL
RT: 13.457 min Scan# 1763
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:216 Resp: 518491
Ion Ratio Lower Upper
216 100
214 78.8 62.6 94.0
218 49.3 38.2 57.2



#76
Pentachlorobenzene
Concen: 19.660 ng/uL
RT: 14.962 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:250 Resp: 558741
Ion Ratio Lower Upper
250 100
248 63.7 50.2 75.2
252 63.9 50.5 75.7





#77

Carbazole

Concen: 18.805 ng/ul

RT: 17.786 min Scan# 2499

Delta R.T. -0.000 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19





Instrument :

BNA_M

ClientSampleId :

Tgt Ion:167 Resp: 1872692

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.2	25.8
139	13.3	10.6	15.8



#78

Di-n-butylphthalate

Concen: 18.439 ng/ul

RT: 18.356 min Scan# 2596




Delta R.T. -0.006 min


Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Tgt Ion:149 Resp: 2081262

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.2	10.8
104	5.6	4.3	6.5









#80
Fluoranthene
Concen: 20.117 ng/ul
RT: 19.433 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:202 Resp: 2269750
Ion Ratio Lower Upper
202 100
101 11.7 9.8 14.6
100 8.9 7.4 11.2



#82
Pyrene
Concen: 20.322 ng/ul
RT: 19.792 min Scan# 2840
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:202 Resp: 2427099
Ion Ratio Lower Upper
202 100
101 14.2 10.9 16.3
100 11.7 8.7 13.1









#83
Butylbenzylphthalate
Concen: 19.350 ng/ul
RT: 20.692 min Scan# 2993
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 875727
Ion Ratio Lower Upper
149 100
91 74.3 55.2 82.8
206 19.6 16.2 24.2



#84
3,3'-Dichlorobenzidine
Concen: 19.281 ng/ul
RT: 21.462 min Scan# 3124
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:252 Resp: 705087
Ion Ratio Lower Upper
252 100
254 63.8 53.0 79.4
126 13.8 10.4 15.6





#85

Benzo(a)anthracene

Concen: 19.860 ng/ul

RT: 21.533 min Scan# 3136

Delta R.T. -0.006 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:228 Resp: 2328921

Ion Ratio Lower Upper

228 100

229 19.8 15.8 23.8

226 27.7 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.171 ng/ul

RT: 21.468 min Scan# 3125

Delta R.T. -0.012 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19




Tgt Ion:149 Resp: 1333490


Ion Ratio Lower Upper

149 100

167 27.3 22.0 33.0

279 3.8 3.3 4.9





#87

Chrysene

Concen: 19.685 ng/ul

RT: 21.586 min Scan# 3145

Delta R.T. -0.006 min


Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:228 Resp: 2191047

Ion Ratio Lower Upper

228 100

226 30.1 24.4 36.6

229 19.8 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 19.704 ng/ul




RT: 22.415 min Scan# 3286


Delta R.T. -0.006 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Tgt Ion:149 Resp: 2134797









#90
Benzo(b)fluoranthene
Concen: 20.301 ng/ul
RT: 23.244 min Scan# 3427
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 2313478
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.0 9.1 13.7



#91
Benzo(k)fluoranthene
Concen: 19.337 ng/ul
RT: 23.297 min Scan# 3436
Delta R.T. -0.006 min
Lab File: BM036573.D
Acq: 19 Sep 2022 10:19

Tgt Ion:252 Resp: 2201864
Ion Ratio Lower Upper
252 100
253 21.4 17.8 26.8
125 10.9 8.2 12.4





#93

Benzo(a)pyrene

Concen: 19.649 ng/ul

RT: 23.885 min Scan# 3536

Delta R.T. -0.006 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:252 Resp: 1968510

Ion Ratio Lower Upper

252 100

253 21.7 17.5 26.3

125 12.3 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.061 ng/ul

RT: 26.538 min Scan# 3987

Delta R.T. -0.012 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19




Tgt Ion:276 Resp: 2112942


Ion Ratio Lower Upper

276 100

138 26.6 20.2 30.4

277 24.8 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 18.275 ng/ul

RT: 26.562 min Scan# 3991

Delta R.T. -0.012 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:278 Resp: 1938920

Ion Ratio Lower Upper

278 100

139 17.3 14.5 21.7

279 23.7 19.7 29.5



#96

Benzo(g,h,i)perylene

Concen: 17.584 ng/ul

RT: 27.320 min Scan# 4120

Delta R.T. -0.012 min

Lab File: BM036573.D

Acq: 19 Sep 2022 10:19

Tgt Ion:276 Resp: 1783369

Ion Ratio Lower Upper

276 100

138 23.4 18.7 28.1

277 24.0 19.2 28.8

